

Data Path : Z:\HPCHEM1\BNA F\DATA\BF031517\
 Data File : BF093688.D
 Acq On : 16 Mar 2017 00:33
 Operator : SJ/MA
 Sample : I2264-16 5X
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 E2-EE9-ESW

Quant Time: Mar 16 03:37:52 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF030717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 13 15:05:08 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.71	152	166501	20.00	ng	-0.01
21) Naphthalene-d8	8.00	136	594259	20.00	ng	-0.01
38) Acenaphthene-d10	9.75	164	220688	20.00	ng	-0.01
63) Phenanthrene-d10	11.25	188	339644	20.00	ng	0.00
75) Chrysene-d12	13.89	240	352282	20.00	ng	0.00
86) Perylene-d12	15.29	264	295992	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.32	112	215695	21.56	ng	0.00
7) Phenol-d6	6.39	99	297702	23.60	ng	0.01
23) Nitrobenzene-d5	7.28	82	175691	16.40	ng	-0.02
41) 2,4,6-Tribromophenol	10.56	330	27601	19.20	ng	0.01
44) 2-Fluorobiphenyl	9.08	172	238598	20.11	ng	-0.01
78) Terphenyl-d14	12.84	244	146205	10.79	ng	0.01
Target Compounds						Qvalue
70) Phenanthrene	11.27	178	105834	5.46	ng	98
74) Fluoranthene	12.46	202	134399	7.18	ng	96
77) Pyrene	12.69	202	136251	5.32	ng	97
80) Benzo(a)anthracene	13.88	228	75466m	3.63	ng	
82) Chrysene	13.91	228	53191	2.76	ng	98
87) Benzo(b)fluoranthene	14.91	252	72238m	4.01	ng	
89) Benzo(a)pyrene	15.24	252	47506	2.88	ng	# 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF031517\
Data File : BF093688.D
Acq On : 16 Mar 2017 00:33
Operator : SJ/MA
Sample : I2264-16 5X
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
E2-EE9-ESW

Quant Time: Mar 16 03:37:52 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF030717.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 13 15:05:08 2017
Response via : Initial Calibration

