

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031522\
 Data File : BF127320.D
 Acq On : 15 Mar 2022 12:12
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 15 14:21:48 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031022.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 10 16:51:53 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	146	0.00
2	1,4-Dioxane	40.000	34.787	13.0	133	0.03
3	Pyridine	40.000	36.529	8.7	131	0.02
4	n-Nitrosodimethylamine	40.000	35.034	12.4	124	0.03
5 S	2-Fluorophenol	80.000	75.087	6.1	142	0.00
6	Aniline	40.000	38.377	4.1	137	0.00
7 S	Phenol-d6	80.000	75.274	5.9	138	0.00
8	2-Chlorophenol	40.000	37.673	5.8	136	0.00
9	Benzaldehyde	40.000	35.044	12.4	131	0.00
10 C	Phenol	40.000	37.720	5.7	136	0.00
11	bis(2-Chloroethyl)ether	40.000	38.420	3.9	136	0.00
12	1,3-Dichlorobenzene	40.000	38.121	4.7	139	0.00
13 C	1,4-Dichlorobenzene	40.000	37.698	5.8	136	0.00
14	1,2-Dichlorobenzene	40.000	36.641	8.4	133	0.00
15	Benzyl Alcohol	40.000	36.863	7.8	129	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	37.856	5.4	134	0.00
17	2-Methylphenol	40.000	37.118	7.2	135	0.00
18	Hexachloroethane	40.000	37.166	7.1	137	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	35.697	10.8	126	0.00
20	3+4-Methylphenols	40.000	34.850	12.9	129	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	147	0.00
22	Acetophenone	40.000	35.108	12.2	129	0.00
23 S	Nitrobenzene-d5	80.000	73.855	7.7	136	0.00
24	Nitrobenzene	40.000	37.535	6.2	136	0.00
25	Isophorone	40.000	37.803	5.5	140	0.00
26 C	2-Nitrophenol	40.000	38.813	3.0	144	0.00
27	2,4-Dimethylphenol	40.000	38.582	3.5	142	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.435	3.9	138	0.00
29 C	2,4-Dichlorophenol	40.000	38.103	4.7	137	0.00
30	1,2,4-Trichlorobenzene	40.000	38.018	5.0	139	0.00
31	Naphthalene	40.000	37.602	6.0	141	0.00
32	Benzoic acid	40.000	37.815	5.5	146	0.00
33	4-Chloroaniline	40.000	37.637	5.9	140	0.00
34 C	Hexachlorobutadiene	40.000	37.824	5.4	136	0.00
35	Caprolactam	40.000	38.018	5.0	130	0.00
36 C	4-Chloro-3-methylphenol	40.000	37.771	5.6	135	0.00
37	2-Methylnaphthalene	40.000	37.448	6.4	138	0.00
38	1-Methylnaphthalene	40.000	37.238	6.9	134	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	140	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.526	6.2	132	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.598	1.0	134	0.00
42 S	2,4,6-Tribromophenol	80.000	73.289	8.4	127	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.500	3.8	131	0.00
44	2,4,5-Trichlorophenol	40.000	39.003	2.5	135	0.00
45 S	2-Fluorobiphenyl	80.000	72.325	9.6	131	0.00
46	1,1'-Biphenyl	40.000	37.663	5.8	135	0.00
47	2-Chloronaphthalene	40.000	37.996	5.0	138	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031522\
 Data File : BF127320.D
 Acq On : 15 Mar 2022 12:12
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 15 14:21:48 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031022.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 10 16:51:53 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	38.335	4.2	133	0.00
49	Acenaphthylene	40.000	37.627	5.9	130	0.00
50	Dimethylphthalate	40.000	37.704	5.7	136	0.00
51	2,6-Dinitrotoluene	40.000	38.896	2.8	134	0.00
52 C	Acenaphthene	40.000	36.597	8.5	130	0.00
53	3-Nitroaniline	40.000	38.040	4.9	134	0.00
54 P	2,4-Dinitrophenol	40.000	35.372	11.6	118	0.00
55	Dibenzofuran	40.000	36.974	7.6	131	0.00
56 P	4-Nitrophenol	40.000	40.068	-0.2	128	0.00
57	2,4-Dinitrotoluene	40.000	37.712	5.7	128	0.00
58	Fluorene	40.000	36.222	9.4	127	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.637	3.4	133	0.00
60	Diethylphthalate	40.000	36.988	7.5	132	0.00
61	4-Chlorophenyl-phenylether	40.000	36.336	9.2	126	0.00
62	4-Nitroaniline	40.000	38.369	4.1	132	0.00
63	Azobenzene	40.000	37.930	5.2	128	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	137	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.301	-0.8	124	0.00
66 c	n-Nitrosodiphenylamine	40.000	37.867	5.3	131	0.00
67	4-Bromophenyl-phenylether	40.000	38.005	5.0	126	0.00
68	Hexachlorobenzene	40.000	37.309	6.7	123	0.00
69	Atrazine	40.000	38.843	2.9	125	0.00
70 C	Pentachlorophenol	40.000	40.861	-2.2	127	0.00
71	Phenanthrene	40.000	36.608	8.5	128	0.00
72	Anthracene	40.000	37.243	6.9	130	0.00
73	Carbazole	40.000	37.150	7.1	128	0.00
74	Di-n-butylphthalate	40.000	37.670	5.8	130	0.00
75 C	Fluoranthene	40.000	36.446	8.9	125	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	134	0.00
77	Benzydine	40.000	34.071	14.8	134	0.00
78	Pyrene	40.000	38.147	4.6	129	0.00
79 S	Terphenyl-d14	80.000	74.710	6.6	128	0.00
80	Butylbenzylphthalate	40.000	39.633	0.9	129	0.00
81	Benzo(a)anthracene	40.000	37.599	6.0	125	0.00
82	3,3'-Dichlorobenzidine	40.000	36.758	8.1	127	0.00
83	Chrysene	40.000	38.273	4.3	130	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	39.348	1.6	128	0.00
85 c	Di-n-octyl phthalate	40.000	40.779	-1.9	136	0.00
86 I	Perylene-d12	20.000	20.000	0.0	146	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	38.373	4.1	136	0.00
88	Benzo(b)fluoranthene	40.000	36.473	8.8	132	0.00
89	Benzo(k)fluoranthene	40.000	40.862	-2.2	151	0.00
90 C	Benzo(a)pyrene	40.000	39.022	2.4	139	0.00
91	Dibenzo(a,h)anthracene	40.000	38.049	4.9	135	0.00
92	Benzo(g,h,i)perylene	40.000	38.289	4.3	135	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031522\
Data File : BF127320.D
Acq On : 15 Mar 2022 12:12
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Mar 15 14:21:48 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031022.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 10 16:51:53 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0