

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031522\
 Data File : BF127323.D
 Acq On : 15 Mar 2022 14:40
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 16 00:58:12 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031022.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 10 16:51:53 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	146	0.00
2	1,4-Dioxane	40.000	32.453	18.9	125	0.03
3	Pyridine	40.000	36.773	8.1	133	0.02
4	n-Nitrosodimethylamine	40.000	35.190	12.0	125	0.03
5 S	2-Fluorophenol	80.000	74.697	6.6	142	0.00
6	Aniline	40.000	38.077	4.8	136	0.00
7 S	Phenol-d6	80.000	74.387	7.0	137	0.00
8	2-Chlorophenol	40.000	37.323	6.7	136	0.00
9	Benzaldehyde	40.000	35.871	10.3	134	0.00
10 C	Phenol	40.000	37.460	6.3	136	0.00
11	bis(2-Chloroethyl)ether	40.000	37.989	5.0	135	0.00
12	1,3-Dichlorobenzene	40.000	37.266	6.8	137	0.00
13 C	1,4-Dichlorobenzene	40.000	37.368	6.6	136	0.00
14	1,2-Dichlorobenzene	40.000	36.932	7.7	135	0.00
15	Benzyl Alcohol	40.000	36.280	9.3	127	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	37.646	5.9	134	0.00
17	2-Methylphenol	40.000	36.566	8.6	134	0.00
18	Hexachloroethane	40.000	37.065	7.3	137	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	35.960	10.1	127	0.00
20	3+4-Methylphenols	40.000	34.772	13.1	130	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	149	0.00
22	Acetophenone	40.000	34.640	13.4	129	0.00
23 S	Nitrobenzene-d5	80.000	72.671	9.2	135	0.00
24	Nitrobenzene	40.000	36.830	7.9	135	0.00
25	Isophorone	40.000	37.507	6.2	141	0.00
26 C	2-Nitrophenol	40.000	38.726	3.2	145	0.00
27	2,4-Dimethylphenol	40.000	38.585	3.5	144	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.156	4.6	138	0.00
29 C	2,4-Dichlorophenol	40.000	37.779	5.6	137	0.00
30	1,2,4-Trichlorobenzene	40.000	37.608	6.0	139	0.00
31	Naphthalene	40.000	37.492	6.3	143	0.00
32	Benzoic acid	40.000	36.953	7.6	143	0.00
33	4-Chloroaniline	40.000	36.074	9.8	136	0.00
34 C	Hexachlorobutadiene	40.000	36.983	7.5	135	0.00
35	Caprolactam	40.000	37.009	7.5	128	0.00
36 C	4-Chloro-3-methylphenol	40.000	37.569	6.1	136	0.00
37	2-Methylnaphthalene	40.000	37.450	6.4	140	0.00
38	1-Methylnaphthalene	40.000	37.154	7.1	136	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	141	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.396	4.0	136	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.099	-0.2	137	0.00
42 S	2,4,6-Tribromophenol	80.000	74.693	6.6	130	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.610	3.5	132	0.00
44	2,4,5-Trichlorophenol	40.000	39.030	2.4	136	0.00
45 S	2-Fluorobiphenyl	80.000	72.356	9.6	131	0.00
46	1,1'-Biphenyl	40.000	37.322	6.7	135	0.00
47	2-Chloronaphthalene	40.000	38.135	4.7	139	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031522\
 Data File : BF127323.D
 Acq On : 15 Mar 2022 14:40
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 16 00:58:12 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031022.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 10 16:51:53 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	38.789	3.0	135	0.00
49	Acenaphthylene	40.000	37.668	5.8	131	0.00
50	Dimethylphthalate	40.000	37.251	6.9	135	0.00
51	2,6-Dinitrotoluene	40.000	38.385	4.0	133	0.00
52 C	Acenaphthene	40.000	36.794	8.0	132	0.00
53	3-Nitroaniline	40.000	37.368	6.6	132	0.00
54 P	2,4-Dinitrophenol	40.000	36.747	8.1	124	0.00
55	Dibenzofuran	40.000	36.716	8.2	130	0.00
56 P	4-Nitrophenol	40.000	39.263	1.8	126	0.00
57	2,4-Dinitrotoluene	40.000	37.794	5.5	129	0.00
58	Fluorene	40.000	36.010	10.0	127	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.201	4.5	132	0.00
60	Diethylphthalate	40.000	37.245	6.9	134	0.00
61	4-Chlorophenyl-phenylether	40.000	36.355	9.1	127	0.00
62	4-Nitroaniline	40.000	37.716	5.7	131	0.00
63	Azobenzene	40.000	37.333	6.7	127	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	136	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.290	-3.2	126	0.00
66 c	n-Nitrosodiphenylamine	40.000	37.329	6.7	128	0.00
67	4-Bromophenyl-phenylether	40.000	38.065	4.8	126	0.00
68	Hexachlorobenzene	40.000	37.413	6.5	123	0.00
69	Atrazine	40.000	38.524	3.7	124	0.00
70 C	Pentachlorophenol	40.000	39.897	0.3	123	0.00
71	Phenanthrene	40.000	36.863	7.8	129	0.00
72	Anthracene	40.000	37.441	6.4	130	0.00
73	Carbazole	40.000	37.596	6.0	129	0.00
74	Di-n-butylphthalate	40.000	37.519	6.2	129	0.00
75 C	Fluoranthene	40.000	36.996	7.5	127	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	135	0.00
77	Benzydine	40.000	36.505	8.7	145	0.00
78	Pyrene	40.000	38.047	4.9	130	0.00
79 S	Terphenyl-d14	80.000	74.196	7.3	128	0.00
80	Butylbenzylphthalate	40.000	39.393	1.5	130	0.00
81	Benzo(a)anthracene	40.000	38.028	4.9	127	0.00
82	3,3'-Dichlorobenzidine	40.000	38.206	4.5	133	0.00
83	Chrysene	40.000	38.827	2.9	134	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	39.118	2.2	128	0.00
85 c	Di-n-octyl phthalate	40.000	41.661	-4.2	140	0.00
86 I	Perylene-d12	20.000	20.000	0.0	149	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	36.855	7.9	134	0.00
88	Benzo(b)fluoranthene	40.000	38.738	3.2	144	0.00
89	Benzo(k)fluoranthene	40.000	37.854	5.4	144	0.00
90 C	Benzo(a)pyrene	40.000	38.045	4.9	138	0.00
91	Dibenzo(a,h)anthracene	40.000	36.841	7.9	134	0.00
92	Benzo(g,h,i)perylene	40.000	37.140	7.1	134	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031522\
Data File : BF127323.D
Acq On : 15 Mar 2022 14:40
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Mar 16 00:58:12 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031022.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 10 16:51:53 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0