

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031524\
 Data File : BF137617.D
 Acq On : 15 Mar 2024 19:05
 Operator : CG\JU
 Sample : P1771-01DL 5X
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 COMP-1DL

Quant Time: Mar 15 23:24:00 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 14 14:45:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.722	152	127819	20.000	ng	0.00
21) Naphthalene-d8	8.004	136	502499	20.000	ng	0.00
39) Acenaphthene-d10	9.757	164	278061	20.000	ng	0.00
64) Phenanthrene-d10	11.239	188	499328	20.000	ng	0.00
76) Chrysene-d12	13.874	240	326790	20.000	ng	0.00
86) Perylene-d12	15.298	264	295085	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.328	112	123982	14.691	ng	0.00
7) Phenol-d6	6.357	99	170881	14.027	ng	-0.01
23) Nitrobenzene-d5	7.287	82	91039	8.418	ng	-0.01
42) 2,4,6-Tribromophenol	10.545	330	34573	14.158	ng	0.00
45) 2-Fluorobiphenyl	9.081	172	168545	9.488	ng	0.00
79) Terphenyl-d14	12.822	244	183679	7.729	ng	0.00
Target Compounds						
						Qvalue
75) Fluoranthene	12.451	202	67193	2.530	ng	96
78) Pyrene	12.674	202	69296	2.161	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031524\
Data File : BF137617.D
Acq On : 15 Mar 2024 19:05
Operator : CG\JU
Sample : P1771-01DL 5X
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
COMP-1DL

Quant Time: Mar 15 23:24:00 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 14 14:45:20 2024
Response via : Initial Calibration

