

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031524\
 Data File : BF137610.D
 Acq On : 15 Mar 2024 15:34
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 15 16:29:26 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 14 14:45:20 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	74	0.00
2	1,4-Dioxane	40.000	38.115	4.7	69	0.00
3	Pyridine	40.000	37.339	6.7	63	0.00
4	n-Nitrosodimethylamine	40.000	35.533	11.2	64	0.00
5 S	2-Fluorophenol	80.000	80.234	-0.3	71	0.00
6	Aniline	40.000	36.590	8.5	64	0.00
7 S	Phenol-d6	80.000	73.718	7.9	68	0.00
8	2-Chlorophenol	40.000	38.363	4.1	70	0.00
9	Benzaldehyde	40.000	37.257	6.9	64	0.00
10 C	Phenol	40.000	37.604	6.0	67	0.00
11	bis(2-Chloroethyl)ether	40.000	38.045	4.9	67	0.00
12	1,3-Dichlorobenzene	40.000	38.114	4.7	70	0.00
13 C	1,4-Dichlorobenzene	40.000	37.956	5.1	69	0.00
14	1,2-Dichlorobenzene	40.000	37.520	6.2	70	0.00
15	Benzyl Alcohol	40.000	37.644	5.9	65	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	33.970	15.1	63	0.00
17	2-Methylphenol	40.000	36.519	8.7	67	0.00
18	Hexachloroethane	40.000	39.080	2.3	71	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	34.459	13.9	65	0.00
20	3+4-Methylphenols	40.000	37.469	6.3	69	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	70	0.00
22	Acetophenone	40.000	38.500	3.8	67	0.00
23 S	Nitrobenzene-d5	80.000	77.451	3.2	66	0.00
24	Nitrobenzene	40.000	39.033	2.4	67	0.00
25	Isophorone	40.000	36.383	9.0	63	0.00
26 C	2-Nitrophenol	40.000	41.907	-4.8	70	0.00
27	2,4-Dimethylphenol	40.000	39.074	2.3	67	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.016	5.0	65	0.00
29 C	2,4-Dichlorophenol	40.000	39.909	0.2	68	0.00
30	1,2,4-Trichlorobenzene	40.000	39.262	1.8	68	0.00
31	Naphthalene	40.000	38.476	3.8	68	0.00
32	Benzoic acid	40.000	29.105	27.2#	51	0.00
33	4-Chloroaniline	40.000	39.530	1.2	65	0.00
34 C	Hexachlorobutadiene	40.000	39.767	0.6	69	0.00
35	Caprolactam	40.000	38.681	3.3	66	0.00
36 C	4-Chloro-3-methylphenol	40.000	37.819	5.5	64	0.00
37	2-Methylnaphthalene	40.000	37.816	5.5	66	0.00
38	1-Methylnaphthalene	40.000	37.538	6.2	66	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	68	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.585	1.0	67	0.00
41 P	Hexachlorocyclopentadiene	40.000	44.495	-11.2	83	0.00
42 S	2,4,6-Tribromophenol	80.000	78.569	1.8	67	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.164	2.1	65	0.00
44	2,4,5-Trichlorophenol	40.000	40.149	-0.4	67	0.00
45 S	2-Fluorobiphenyl	80.000	75.713	5.4	67	0.00
46	1,1'-Biphenyl	40.000	38.597	3.5	66	0.00
47	2-Chloronaphthalene	40.000	38.579	3.6	66	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031524\
 Data File : BF137610.D
 Acq On : 15 Mar 2024 15:34
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 15 16:29:26 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 14 14:45:20 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.215	-0.5	65	0.00
49	Acenaphthylene	40.000	38.091	4.8	65	0.00
50	Dimethylphthalate	40.000	37.831	5.4	65	0.00
51	2,6-Dinitrotoluene	40.000	39.644	0.9	66	0.00
52 C	Acenaphthene	40.000	37.663	5.8	65	0.00
53	3-Nitroaniline	40.000	39.852	0.4	64	0.00
54 P	2,4-Dinitrophenol	40.000	37.044	7.4	64	0.00
55	Dibenzofuran	40.000	38.082	4.8	65	0.00
56 P	4-Nitrophenol	40.000	37.903	5.2	64	0.00
57	2,4-Dinitrotoluene	40.000	40.881	-2.2	66	0.00
58	Fluorene	40.000	36.794	8.0	65	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	36.885	7.8	62	0.00
60	Diethylphthalate	40.000	38.578	3.6	66	0.00
61	4-Chlorophenyl-phenylether	40.000	37.190	7.0	66	0.00
62	4-Nitroaniline	40.000	35.520	11.2	59	0.00
63	Azobenzene	40.000	35.561	11.1	60	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	65	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.172	-0.4	62	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.699	3.3	64	0.00
67	4-Bromophenyl-phenylether	40.000	39.628	0.9	64	0.00
68	Hexachlorobenzene	40.000	39.700	0.7	66	0.00
69	Atrazine	40.000	39.759	0.6	65	0.00
70 C	Pentachlorophenol	40.000	36.106	9.7	56	0.00
71	Phenanthrene	40.000	37.935	5.2	64	0.00
72	Anthracene	40.000	38.944	2.6	65	0.00
73	Carbazole	40.000	38.481	3.8	63	0.00
74	Di-n-butylphthalate	40.000	38.716	3.2	63	0.00
75 C	Fluoranthene	40.000	38.455	3.9	65	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	79	0.00
77	Benzydine	40.000	28.694	28.3#	49	0.00
78	Pyrene	40.000	34.881	12.8	65	0.00
79 S	Terphenyl-d14	80.000	68.907	13.9	66	0.00
80	Butylbenzylphthalate	40.000	49.059	-22.6	92	0.00
81	Benzo(a)anthracene	40.000	36.770	8.1	72	0.00
82	3,3'-Dichlorobenzidine	40.000	41.452	-3.6	73	0.00
83	Chrysene	40.000	37.827	5.4	73	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	50.042	-25.1#	93	0.00
85 c	Di-n-octyl phthalate	40.000	41.461	-3.7	82	0.00
86 I	Perylene-d12	20.000	20.000	0.0	70	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	36.607	8.5	60	0.00
88	Benzo(b)fluoranthene	40.000	39.536	1.2	68	0.00
89	Benzo(k)fluoranthene	40.000	39.113	2.2	70	0.00
90 C	Benzo(a)pyrene	40.000	38.678	3.3	65	0.00
91	Dibenzo(a,h)anthracene	40.000	36.238	9.4	59	0.00
92	Benzo(g,h,i)perylene	40.000	37.202	7.0	61	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031524\
Data File : BF137610.D
Acq On : 15 Mar 2024 15:34
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Mar 15 16:29:26 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 14 14:45:20 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0