




Data Path : Z:\HPCHEM1\BNA_F\DATA\BF031616\
Data File : BF085467.D
Acq On : 16 Mar 2016 17:43
Operator : UM/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Mar 17 00:49:53 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF031516.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 15 18:57:05 2016
Response via : Initial Calibration



TIC: BF085467.D

(69) Pentachlorophenol (C)

11.236min (-0.012) 33691.08ng




response 99962

Ion	Exp%	Act%
265.70	100	100
268.00	62.50	65.37
264.00	64.30	63.13
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF031616\
 Data File : BF085467.D
 Acq On : 16 Mar 2016 17:43
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Quant Time: Mar 17 00:49:53 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF031516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 15 18:57:05 2016
 Response via : Initial Calibration



TIC: BF085467.D

(76) Benzidine

12.756min (-0.012) 18930.69ng




response 371623

Ion	Exp%	Act%
184.00	100	100
185.00	15.50	17.05
183.00	11.80	11.61
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF031616\
Data File : BF085467.D
Acq On : 16 Mar 2016 17:43
Operator : UM/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Mar 17 00:52:35 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF031516.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 15 18:57:05 2016
Response via : Initial Calibration



TIC: BF085467.D

(69) Pentachlorophenol (C)

11.236min (-0.012) 33691.08ng




response 99962

Ion	Exp%	Act%
265.70	100	100
268.00	62.50	65.37
264.00	64.30	63.13
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF031616\
 Data File : BF085467.D
 Acq On : 16 Mar 2016 17:43
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Quant Time: Mar 17 00:52:35 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF031516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 15 18:57:05 2016
 Response via : Initial Calibration



TIC: BF085467.D

(76) Benzidine

12.756min (-0.012) 18930.69ng




response 371623

Ion	Exp%	Act%
184.00	100	100
185.00	15.50	17.05
183.00	11.80	11.61
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF031616\
Data File : BF085483.D
Acq On : 17 Mar 2016 1:20
Operator : UM/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Mar 17 00:53:58 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF031516.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 15 18:57:05 2016
Response via : Initial Calibration



TIC: BF085483.D

(69) Pentachlorophenol (C)

11.225min (-0.023) 42445.00ng




response 136273

Ion	Exp%	Act%
265.70	100	100
268.00	62.50	61.37
264.00	64.30	65.09
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF031616\
Data File : BF085483.D
Acq On : 17 Mar 2016 1:20
Operator : UM/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Mar 17 00:53:58 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF031516.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 15 18:57:05 2016
Response via : Initial Calibration



TIC: BF085483.D

(76) Benzidine

12.756min (-0.012) 0.00ng

response 826195

Ion	Exp%	Act%
184.00	100	100
185.00	15.50	15.29
183.00	11.80	11.90
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF031616\
 Data File : BF085484.D
 Acq On : 17 Mar 2016 1:49
 Operator : UM/SJ
 Sample : SSTDC0040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Quant Time: Mar 17 03:05:46 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF031516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 15 18:57:05 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.90	152	231128	20.00	ng	-0.01
21) Naphthalene-d8	8.20	136	932848	20.00	ng	-0.01
38) Acenaphthene-d10	9.94	164	407302	20.00	ng	-0.01
63) Phenanthrene-d10	11.43	188	738611	20.00	ng	-0.01
75) Chrysene-d12	14.07	240	425669	20.00	ng	-0.01
86) Perylene-d12	15.51	264	380155	20.00	ng	-0.02

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
5) 2-Fluorophenol	5.50	112	973449	72.40	ng	-0.01
7) Phenol-d6	6.54	99	1313238	75.29	ng	-0.01
23) Nitrobenzene-d5	7.48	82	1165266	73.86	ng	-0.01
41) 2,4,6-Tribromophenol	10.74	330	322728	86.03	ng	-0.01
44) 2-Fluorobiphenyl	9.27	172	1734419	73.88	ng	-0.01
78) Terphenyl-d14	13.02	244	1760200	93.54	ng	-0.01

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) 1,4-Dioxane	2.52	88	267983	39.94	ng	96
3) Pyridine	3.27	79	612706	38.47	ng	97
4) n-Nitrosodimethylamine	3.22	42	285734	38.85	ng	98
6) Aniline	6.57	93	967134	40.17	ng	96
8) 2-Chlorophenol	6.69	128	595913	37.80	ng	98
9) Benzaldehyde	6.45	77	317671	32.62	ng	99
10) Phenol	6.55	94	714114	36.53	ng	94
11) bis(2-Chloroethyl)ether	6.64	93	552494	36.75	ng	96
12) 1,3-Dichlorobenzene	6.85	146	697068	39.89	ng	98
13) 1,4-Dichlorobenzene	6.85	146	697068	39.83	ng	96
14) 1,2-Dichlorobenzene	7.08	146	586508	38.18	ng	98
15) Benzyl Alcohol	7.05	79	501857	40.29	ng	96
16) 2,2'-oxybis(1-Chloropropan	7.18	45	819406	40.02	ng	96
17) 2-Methylphenol	7.17	107	545132	41.30	ng	97
18) Hexachloroethane	7.42	117	252455	39.32	ng	98
19) n-Nitroso-di-n-propylamine	7.33	70	371609	35.58	ng	# 92
20) 3+4-Methylphenols	7.32	107	525913	37.17	ng	# 84
22) Acetophenone	7.33	105	838678	36.69	ng	# 92
24) Nitrobenzene	7.50	77	711446	40.02	ng	# 84
25) Isophorone	7.74	82	1258992	39.13	ng	96
26) 2-Nitrophenol	7.81	139	335639	39.41	ng	# 75
27) 2,4-Dimethylphenol	7.85	122	607735	40.12	ng	99
28) bis(2-Chloroethoxy)methane	7.94	93	749458	38.96	ng	98
29) 2,4-Dichlorophenol	8.05	162	457925	37.27	ng	89
30) 1,2,4-Trichlorobenzene	8.14	180	535377	37.71	ng	94
31) Naphthalene	8.22	128	1778939	39.03	ng	99
32) Benzoic acid	7.99	122	431310	49.52	ng	97
33) 4-Chloroaniline	8.26	127	765149	39.02	ng	95
34) Hexachlorobutadiene	8.33	225	301150	40.57	ng	99
35) Caprolactam	8.65	113	162614	40.53	ng	93
36) 4-Chloro-3-methylphenol	8.74	107	503639	35.89	ng	99
37) 2-Methylnaphthalene	8.90	142	1011247	35.03	ng	97
39) 1,2,4,5-Tetrachlorobenzene	9.08	216	536001	42.31	ng	97
40) Hexachlorocyclopentadiene	9.06	237	262434	42.86	ng	97

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF031616\
 Data File : BF085484.D
 Acq On : 17 Mar 2016 1:49
 Operator : UM/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Quant Time: Mar 17 03:05:46 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF031516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 15 18:57:05 2016
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
42) 2,4,6-Trichlorophenol	9.18	196	340788	40.74	ng	97
43) 2,4,5-Trichlorophenol	9.22	196	350421	41.47	ng	97
45) 1,1'-Biphenyl	9.37	154	1362382	40.98	ng	95
46) 2-Chloronaphthalene	9.40	162	995963	39.98	ng	94
47) 2-Nitroaniline	9.49	65	293851	38.96	ng	95
48) Acenaphthylene	9.81	152	1512776	37.54	ng	99
49) Dimethylphthalate	9.67	163	1246086	41.93	ng	98
50) 2,6-Dinitrotoluene	9.74	165	308284	46.96	ng	# 86
51) Acenaphthene	9.98	154	991278	39.13	ng	99
52) 3-Nitroaniline	9.91	138	338802	43.16	ng	85
53) 2,4-Dinitrophenol	10.01	184	146608	44.14	ng	# 33
54) Dibenzofuran	10.15	168	1183369	38.02	ng	98
55) 4-Nitrophenol	10.06	139	212183	39.27	ng	# 73
56) 2,4-Dinitrotoluene	10.14	165	355448	43.17	ng	96
57) Fluorene	10.49	166	914414	36.84	ng	99
58) 2,3,4,6-Tetrachlorophenol	10.28	232	241696	41.65	ng	93
59) Diethylphthalate	10.37	149	1082216	37.62	ng	97
60) 4-Chlorophenyl-phenylether	10.48	204	486190	38.12	ng	92
61) 4-Nitroaniline	10.52	138	325446	44.09	ng	98
62) Azobenzene	10.64	77	1114322	39.01	ng	96
64) 4,6-Dinitro-2-methylphenol	10.55	198	208795	47.32	ng	93
65) n-Nitrosodiphenylamine	10.61	169	853966	37.52	ng	100
66) 4-Bromophenyl-phenylether	10.97	248	295814	39.57	ng	# 88
67) Hexachlorobenzene	11.04	284	302635	39.23	ng	# 84
68) Atrazine	11.13	200	236100	36.25	ng	97
69) Pentachlorophenol	11.24	266	190419	46.57	ng	98
70) Phenanthrene	11.45	178	1388258	36.09	ng	99
71) Anthracene	11.51	178	1576918	40.24	ng	99
72) Carbazole	11.66	167	1220786	36.82	ng	99
73) Di-n-butylphthalate	11.99	149	1641931	39.97	ng	99
74) Fluoranthene	12.64	202	1363837	37.71	ng	98
76) Benzidine	12.76	184	556118	42.06	ng	100
77) Pyrene	12.87	202	1360436	42.61	ng	99
79) Butylbenzylphthalate	13.49	149	653841	46.84	ng	# 81
80) Benzo(a)anthracene	14.06	228	953256	39.06	ng	99
81) 3,3'-Dichlorobenzidine	14.01	252	338858	41.02	ng	98
82) Chrysene	14.09	228	900107	41.00	ng	100
83) Bis(2-ethylhexyl)phthalate	14.05	149	771760	45.32	ng	# 95
84) Di-n-octyl phthalate	14.65	149	1123326	45.18	ng	99
85) Indeno(1,2,3-cd)pyrene	16.95	276	1136909	54.63	ng	97
87) Benzo(b)fluoranthene	15.09	252	967640	38.98	ng	100
88) Benzo(k)fluoranthene	15.09	252	967640	49.76	ng	99
89) Benzo(a)pyrene	15.45	252	827024	38.56	ng	99
90) Dibenzo(a,h)anthracene	16.97	278	932101	44.49	ng	96
91) Benzo(g,h,i)perylene	17.40	276	975561	45.59	ng	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

```
Data Path : Z:\HPCHEM1\BNA_F\DATA\BF031616\  
Data File : BF085484.D  
Acq On    : 17 Mar 2016    1:49  
Operator  : UM/SJ  
Sample    : SSTCCCC040  
Misc      :  
ALS Vial  : 2    Sample Multiplier: 1
```

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Mar 17 03:05:46 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF031516.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 15 18:57:05 2016
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 20.00 ng

RT: 6.90 min Scan# 404

Delta R.T. -0.01 min

Lab File: BF085484.D

Acq: 17 Mar 2016 1:49

Instrument :

BNA_F

ClientSampleId :

DFTPP


Tgt Ion:152 Resp: 231128

Ion Ratio Lower Upper

152 100

150 155.7 124.6 186.8




115 63.0 46.6 70.0



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane

Concen: 39.94 ng

RT: 2.52 min Scan# 20

Delta R.T. -0.03 min

Lab File: BF085484.D

Acq: 17 Mar 2016 1:49


Tgt Ion: 88 Resp: 267983

Ion Ratio Lower Upper

88 100

58 67.3 57.0 85.6



43 24.3 20.5 30.7




Abundance Ion 88.00 (87.70 to 88.70): BF0

Ion 58.00 (57.70 to 58.70): BF0

Ion 43.00 (42.70 to 43.70): BF0





#3

Pyridine

Concen: 38.47 ng

RT: 3.27 min Scan# 86

Delta R.T. -0.03 min

Lab File: BF085484.D

Acq: 17 Mar 2016 1:49

Instrument :

BNA_F

ClientSampleId :

DFTPP


Tgt Ion: 79 Resp: 612706

Ion Ratio Lower Upper

79 100

52 62.8 52.6 79.0




51 31.0 25.8 38.8



Abundance Ion 79.00 (78.70 to 79.70): BF0

Ion 52.00 (51.70 to 52.70): BF0

Ion 51.00 (50.70 to 51.70): BF0



#4

n-Nitrosodimethylamine

Concen: 38.85 ng

RT: 3.22 min Scan# 82

Delta R.T. -0.03 min

Lab File: BF085484.D

Acq: 17 Mar 2016 1:49


Tgt Ion: 42 Resp: 285734

Ion Ratio Lower Upper

42 100

74 149.2 116.8 175.2



44 7.1 5.5 8.3




Abundance Ion 42.00 (41.70 to 42.70): BF0

Ion 74.00 (73.70 to 74.70): BF0

Ion 44.00 (43.70 to 44.70): BF0





#5

2-Fluorophenol

Concen: 72.40 ng

RT: 5.50 min Scan# 281

Delta R.T. -0.01 min

Lab File: BF085484.D

Acq: 17 Mar 2016 1:49

Instrument :

BNA_F

ClientSampled :

DFTPP


Tgt Ion: 112 Resp: 973449

Ion Ratio Lower Upper

112 100

64 56.8 49.0 73.4




63 28.6 24.8 37.2



Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BF0

Ion 63.00 (62.70 to 63.70): BF0



#6

Aniline

Concen: 40.17 ng

RT: 6.57 min Scan# 375

Delta R.T. -0.01 min

Lab File: BF085484.D

Acq: 17 Mar 2016 1:49


Tgt Ion: 93 Resp: 967134

Ion Ratio Lower Upper

93 100

66 36.9 31.6 47.4



65 18.5 15.8 23.8




Abundance Ion 93.00 (92.70 to 93.70): BF0

Ion 66.00 (65.70 to 66.70): BF0

Ion 65.00 (64.70 to 65.70): BF0





#7

Phenol-d6

Concen: 75.29 ng

RT: 6.54 min Scan# 372

Delta R.T. -0.01 min

Lab File: BF085484.D

Acq: 17 Mar 2016 1:49

Instrument :

BNA_F

ClientSampleId :

DFTPP


Tgt Ion: 99 Resp: 1313238

Ion Ratio Lower Upper

99 100

42 15.3 13.6 20.4




71 32.5 26.7 40.1



Abundance Ion 99.00 (98.70 to 99.70): BF0

Ion 42.00 (41.70 to 42.70): BF0

Ion 71.00 (70.70 to 71.70): BF0



#8

2-Chlorophenol

Concen: 37.80 ng

RT: 6.69 min Scan# 385

Delta R.T. -0.01 min

Lab File: BF085484.D

Acq: 17 Mar 2016 1:49


Tgt Ion: 128 Resp: 595913

Ion Ratio Lower Upper

128 100

130 32.9 13.3 53.3



64 48.1 25.7 65.7




Abundance Ion 128.00 (127.70 to 128.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 64.00 (63.70 to 64.70): BF0





#9

Benzaldehyde

Concen: 32.62 ng

RT: 6.45 min Scan# 364

Delta R.T. -0.01 min

Lab File: BF085484.D

Acq: 17 Mar 2016 1:49

Instrument :

BNA_F

ClientSampleId :

DFTPP


Tgt Ion: 77 Resp: 317671

Ion Ratio Lower Upper

77 100

105 97.2 78.2 118.2




106 93.5 72.9 112.9



Abundance Ion 77.00 (76.70 to 77.70): BF0

Ion 105.00 (104.70 to 105.70): E

Ion 106.00 (105.70 to 106.70): E



#10

Phenol

Concen: 36.53 ng

RT: 6.55 min Scan# 373

Delta R.T. -0.01 min

Lab File: BF085484.D

Acq: 17 Mar 2016 1:49


Tgt Ion: 94 Resp: 714114

Ion Ratio Lower Upper

94 100

65 30.6 9.1 49.1



66 44.9 19.3 59.3




Abundance Ion 94.00 (93.70 to 94.70): BF0

Ion 65.00 (64.70 to 65.70): BF0

Ion 66.00 (65.70 to 66.70): BF0









#11
bis(2-Chloroethyl)ether
Concen: 36.75 ng
RT: 6.64 min Scan# 381
Delta R.T. -0.01 min
Lab File: BF085484.D
Acq: 17 Mar 2016 1:49

Instrument :
BNA_F
ClientSampleId :
DFTPP

Tgt Ion: 93 Resp: 552494
Ion Ratio Lower Upper
93 100
63 75.6 50.9 90.9
95 33.0 13.7 53.7




Abundance Ion 93.00 (92.70 to 93.70): BF0
Ion 63.00 (62.70 to 63.70): BF0
Ion 95.00 (94.70 to 95.70): BF0






#12
1,3-Dichlorobenzene
Concen: 39.89 ng
RT: 6.85 min Scan# 399
Delta R.T. -0.01 min
Lab File: BF085484.D
Acq: 17 Mar 2016 1:49

Tgt Ion: 146 Resp: 697068
Ion Ratio Lower Upper
146 100
148 64.8 52.1 78.1
75 26.8 19.5 29.3



Abundance Ion 146.00 (145.70 to 146.70): E
Ion 148.00 (147.70 to 148.70): E
Ion 75.00 (74.70 to 75.70): BF0









#13
 1,4-Dichlorobenzene
 Concen: 39.83 ng
 RT: 6.85 min Scan# 399
 Delta R.T. -0.08 min
 Lab File: BF085484.D
 Acq: 17 Mar 2016 1:49




Instrument :
 BNA_F
 ClientSampleId :
 DFTPP


Tgt Ion:146 Resp: 697068
 Ion Ratio Lower Upper
 146 100
 148 64.8 51.3 76.9
 111 42.2 38.4 57.6



#14
 1,2-Dichlorobenzene
 Concen: 38.18 ng
 RT: 7.08 min Scan# 419
 Delta R.T. -0.01 min
 Lab File: BF085484.D
 Acq: 17 Mar 2016 1:49

Tgt Ion:146 Resp: 586508
 Ion Ratio Lower Upper
 146 100
 148 63.8 51.4 77.2
 111 48.2 36.7 55.1









#15
Benzyl Alcohol
Concen: 40.29 ng
RT: 7.05 min Scan# 417
Delta R.T. -0.01 min
Lab File: BF085484.D
Acq: 17 Mar 2016 1:49




Instrument :
BNA_F
ClientSampleId :
DFTPP


Tgt Ion: 79 Resp: 501857
Ion Ratio Lower Upper
79 100
108 84.1 62.9 94.3
77 61.5 49.5 74.3



#16
2,2'-oxybis(1-Chloropropane)
Concen: 40.02 ng
RT: 7.18 min Scan# 428
Delta R.T. -0.01 min
Lab File: BF085484.D
Acq: 17 Mar 2016 1:49

Tgt Ion: 45 Resp: 819406
Ion Ratio Lower Upper
45 100
77 16.7 0.0 35.1
79 12.8 0.0 31.3









#17
2-Methylphenol
Concen: 41.30 ng
RT: 7.17 min Scan# 427
Delta R.T. -0.00 min
Lab File: BF085484.D
Acq: 17 Mar 2016 1:49

Instrument :
BNA_F
ClientSampleId :
DFTPP

Tgt Ion:107 Resp: 545132
Ion Ratio Lower Upper
107 100
108 115.1 90.9 136.3
77 44.6 32.6 48.8
79 44.0 32.8 49.2




Abundance Ion 107.00 (106.70 to 107.70): E
Ion 108.00 (107.70 to 108.70): E
Ion 77.00 (76.70 to 77.70): BF0
Ion 79.00 (78.70 to 79.70): BF0






#18
Hexachloroethane
Concen: 39.32 ng
RT: 7.42 min Scan# 449
Delta R.T. -0.01 min
Lab File: BF085484.D
Acq: 17 Mar 2016 1:49

Tgt Ion:117 Resp: 252455
Ion Ratio Lower Upper
117 100
119 96.1 78.2 117.4
201 91.3 72.0 108.0



Abundance Ion 117.00 (116.70 to 117.70): E
Ion 119.00 (118.70 to 119.70): E
Ion 201.00 (200.70 to 201.70): E









#19
n-Nitroso-di-n-propylamine
Concen: 35.58 ng
RT: 7.33 min Scan# 441
Delta R.T. -0.01 min
Lab File: BF085484.D
Acq: 17 Mar 2016 1:49




Instrument :
BNA_F
ClientSampleId :
DFTPP


Tgt Ion: 70 Resp: 371609
Ion Ratio Lower Upper
70 100
42 51.1 37.7 56.5
101 9.4 8.2 12.4
130 18.8 20.5 30.7#



#20
3+4-Methylphenols
Concen: 37.17 ng
RT: 7.32 min Scan# 440
Delta R.T. -0.01 min
Lab File: BF085484.D
Acq: 17 Mar 2016 1:49

Tgt Ion: 107 Resp: 525913
Ion Ratio Lower Upper
107 100
108 86.1 68.8 108.8
77 104.2 53.6 93.6#
79 26.1 8.0 48.0









#21
Naphthalene-d8
Concen: 20.00 ng
RT: 8.20 min Scan# 517
Delta R.T. -0.01 min
Lab File: BF085484.D
Acq: 17 Mar 2016 1:49

Instrument :
BNA_F
ClientSampleId :
DFTPP

Tgt Ion:	136	Resp:	932848
Ion Ratio	Lower	Upper	
136	100		
137	11.0	9.3	13.9
54	5.6	7.4	11.2#
68	5.0	6.0	9.0#




Abundance Ion 136.00 (135.70 to 136.70): E
Ion 137.00 (136.70 to 137.70): E
Ion 54.00 (53.70 to 54.70): BF0
Ion 68.00 (67.70 to 68.70): BF0






#22
Acetophenone
Concen: 36.69 ng
RT: 7.33 min Scan# 441
Delta R.T. -0.00 min
Lab File: BF085484.D
Acq: 17 Mar 2016 1:49

Tgt Ion:	105	Resp:	838678
Ion Ratio	Lower	Upper	
105	100		
71	9.3	4.1	6.1#
51	20.1	21.1	31.7#
120	22.9	17.3	25.9



Abundance Ion 105.00 (104.70 to 105.70): E
Ion 71.00 (70.70 to 71.70): BF0
Ion 51.00 (50.70 to 51.70): BF0
Ion 120.00 (119.70 to 120.70): E









#23
 Nitrobenzene-d5
 Concen: 73.86 ng
 RT: 7.48 min Scan# 454
 Delta R.T. -0.01 min
 Lab File: BF085484.D
 Acq: 17 Mar 2016 1:49




Instrument :
 BNA_F
 ClientSampleId :
 DFTPP


Tgt Ion: 82 Resp: 1165266
 Ion Ratio Lower Upper
 82 100
 128 46.8 34.5 51.7
 54 45.9 39.3 58.9



#24
 Nitrobenzene
 Concen: 40.02 ng
 RT: 7.50 min Scan# 456
 Delta R.T. -0.01 min
 Lab File: BF085484.D
 Acq: 17 Mar 2016 1:49

Tgt Ion: 77 Resp: 711446
 Ion Ratio Lower Upper
 77 100
 123 57.7 35.4 53.2#
 65 12.4 10.6 16.0









#25
Isophorone
Concen: 39.13 ng
RT: 7.74 min Scan# 477
Delta R.T. -0.01 min
Lab File: BF085484.D
Acq: 17 Mar 2016 1:49




Instrument :
BNA_F
ClientSampleId :
DFTPP


Tgt Ion: 82 Resp: 1258992
Ion Ratio Lower Upper
82 100
95 7.1 5.4 8.2
138 18.7 13.3 19.9



#26
2-Nitrophenol
Concen: 39.41 ng
RT: 7.81 min Scan# 483
Delta R.T. -0.01 min
Lab File: BF085484.D
Acq: 17 Mar 2016 1:49

Tgt Ion: 139 Resp: 335639
Ion Ratio Lower Upper
139 100
109 29.4 19.1 28.7#
65 49.4 23.9 35.9#





#27

2,4-Dimethylphenol

Concen: 40.12 ng

RT: 7.85 min Scan# 487

Delta R.T. -0.00 min

Lab File: BF085484.D

Acq: 17 Mar 2016 1:49

Instrument :

BNA_F

ClientSampleId :

DFTPP


Tgt Ion:122 Resp: 607735

Ion Ratio Lower Upper

122 100

107 104.8 84.8 127.2




121 58.6 47.0 70.6



Abundance Ion 122.00 (121.70 to 122.70): E

Ion 107.00 (106.70 to 107.70): E

Ion 121.00 (120.70 to 121.70): E



#28

bis(2-Chloroethoxy)methane

Concen: 38.96 ng

RT: 7.94 min Scan# 495

Delta R.T. -0.01 min

Lab File: BF085484.D

Acq: 17 Mar 2016 1:49


Tgt Ion: 93 Resp: 749458

Ion Ratio Lower Upper

93 100

95 33.3 25.9 38.9



123 11.9 8.8 13.2




Abundance Ion 93.00 (92.70 to 93.70): BF0

Ion 95.00 (94.70 to 95.70): BF0

Ion 123.00 (122.70 to 123.70): E









#29
 2,4-Dichlorophenol
 Concen: 37.27 ng
 RT: 8.05 min Scan# 504
 Delta R.T. -0.01 min
 Lab File: BF085484.D
 Acq: 17 Mar 2016 1:49

Instrument :
 BNA_F
 ClientSampled :
 DFTPP

Tgt Ion	Resp	Lower	Upper
162	100		
164	63.9	45.7	85.7
98	45.8	9.4	49.4




Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 164.00 (163.70 to 164.70): E
 Ion 98.00 (97.70 to 98.70): BF0






#30
 1,2,4-Trichlorobenzene
 Concen: 37.71 ng
 RT: 8.14 min Scan# 512
 Delta R.T. -0.00 min
 Lab File: BF085484.D
 Acq: 17 Mar 2016 1:49

Tgt Ion	Resp	Lower	Upper
180	100		
182	97.8	74.9	112.3
145	28.6	27.9	41.9



Abundance Ion 180.00 (179.70 to 180.70): E
 Ion 182.00 (181.70 to 182.70): E
 Ion 145.00 (144.70 to 145.70): E





#31

Naphthalene

Concen: 39.03 ng

RT: 8.22 min Scan# 519

Delta R.T. -0.01 min

Lab File: BF085484.D

Acq: 17 Mar 2016 1:49

Instrument :

BNA_F

ClientSampleId :

DFTPP


Tgt Ion:128 Resp: 1778939

Ion Ratio Lower Upper

128 100

129 11.3 9.7 14.5




127 13.9 11.2 16.8



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#32

Benzoic acid

Concen: 49.52 ng

RT: 7.99 min Scan# 499

Delta R.T. -0.00 min

Lab File: BF085484.D

Acq: 17 Mar 2016 1:49


Tgt Ion:122 Resp: 431310

Ion Ratio Lower Upper

122 100

105 122.9 101.3 141.3



77 90.5 74.7 114.7




Abundance Ion 122.00 (121.70 to 122.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 77.00 (76.70 to 77.70): BF0





#33

4-Chloroaniline

Concen: 39.02 ng

RT: 8.26 min Scan# 523

Delta R.T. -0.01 min

Lab File: BF085484.D

Acq: 17 Mar 2016 1:49

Instrument :

BNA_F

ClientSampleId :

DFTPP

Tgt Ion:127 Resp: 765149


Ion Ratio Lower Upper

127 100

129 32.6 26.1 39.1

65 27.3 27.0 40.6

92 20.4 17.4 26.0






Abundance Ion 127.00 (126.70 to 127.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 65.00 (64.70 to 65.70): BF0

Ion 92.00 (91.70 to 92.70): BF0



#34

Hexachlorobutadiene

Concen: 40.57 ng

RT: 8.33 min Scan# 529

Delta R.T. -0.01 min

Lab File: BF085484.D

Acq: 17 Mar 2016 1:49


Tgt Ion:225 Resp: 301150

Ion Ratio Lower Upper

225 100

223 63.0 49.7 74.5



227 64.5 52.2 78.4




Abundance Ion 225.00 (224.70 to 225.70): E

Ion 223.00 (222.70 to 223.70): E

Ion 227.00 (226.70 to 227.70): E









#35
 Caprolactam
 Concen: 40.53 ng
 RT: 8.65 min Scan# 557
 Delta R.T. -0.01 min
 Lab File: BF085484.D
 Acq: 17 Mar 2016 1:49

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Tgt Ion:113 Resp: 162614
 Ion Ratio Lower Upper
 113 100
 55 161.3 152.7 192.7
 56 122.3 109.0 149.0




Abundance Ion 113.00 (112.70 to 113.70): E
 Ion 55.00 (54.70 to 55.70): BF0
 Ion 56.00 (55.70 to 56.70): BF0






#36
 4-Chloro-3-methylphenol
 Concen: 35.89 ng
 RT: 8.74 min Scan# 565
 Delta R.T. -0.01 min
 Lab File: BF085484.D
 Acq: 17 Mar 2016 1:49

Tgt Ion:107 Resp: 503639
 Ion Ratio Lower Upper
 107 100
 144 23.8 19.4 29.2
 142 73.3 59.4 89.2



Abundance Ion 107.00 (106.70 to 107.70): E
 Ion 144.00 (143.70 to 144.70): E
 Ion 142.00 (141.70 to 142.70): E









#37
 2-Methylnaphthalene
 Concen: 35.03 ng
 RT: 8.90 min Scan# 579
 Delta R.T. -0.01 min
 Lab File: BF085484.D
 Acq: 17 Mar 2016 1:49

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Tgt Ion:142 Resp: 1011247
 Ion Ratio Lower Upper
 142 100
 141 90.3 71.7 107.5
 115 37.0 26.2 39.4




Abundance Ion 142.00 (141.70 to 142.70): E
 Ion 141.00 (140.70 to 141.70): E
 Ion 115.00 (114.70 to 115.70): E






#38
 Acenaphthene-d10
 Concen: 20.00 ng
 RT: 9.94 min Scan# 670
 Delta R.T. -0.01 min
 Lab File: BF085484.D
 Acq: 17 Mar 2016 1:49

Tgt Ion:164 Resp: 407302
 Ion Ratio Lower Upper
 164 100
 162 103.7 83.3 124.9
 160 46.4 37.9 56.9



Abundance Ion 164.00 (163.70 to 164.70): E
 Ion 162.00 (161.70 to 162.70): E
 Ion 160.00 (159.70 to 160.70): E





#39

1,2,4,5-Tetrachlorobenzene

Concen: 42.31 ng

RT: 9.08 min Scan# 594

Delta R.T. -0.01 min

Lab File: BF085484.D

Acq: 17 Mar 2016 1:49

Instrument :

BNA_F

ClientSampleId :

DFTPP

Tgt Ion:216 Resp: 536001


Ion Ratio Lower Upper

216 100

214 79.4 63.8 95.6

179 22.8 19.6 29.4

108 20.0 19.6 29.4






Abundance Ion 216.00 (215.70 to 216.70): E

Ion 214.00 (213.70 to 214.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 108.00 (107.70 to 108.70): E



#40

Hexachlorocyclopentadiene

Concen: 42.86 ng

RT: 9.06 min Scan# 593

Delta R.T. -0.00 min

Lab File: BF085484.D

Acq: 17 Mar 2016 1:49


Tgt Ion:237 Resp: 262434

Ion Ratio Lower Upper

237 100

235 61.9 43.7 83.7



272 14.7 0.0 31.5




Abundance Ion 236.80 (236.50 to 237.50): E

Ion 234.90 (234.60 to 235.60): E

Ion 271.75 (271.45 to 272.45): E









#41
 2,4,6-Tribromophenol
 Concen: 86.03 ng
 RT: 10.74 min Scan# 740
 Delta R.T. -0.01 min
 Lab File: BF085484.D
 Acq: 17 Mar 2016 1:49




Instrument :
 BNA_F
 ClientSampleId :
 DFTPP


Tgt Ion:330 Resp: 322728
 Ion Ratio Lower Upper
 330 100
 332 97.0 77.1 115.7
 141 31.4 35.0 52.6#



#42
 2,4,6-Trichlorophenol
 Concen: 40.74 ng
 RT: 9.18 min Scan# 603
 Delta R.T. -0.01 min
 Lab File: BF085484.D
 Acq: 17 Mar 2016 1:49

Tgt Ion:196 Resp: 340788
 Ion Ratio Lower Upper
 196 100
 198 94.1 77.8 116.6
 200 28.9 25.0 37.6









#43
 2,4,5-Trichlorophenol
 Concen: 41.47 ng
 RT: 9.22 min Scan# 607
 Delta R.T. -0.01 min
 Lab File: BF085484.D
 Acq: 17 Mar 2016 1:49

Instrument :
 BNA_F
 ClientSampleID :
 DFTPP

Tgt Ion:196 Resp: 350421
 Ion Ratio Lower Upper
 196 100
 198 94.2 75.7 113.5
 97 52.4 46.4 69.6
 132 31.2 26.2 39.4




Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 97.00 (96.70 to 97.70): BF0
 Ion 132.00 (131.70 to 132.70): E






#44
 2-Fluorobiphenyl
 Concen: 73.88 ng
 RT: 9.27 min Scan# 611
 Delta R.T. -0.01 min
 Lab File: BF085484.D
 Acq: 17 Mar 2016 1:49

Tgt Ion:172 Resp: 1734419
 Ion Ratio Lower Upper
 172 100
 171 37.7 29.3 43.9
 170 25.4 20.0 30.0



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E









#45
 1,1'-Biphenyl
 Concen: 40.98 ng
 RT: 9.37 min Scan# 620
 Delta R.T. -0.01 min
 Lab File: BF085484.D
 Acq: 17 Mar 2016 1:49

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Tgt Ion:154 Resp: 1362382
 Ion Ratio Lower Upper
 154 100
 153 42.9 21.4 61.4
 76 12.4 0.0 37.5




Abundance Ion 154.00 (153.70 to 154.70): E
 Ion 153.00 (152.70 to 153.70): E
 Ion 76.00 (75.70 to 76.70): BF0






#46
 2-Chloronaphthalene
 Concen: 39.98 ng
 RT: 9.40 min Scan# 622
 Delta R.T. -0.01 min
 Lab File: BF085484.D
 Acq: 17 Mar 2016 1:49

Tgt Ion:162 Resp: 995963
 Ion Ratio Lower Upper
 162 100
 127 45.9 31.6 47.4
 164 33.5 27.1 40.7



Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 127.00 (126.70 to 127.70): E
 Ion 164.00 (163.70 to 164.70): E









#47
2-Nitroaniline
Concen: 38.96 ng
RT: 9.49 min Scan# 630
Delta R.T. -0.01 min
Lab File: BF085484.D
Acq: 17 Mar 2016 1:49




Instrument :
BNA_F
ClientSampleId :
DFTPP


Tgt Ion: 65 Resp: 293851
Ion Ratio Lower Upper
65 100
92 69.8 55.4 83.2
138 102.6 75.7 113.5



#48
Acenaphthylene
Concen: 37.54 ng
RT: 9.81 min Scan# 658
Delta R.T. -0.01 min
Lab File: BF085484.D
Acq: 17 Mar 2016 1:49

Tgt Ion: 152 Resp: 1512776
Ion Ratio Lower Upper
152 100
151 21.2 17.5 26.3
153 14.1 11.3 16.9









#49
Dimethylphthalate
Concen: 41.93 ng
RT: 9.67 min Scan# 646
Delta R.T. -0.01 min
Lab File: BF085484.D
Acq: 17 Mar 2016 1:49

Instrument :
BNA_F
ClientSampled :
DFTPP

Tgt Ion:163 Resp: 1246086
Ion Ratio Lower Upper
163 100
194 3.3 3.1 4.7
164 10.9 8.2 12.4




Abundance Ion 163.00 (162.70 to 163.70): E
Ion 194.00 (193.70 to 194.70): E
Ion 164.00 (163.70 to 164.70): E






#50
2,6-Dinitrotoluene
Concen: 46.96 ng
RT: 9.74 min Scan# 652
Delta R.T. -0.01 min
Lab File: BF085484.D
Acq: 17 Mar 2016 1:49

Tgt Ion:165 Resp: 308284
Ion Ratio Lower Upper
165 100
63 40.9 41.8 62.8#
89 49.0 46.7 70.1



Abundance Ion 165.00 (164.70 to 165.70): E
Ion 63.00 (62.70 to 63.70): BF0
Ion 89.00 (88.70 to 89.70): BF0





#51

Acenaphthene

Concen: 39.13 ng

RT: 9.98 min Scan# 673

Delta R.T. -0.01 min

Lab File: BF085484.D

Acq: 17 Mar 2016 1:49

Instrument :

BNA_F

ClientSampleId :

DFTPP


Tgt Ion:154 Resp: 991278

Ion Ratio Lower Upper

154 100

153 112.6 89.8 134.8




152 55.0 44.6 66.8



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#52

3-Nitroaniline

Concen: 43.16 ng

RT: 9.91 min Scan# 667

Delta R.T. -0.01 min

Lab File: BF085484.D

Acq: 17 Mar 2016 1:49


Tgt Ion:138 Resp: 338802

Ion Ratio Lower Upper

138 100

108 8.7 7.8 11.6



92 85.5 81.3 121.9




Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BF0









#53
2,4-Dinitrophenol
Concen: 44.14 ng
RT: 10.01 min Scan# 676
Delta R.T. -0.01 min
Lab File: BF085484.D
Acq: 17 Mar 2016 1:49

Instrument :
BNA_F
ClientSampleId :
DFTPP

Tgt Ion:184 Resp: 146608
Ion Ratio Lower Upper
184 100
63 33.3 82.2 123.4#
154 57.5 108.9 163.3#




Abundance Ion 184.00 (183.70 to 184.70): E
Ion 63.00 (62.70 to 63.70): BF0
Ion 154.00 (153.70 to 154.70): E






#54
Dibenzofuran
Concen: 38.02 ng
RT: 10.15 min Scan# 688
Delta R.T. -0.01 min
Lab File: BF085484.D
Acq: 17 Mar 2016 1:49

Tgt Ion:168 Resp: 1183369
Ion Ratio Lower Upper
168 100
139 43.5 33.6 50.4
169 14.1 11.2 16.8



Abundance Ion 168.00 (167.70 to 168.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 169.00 (168.70 to 169.70): E









#55
4-Nitrophenol
Concen: 39.27 ng
RT: 10.06 min Scan# 680
Delta R.T. -0.01 min
Lab File: BF085484.D
Acq: 17 Mar 2016 1:49




Instrument :
BNA_F
ClientSampleId :
DFTPP


Tgt Ion:139 Resp: 212183
Ion Ratio Lower Upper
139 100
109 73.1 41.9 81.9
65 99.1 48.3 88.3#



#56
2,4-Dinitrotoluene
Concen: 43.17 ng
RT: 10.14 min Scan# 687
Delta R.T. -0.01 min
Lab File: BF085484.D
Acq: 17 Mar 2016 1:49

Tgt Ion:165 Resp: 355448
Ion Ratio Lower Upper
165 100
63 41.7 35.4 53.0
89 73.2 61.9 92.9
182 2.5 1.9 2.9









#57
 Fluorene
 Concen: 36.84 ng
 RT: 10.49 min Scan# 718
 Delta R.T. -0.01 min
 Lab File: BF085484.D
 Acq: 17 Mar 2016 1:49




Instrument :
 BNA_F
 ClientSampleId :
 DFTPP


Tgt Ion:166 Resp: 914414
 Ion Ratio Lower Upper
 166 100
 165 100.3 79.4 119.0
 167 13.9 11.0 16.6



#58
 2,3,4,6-Tetrachlorophenol
 Concen: 41.65 ng
 RT: 10.28 min Scan# 699
 Delta R.T. -0.01 min
 Lab File: BF085484.D
 Acq: 17 Mar 2016 1:49

Tgt Ion:232 Resp: 241696
 Ion Ratio Lower Upper
 232 100
 131 52.2 46.0 69.0
 130 2.5 2.2 3.4
 166 31.9 28.8 43.2









#59
Diethylphthalate
Concen: 37.62 ng
RT: 10.37 min Scan# 707
Delta R.T. -0.01 min
Lab File: BF085484.D
Acq: 17 Mar 2016 1:49

Instrument :
BNA_F
ClientSampleId :
DFTPP

Tgt Ion:149 Resp: 1082216
Ion Ratio Lower Upper
149 100
177 22.0 19.2 28.8
150 13.3 9.9 14.9




Abundance Ion 149.00 (148.70 to 149.70): E
Ion 177.00 (176.70 to 177.70): E
Ion 150.00 (149.70 to 150.70): E






#60
4-Chlorophenyl-phenylether
Concen: 38.12 ng
RT: 10.48 min Scan# 717
Delta R.T. -0.01 min
Lab File: BF085484.D
Acq: 17 Mar 2016 1:49

Tgt Ion:204 Resp: 486190
Ion Ratio Lower Upper
204 100
206 32.8 26.5 39.7
141 78.2 55.4 83.2



Abundance Ion 204.00 (203.70 to 204.70): E
Ion 206.00 (205.70 to 206.70): E
Ion 141.00 (140.70 to 141.70): E





#61

4-Nitroaniline

Concen: 44.09 ng

RT: 10.52 min Scan# 720

Delta R.T. -0.01 min

Lab File: BF085484.D

Acq: 17 Mar 2016 1:49

Instrument :

BNA_F

ClientSampleId :

DFTPP


Tgt Ion:138 Resp: 325446

Ion Ratio Lower Upper

138 100

92 51.2 34.2 74.2




108 73.8 53.9 93.9



Abundance Ion 138.00 (137.70 to 138.70): E

Ion 92.00 (91.70 to 92.70): BF0

Ion 108.00 (107.70 to 108.70): E



#62

Azobenzene

Concen: 39.01 ng

RT: 10.64 min Scan# 731

Delta R.T. -0.01 min

Lab File: BF085484.D

Acq: 17 Mar 2016 1:49

Tgt Ion: 77 Resp: 1114322


Ion Ratio Lower Upper

77 100

182 18.1 2.1 42.1

105 22.1 3.8 43.8

51 27.6 7.8 47.8






Abundance Ion 77.00 (76.70 to 77.70): BF0

Ion 182.00 (181.70 to 182.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 51.00 (50.70 to 51.70): BF0









#63
 Phenanthrene-d10
 Concen: 20.00 ng
 RT: 11.43 min Scan# 800
 Delta R.T. -0.01 min
 Lab File: BF085484.D
 Acq: 17 Mar 2016 1:49




Instrument :
 BNA_F
 ClientSampleId :
 DFTPP


Tgt Ion:188 Resp: 738611
 Ion Ratio Lower Upper
 188 100
 94 10.7 7.0 10.6#
 80 11.0 7.4 11.0



#64
 4,6-Dinitro-2-methylphenol
 Concen: 47.32 ng
 RT: 10.55 min Scan# 723
 Delta R.T. -0.01 min
 Lab File: BF085484.D
 Acq: 17 Mar 2016 1:49

Tgt Ion:198 Resp: 208795
 Ion Ratio Lower Upper
 198 100
 51 33.5 9.5 49.5
 105 44.5 20.5 60.5





#65

n-Nitrosodiphenylamine

Concen: 37.52 ng

RT: 10.61 min Scan# 728

Delta R.T. -0.01 min

Lab File: BF085484.D

Acq: 17 Mar 2016 1:49

Instrument :

BNA_F

ClientSampleId :

DFTPP


Tgt Ion:169 Resp: 853966

Ion Ratio Lower Upper

169 100

168 68.0 54.8 82.2




167 38.0 30.3 45.5



Abundance Ion 169.00 (168.70 to 169.70): E

Ion 168.00 (167.70 to 168.70): E

Ion 167.00 (166.70 to 167.70): E



#66

4-Bromophenyl-phenylether

Concen: 39.57 ng

RT: 10.97 min Scan# 760

Delta R.T. -0.01 min

Lab File: BF085484.D

Acq: 17 Mar 2016 1:49


Tgt Ion:248 Resp: 295814

Ion Ratio Lower Upper

248 100

250 95.4 77.1 115.7



141 104.8 66.2 99.2#




Abundance Ion 248.00 (247.70 to 248.70): E

Ion 250.00 (249.70 to 250.70): E

Ion 141.00 (140.70 to 141.70): E









#67
Hexachlorobenzene
Concen: 39.23 ng
RT: 11.04 min Scan# 766
Delta R.T. -0.01 min
Lab File: BF085484.D
Acq: 17 Mar 2016 1:49

Instrument :
BNA_F
ClientSampleId :
DFTPP

Tgt Ion	Ratio	Lower	Upper
284	100		
142	46.1	24.9	37.3#
249	33.7	25.0	37.4




Abundance Ion 283.80 (283.50 to 284.50): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E






#68
Atrazine
Concen: 36.25 ng
RT: 11.13 min Scan# 774
Delta R.T. -0.01 min
Lab File: BF085484.D
Acq: 17 Mar 2016 1:49

Tgt Ion	Ratio	Lower	Upper
200	100		
173	28.8	10.5	50.5
215	45.2	23.4	63.4



Abundance Ion 200.00 (199.70 to 200.70): E
Ion 173.00 (172.70 to 173.70): E
Ion 215.00 (214.70 to 215.70): E









#69
 Pentachlorophenol
 Concen: 46.57 ng
 RT: 11.24 min Scan# 783
 Delta R.T. -0.01 min
 Lab File: BF085484.D
 Acq: 17 Mar 2016 1:49

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Tgt Ion:266 Resp: 190419
 Ion Ratio Lower Upper
 266 100
 268 63.5 50.0 75.0
 264 62.9 51.4 77.2




Abundance Ion 265.70 (265.40 to 266.40): E
 Ion 268.00 (267.70 to 268.70): E
 Ion 264.00 (263.70 to 264.70): E






#70
 Phenanthrene
 Concen: 36.09 ng
 RT: 11.45 min Scan# 802
 Delta R.T. -0.01 min
 Lab File: BF085484.D
 Acq: 17 Mar 2016 1:49

Tgt Ion:178 Resp: 1388258
 Ion Ratio Lower Upper
 178 100
 176 20.6 16.8 25.2
 179 16.4 13.1 19.7



Abundance Ion 178.00 (177.70 to 178.70): E
 Ion 176.00 (175.70 to 176.70): E
 Ion 179.00 (178.70 to 179.70): E





#71

Anthracene

Concen: 40.24 ng

RT: 11.51 min Scan# 807

Delta R.T. -0.01 min

Lab File: BF085484.D

Acq: 17 Mar 2016 1:49

Instrument :

BNA_F

ClientSampleId :

DFTPP


Tgt Ion:178 Resp: 1576918

Ion Ratio Lower Upper

178 100

176 20.0 15.9 23.9




179 16.2 12.5 18.7



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#72

Carbazole

Concen: 36.82 ng

RT: 11.66 min Scan# 820

Delta R.T. -0.01 min

Lab File: BF085484.D

Acq: 17 Mar 2016 1:49


Tgt Ion:167 Resp: 1220786

Ion Ratio Lower Upper

167 100

166 22.7 18.2 27.2



139 15.1 11.4 17.0




Abundance Ion 167.00 (166.70 to 167.70): E

Ion 166.00 (165.70 to 166.70): E

Ion 139.00 (138.70 to 139.70): E





#73

Di-n-butylphthalate

Concen: 39.97 ng

RT: 11.99 min Scan# 849

Delta R.T. -0.01 min

Lab File: BF085484.D

Acq: 17 Mar 2016 1:49

Instrument :

BNA_F

ClientSampleId :

DFTPP


Tgt Ion:149 Resp: 1641931

Ion Ratio Lower Upper

149 100

150 9.9 7.8 11.6




104 5.3 4.7 7.1



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E



#74

Fluoranthene

Concen: 37.71 ng

RT: 12.64 min Scan# 906

Delta R.T. -0.01 min

Lab File: BF085484.D

Acq: 17 Mar 2016 1:49


Tgt Ion:202 Resp: 1363837

Ion Ratio Lower Upper

202 100

101 13.5 0.0 31.8



203 18.6 0.0 38.4




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E





#75

Chrysene-d12

Concen: 20.00 ng

RT: 14.07 min Scan# 1031

Delta R.T. -0.01 min

Lab File: BF085484.D

Acq: 17 Mar 2016 1:49

Instrument :

BNA_F

ClientSampleId :

DFTPP


Tgt Ion:240 Resp: 425669

Ion Ratio Lower Upper

240 100

120 8.4 5.3 7.9#




236 26.5 20.7 31.1



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#76

Benzidine

Concen: 42.06 ng

RT: 12.76 min Scan# 916

Delta R.T. -0.01 min

Lab File: BF085484.D

Acq: 17 Mar 2016 1:49


Tgt Ion:184 Resp: 556118

Ion Ratio Lower Upper

184 100

185 15.5 12.4 18.6



183 11.7 9.4 14.2




Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E





#77

Pyrene

Concen: 42.61 ng

RT: 12.87 min Scan# 926

Delta R.T. -0.01 min

Lab File: BF085484.D

Acq: 17 Mar 2016 1:49

Instrument :

BNA_F

ClientSampleId :

DFTPP


Tgt Ion:202 Resp: 1360436

Ion Ratio Lower Upper

202 100

200 22.4 18.2 27.4




203 18.1 14.7 22.1



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E



#78

Terphenyl-d14

Concen: 93.54 ng

RT: 13.02 min Scan# 939

Delta R.T. -0.01 min

Lab File: BF085484.D

Acq: 17 Mar 2016 1:49


Tgt Ion:244 Resp: 1760200

Ion Ratio Lower Upper

244 100

212 6.8 6.4 9.6



122 8.0 11.4 17.2#




Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E





#79

Butylbenzylphthalate

Concen: 46.84 ng

RT: 13.49 min Scan# 980

Delta R.T. -0.01 min

Lab File: BF085484.D

Acq: 17 Mar 2016 1:49

Instrument :

BNA_F

ClientSampleId :

DFTPP


Tgt Ion:149 Resp: 653841

Ion Ratio Lower Upper

149 100

91 67.1 69.0 103.4#




206 19.7 12.3 18.5#



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 91.00 (90.70 to 91.70): BF0

Ion 206.00 (205.70 to 206.70): E



#80

Benzo(a)anthracene

Concen: 39.06 ng

RT: 14.06 min Scan# 1030

Delta R.T. -0.01 min

Lab File: BF085484.D

Acq: 17 Mar 2016 1:49


Tgt Ion:228 Resp: 953256

Ion Ratio Lower Upper

228 100

226 28.7 22.8 34.2



229 20.2 16.6 24.8




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E









#81
 3,3'-Dichlorobenzidine
 Concen: 41.02 ng
 RT: 14.01 min Scan# 1026
 Delta R.T. -0.02 min
 Lab File: BF085484.D
 Acq: 17 Mar 2016 1:49




Instrument :
 BNA_F
 ClientSampleId :
 DFTPP


Tgt Ion: 252 Resp: 338858
 Ion Ratio Lower Upper
 252 100
 254 63.7 51.4 77.0
 126 18.7 12.9 19.3



#82
 Chrysene
 Concen: 41.00 ng
 RT: 14.09 min Scan# 1033
 Delta R.T. -0.01 min
 Lab File: BF085484.D
 Acq: 17 Mar 2016 1:49

Tgt Ion: 228 Resp: 900107
 Ion Ratio Lower Upper
 228 100
 226 31.5 25.2 37.8
 229 20.0 16.0 24.0









#83
 Bis(2-ethylhexyl)phthalate
 Concen: 45.32 ng
 RT: 14.05 min Scan# 1029
 Delta R.T. -0.01 min
 Lab File: BF085484.D
 Acq: 17 Mar 2016 1:49




Instrument :
 BNA_F
 ClientSampleId :
 DFTPP


Tgt Ion:149 Resp: 771760
 Ion Ratio Lower Upper
 149 100
 167 27.6 20.3 30.5
 279 4.5 1.9 2.9#



#84
 Di-n-octyl phthalate
 Concen: 45.18 ng
 RT: 14.65 min Scan# 1082
 Delta R.T. -0.01 min
 Lab File: BF085484.D
 Acq: 17 Mar 2016 1:49

Tgt Ion:149 Resp: 1123326
 Ion Ratio Lower Upper
 149 100
 167 1.5 1.1 1.7
 43 9.4 7.8 11.8









#85
 Indeno(1,2,3-cd)pyrene
 Concen: 54.63 ng
 RT: 16.95 min Scan# 1283
 Delta R.T. -0.03 min
 Lab File: BF085484.D
 Acq: 17 Mar 2016 1:49




Instrument :
 BNA_F
 ClientSampled :
 DFTPP


Tgt Ion:276 Resp: 1136909
 Ion Ratio Lower Upper
 276 100
 138 30.2 26.4 39.6
 277 25.5 20.2 30.2



#86
 Perylene-d12
 Concen: 20.00 ng
 RT: 15.51 min Scan# 1157
 Delta R.T. -0.02 min
 Lab File: BF085484.D
 Acq: 17 Mar 2016 1:49

Tgt Ion:264 Resp: 380155
 Ion Ratio Lower Upper
 264 100
 260 24.4 19.2 28.8
 265 21.8 17.3 25.9









#87
Benzo(b)fluoranthene
Concen: 38.98 ng
RT: 15.09 min Scan# 1120
Delta R.T. -0.02 min
Lab File: BF085484.D
Acq: 17 Mar 2016 1:49




Instrument :
BNA_F
ClientSampleId :
DFTPP


Tgt Ion:252 Resp: 967640
Ion Ratio Lower Upper
252 100
253 22.3 17.8 26.6
125 13.8 11.3 16.9



#88
Benzo(k)fluoranthene
Concen: 49.76 ng
RT: 15.09 min Scan# 1120
Delta R.T. -0.06 min
Lab File: BF085484.D
Acq: 17 Mar 2016 1:49

Tgt Ion:252 Resp: 967640
Ion Ratio Lower Upper
252 100
253 22.3 17.8 26.6
125 13.8 10.4 15.6





#89

Benzo(a)pyrene

Concen: 38.56 ng

RT: 15.45 min Scan# 1152

Delta R.T. -0.02 min

Lab File: BF085484.D

Acq: 17 Mar 2016 1:49

Instrument :

BNA_F

ClientSampleId :

DFTPP


Tgt Ion:252 Resp: 827024

Ion Ratio Lower Upper

252 100

253 21.7 17.4 26.0




125 11.8 8.2 12.4



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#90

Dibenzo(a,h)anthracene

Concen: 44.49 ng

RT: 16.97 min Scan# 1285

Delta R.T. -0.02 min

Lab File: BF085484.D

Acq: 17 Mar 2016 1:49


Tgt Ion:278 Resp: 932101

Ion Ratio Lower Upper

278 100

139 18.0 17.3 25.9



279 24.1 19.2 28.8




Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E





#91

Benzo(g,h,i)perylene

Concen: 45.59 ng

RT: 17.40 min Scan# 1322

Delta R.T. -0.02 min

Lab File: BF085484.D

Acq: 17 Mar 2016 1:49

Instrument :

BNA_F

ClientSampleId :

DFTPP


Tgt Ion: 276 Resp: 975561

Ion Ratio Lower Upper

276 100

277 23.5 19.0 28.4

138 23.8 22.5 33.7



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

