

Data Path : Z:\HPCHEM1\BNA F\DATA\BF031616\
 Data File : BF085482.D
 Acq On : 17 Mar 2016 00:52
 Operator : UM/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 17 00:47:52 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF031516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 15 18:57:05 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	90	-0.01
2	1,4-Dioxane	40.000	39.457	1.4	93	-0.03
3	Pyridine	40.000	39.289	1.8	89	-0.03
4	n-Nitrosodimethylamine	40.000	39.210	2.0	90	-0.03
5 S	2-Fluorophenol	80.000	72.826	9.0	92	-0.01
6	Aniline	40.000	39.974	0.1	90	-0.01
7 S	Phenol-d6	80.000	76.645	4.2	91	-0.01
8	2-Chlorophenol	40.000	38.060	4.8	94	-0.01
9	Benzaldehyde	40.000	32.958	17.6	94	-0.01
10 C	Phenol	40.000	36.765	8.1	95	-0.01
11	bis(2-Chloroethyl)ether	40.000	37.078	7.3	89	-0.01
12	1,3-Dichlorobenzene	40.000	39.893	0.3	92	-0.01
13 C	1,4-Dichlorobenzene	40.000	39.061	2.3	96	0.00
14	1,2-Dichlorobenzene	40.000	38.629	3.4	92	-0.01
15	Benzyl Alcohol	40.000	40.008	-0.0	90	-0.01
16	2,2'-oxybis(1-Chloropropane)	40.000	40.332	-0.8	93	-0.01
17	2-Methylphenol	40.000	41.588	-4.0	92	-0.01
18	Hexachloroethane	40.000	39.533	1.2	91	-0.01
19 P	n-Nitroso-di-n-propylamine	40.000	36.002	10.0	91	-0.01
20	3+4-Methylphenols	40.000	37.686	5.8	95	-0.01
21 I	Naphthalene-d8	20.000	20.000	0.0	94	-0.01
22	Acetophenone	40.000	35.648	10.9	94	-0.01
23 S	Nitrobenzene-d5	80.000	73.840	7.7	90	-0.01
24	Nitrobenzene	40.000	40.247	-0.6	94	-0.01
25	Isophorone	40.000	38.374	4.1	91	-0.01
26 C	2-Nitrophenol	40.000	40.230	-0.6	90	-0.01
27	2,4-Dimethylphenol	40.000	39.294	1.8	94	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.278	4.3	91	-0.01
29 C	2,4-Dichlorophenol	40.000	37.582	6.0	93	-0.01
30	1,2,4-Trichlorobenzene	40.000	38.404	4.0	93	-0.01
31	Naphthalene	40.000	38.650	3.4	92	-0.01
32	Benzoic acid	40.000	40.173	-0.4	87	0.00
33	4-Chloroaniline	40.000	38.001	5.0	90	-0.01
34 C	Hexachlorobutadiene	40.000	41.051	-2.6	94	-0.01
35	Caprolactam	40.000	37.469	6.3	90	-0.01
36 C	4-Chloro-3-methylphenol	40.000	35.042	12.4	94	-0.01
37	2-Methylnaphthalene	40.000	35.392	11.5	95	-0.01
38 I	Acenaphthene-d10	20.000	20.000	0.0	93	-0.01
39	1,2,4,5-Tetrachlorobenzene	40.000	42.609	-6.5	93	-0.01
40 P	Hexachlorocyclopentadiene	40.000	43.162	-7.9	91	0.00
41 S	2,4,6-Tribromophenol	80.000	83.892	-4.9	88	-0.01
42 C	2,4,6-Trichlorophenol	40.000	40.855	-2.1	94	-0.01
43	2,4,5-Trichlorophenol	40.000	42.423	-6.1	94	-0.01
44 S	2-Fluorobiphenyl	80.000	76.722	4.1	95	-0.01

Data Path : Z:\HPCHEM1\BNA F\DATA\BF031616\
 Data File : BF085482.D
 Acq On : 17 Mar 2016 00:52
 Operator : UM/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 17 00:47:52 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF031516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 15 18:57:05 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	40.000	41.070	-2.7	94	-0.01
46	2-Chloronaphthalene	40.000	40.128	-0.3	96	-0.01
47	2-Nitroaniline	40.000	38.848	2.9	92	-0.01
48	Acenaphthylene	40.000	37.380	6.5	91	-0.01
49	Dimethylphthalate	40.000	40.973	-2.4	93	-0.01
50	2,6-Dinitrotoluene	40.000	46.399	-16.0	93	-0.01
51 C	Acenaphthene	40.000	39.123	2.2	91	-0.01
52	3-Nitroaniline	40.000	40.835	-2.1	86	-0.02
53 P	2,4-Dinitrophenol	40.000	42.622	-6.6	94	-0.01
54	Dibenzofuran	40.000	38.020	4.9	93	-0.01
55 P	4-Nitrophenol	40.000	38.150	4.6	89	-0.01
56	2,4-Dinitrotoluene	40.000	44.561	-11.4	97	-0.01
57	Fluorene	40.000	36.902	7.7	92	-0.01
58	2,3,4,6-Tetrachlorophenol	40.000	41.772	-4.4	93	-0.01
59	Diethylphthalate	40.000	38.144	4.6	96	-0.01
60	4-Chlorophenyl-phenylether	40.000	38.598	3.5	92	-0.01
61	4-Nitroaniline	40.000	40.732	-1.8	89	-0.01
62	Azobenzene	40.000	39.621	0.9	92	-0.01
63 I	Phenanthrene-d10	20.000	20.000	0.0	91	-0.01
64	4,6-Dinitro-2-methylphenol	40.000	47.750	-19.4	95	-0.01
65 c	n-Nitrosodiphenylamine	40.000	38.352	4.1	93	-0.01
66	4-Bromophenyl-phenylether	40.000	38.646	3.4	91	-0.01
67	Hexachlorobenzene	40.000	38.773	3.1	90	-0.01
68	Atrazine	40.000	36.371	9.1	89	-0.01
69 C	Pentachlorophenol	40.000	44.606	-11.5	93	-0.01
70	Phenanthrene	40.000	36.687	8.3	90	-0.01
71	Anthracene	40.000	40.084	-0.2	94	-0.01
72	Carbazole	40.000	36.033	9.9	89	-0.01
73	Di-n-butylphthalate	40.000	39.112	2.2	92	-0.01
74 C	Fluoranthene	40.000	36.554	8.6	91	-0.01
75 I	Chrysene-d12	20.000	20.000	0.0	75	-0.01
76	Benzidine	40.000	37.350	6.6	75	-0.01
77	Pyrene	40.000	44.495	-11.2	90	-0.01
78 S	Terphenyl-d14	80.000	93.739	-17.2	88	-0.01
79	Butylbenzylphthalate	40.000	47.220	-18.0	87	-0.01
80	Benzo(a)anthracene	40.000	40.179	-0.4	78	-0.01
81	3,3'-Dichlorobenzidine	40.000	41.427	-3.6	75	-0.02
82	Chrysene	40.000	42.553	-6.4	91	-0.01
83	Bis(2-ethylhexyl)phthalate	40.000	44.750	-11.9	84	-0.01
84 c	Di-n-octyl phthalate	40.000	47.236	-18.1	90	-0.01
85	Indeno(1,2,3-cd)pyrene	40.000	59.920	-49.8#	104	-0.03
86 I	Perylene-d12	20.000	20.000	0.0	95	-0.02
87	Benzo(b)fluoranthene	40.000	38.709	3.2	105	-0.02

Data Path : Z:\HPCHEM1\BNA F\DATA\BF031616\
Data File : BF085482.D
Acq On : 17 Mar 2016 00:52
Operator : UM/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Mar 17 00:47:52 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF031516.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 15 18:57:05 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	40.000	32.813	18.0	71	-0.02
89 C	Benzo(a)pyrene	40.000	37.763	5.6	92	-0.02
90	Dibenzo(a,h)anthracene	40.000	44.757	-11.9	103	-0.02
91	Benzo(a,h,i)perylene	40.000	45.603	-14.0	105	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0