


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF031616\
 Data File : BF085483.D
 Acq On : 17 Mar 2016 1:20
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF031516.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Mar 15 18:57:05 2016



AutoFind: Scans 821, 822, 823; Background Corrected with Scan 818

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.2	30850	PASS
68	69	0.00	2	1.8	552	PASS
69	198	0.00	100	36.2	30834	PASS
70	69	0.00	2	0.5	143	PASS
127	198	10	80	52.6	44872	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	85241	PASS
199	198	5	9	7.0	5944	PASS
275	198	10	60	23.0	19577	PASS
365	198	1	100	2.6	2251	PASS
441	442	0.01	24	15.5	10662	PASS
442	198	50	100	80.9	68972	PASS
443	442	15	24	18.5	12765	PASS