


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF031715\
 Data File : BF077789.D
 Acq On : 18 Mar 2015 1:01
 Operator : TP/IZ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF031115.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Mar 13 14:04:45 2015



AutoFind: Scans 1070, 1071, 1072; Background Corrected with Scan 1067

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.2	9505	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	34.8	9976	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	49.3	14142	PASS
197	198	0.00	2	0.5	146	PASS
198	198	100	100	100.0	28672	PASS
199	198	5	9	7.0	2005	PASS
275	198	10	60	24.9	7133	PASS
365	198	1	100	2.5	723	PASS
441	442	0.01	24	16.2	3668	PASS
442	198	50	100	79.1	22687	PASS
443	442	15	24	19.0	4316	PASS