

Data Path : Z:\HPCHEM1\BNA F\DATA\BF031815\
 Data File : BF077824.D
 Acq On : 19 Mar 2015 6:38
 Operator : TP/IZ
 Sample : G1535-01 5X
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 COMP1B-1-B-5(0-6)

Manual Integrations
 APPROVED

mohammad
 3/20/2015 1:48:40 PM

Quant Time: Mar 19 08:30:30 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF031115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 19 04:21:27 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.16	152	38589	20.00	ng	0.00
21) Naphthalene-d8	8.74	136	155774	20.00	ng	0.00
38) Acenaphthene-d10	10.91	164	79547	20.00	ng	0.00
63) Phenanthrene-d10	12.75	188	156749	20.00	ng	0.01
75) Chrysene-d12	16.04	240	181242	20.00	ng	0.01
86) Perylene-d12	17.84	264	182362	20.00	ng	0.09
System Monitoring Compounds						
5) 2-Fluorophenol	5.46	112	19030	8.61	ng	0.02
7) Phenol-d6	6.74	99	29834	10.92	ng	0.01
23) Nitrobenzene-d5	7.86	82	14852	6.25	ng	0.00
41) 2,4,6-Tribromophenol	11.89	330	3679	4.83	ng	0.01
44) 2-Fluorobiphenyl	10.09	172	38630	7.14	ng	0.00
78) Terphenyl-d14	14.74	244	50316	6.78	ng	0.00
Target Compounds						Qvalue
70) Phenanthrene	12.77	178	20720	2.30	ng	97
74) Fluoranthene	14.25	202	55718	5.61	ng	93
77) Pyrene	14.53	202	49652	4.36	ng	99
80) Benzo(a)anthracene	16.03	228	22692m	2.11	ng	
82) Chrysene	16.08	228	24163	2.50	ng	95
87) Benzo(b)fluoranthene	17.37	252	31539m	2.65	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF031815\
Data File : BF077824.D
Acq On : 19 Mar 2015 6:38
Operator : TP/IZ
Sample : G1535-01 5X
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
COMP1B-1-B-5(0-6)

Manual Integrations
APPROVED

mohammad
3/20/2015 1:48:40 PM

Quant Time: Mar 19 08:30:30 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF031115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 19 04:21:27 2015
Response via : Initial Calibration

