


Data Path : Z:\HPCHEM1\BNA_F\Data\BF031815\
 Data File : BF077789.D
 Acq On : 18 Mar 2015 8:56
 Operator : TP/IZ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF031115.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Mar 19 02:09:19 2015



AutoFind: Scans 1071, 1072, 1073; Background Corrected with Scan 1067

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.9	10581	PASS
68	69	0.00	2	1.2	122	PASS
69	198	0.00	100	34.5	10457	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	49.0	14826	PASS
197	198	0.00	2	0.3	103	PASS
198	198	100	100	100.0	30285	PASS
199	198	5	9	6.7	2039	PASS
275	198	10	60	25.4	7700	PASS
365	198	1	100	2.7	825	PASS
441	442	0.01	24	15.1	3842	PASS
442	198	50	100	84.0	25436	PASS
443	442	15	24	18.8	4772	PASS