

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF031816\
 Data File : BF085610.D
 Acq On : 20 Mar 2016 10:49
 Operator : UM/SJ
 Sample : H1815-12
 Misc :
 ALS Vial : 68 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-01-COMP

Quant Time: Mar 21 00:22:24 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF031816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 18 23:31:39 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.86	152	141320	20.00	ng	0.00
21) Naphthalene-d8	8.14	136	491222	20.00	ng	0.00
38) Acenaphthene-d10	9.90	164	214886	20.00	ng	0.00
63) Phenanthrene-d10	11.38	188	380452	20.00	ng	0.00
75) Chrysene-d12	14.01	240	273387	20.00	ng	0.00
86) Perylene-d12	15.44	264	237102	20.00	ng	0.00

System Monitoring Compounds

5) 2-Fluorophenol	5.46	112	984153	116.97	ng	0.01
7) Phenol-d6	6.49	99	1169824	108.41	ng	0.00
23) Nitrobenzene-d5	7.42	82	661300	78.14	ng	-0.01
41) 2,4,6-Tribromophenol	10.69	330	239878	113.96	ng	0.00
44) 2-Fluorobiphenyl	9.22	172	943955	72.13	ng	0.00
78) Terphenyl-d14	12.96	244	648295	57.06	ng	0.00

Target Compounds


						Qvalue
9) Benzaldehyde	6.40	77	4346	Below Cal	#	1
22) Acetophenone	7.25	105	24223	2.08	ng	# 1
25) Isophorone	7.67	82	74746	4.41	ng	# 1
31) Naphthalene	8.16	128	82387	3.32	ng	# 62
35) Caprolactam	8.57	113	49409	21.48	ng	# 38
49) Dimethylphthalate	9.61	163	140910	8.69	ng	# 91
50) 2,6-Dinitrotoluene	9.67	165	7356	2.15	ng	# 61
51) Acenaphthene	9.93	154	33806	2.46	ng	# 88
65) n-Nitrosodiphenylamine	10.56	169	53008	4.47	ng	# 51
68) Atrazine	11.26	200	17788	4.93	ng	74
70) Phenanthrene	11.41	178	158725	7.91	ng	96
71) Anthracene	11.41	178	158725	7.55	ng	97
74) Fluoranthene	12.60	202	264631	12.60	ng	92
77) Pyrene	12.82	202	239143	12.18	ng	99
80) Benzo(a)anthracene	14.00	228	112122	7.06	ng	95
82) Chrysene	14.04	228	72811	4.77	ng	97
87) Benzo(b)fluoranthene	15.03	252	140836	9.10	ng	# 95
88) Benzo(k)fluoranthene	15.03	252	140836	11.05	ng	# 93
89) Benzo(a)pyrene	15.37	252	78579	5.99	ng	# 89
91) Benzo(g,h,i)perylene	17.26	276	37080	3.50	ng	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF031816\
Data File : BF085610.D
Acq On : 20 Mar 2016 10:49
Operator : UM/SJ
Sample : H1815-12
Misc :
ALS Vial : 68 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-01-COMP

Quant Time: Mar 21 00:22:24 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF031816.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 18 23:31:39 2016
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 6.86 min Scan# 400
Delta R.T. 0.00 min
Lab File: BF085610.D
Acq: 20 Mar 2016 10:49

Instrument :
BNA_F
ClientSampleId :
SB-01-COMP

Tgt Ion:152 Resp: 141320
Ion Ratio Lower Upper
152 100
150 150.8 124.2 186.2
115 54.1 47.0 70.6




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5


2-Fluorophenol
Concen: 116.97 ng
RT: 5.46 min Scan# 278
Delta R.T. 0.01 min
Lab File: BF085610.D
Acq: 20 Mar 2016 10:49

Tgt Ion:112 Resp: 984153
Ion Ratio Lower Upper
112 100
64 48.7 39.9 59.9
63 24.6 20.2 30.2



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BF0
Ion 63.00 (62.70 to 63.70): BF0





#7

Phenol-d6

Concen: 108.41 ng

RT: 6.49 min Scan# 368

Delta R.T. 0.00 min

Lab File: BF085610.D

Acq: 20 Mar 2016 10:49

Instrument :

BNA_F

ClientSampleId :

SB-01-COMP


Tgt Ion: 99 Resp: 1169824

Ion Ratio Lower Upper

99 100

42 14.6 11.1 16.7




71 33.1 24.9 37.3



Abundance Ion 99.00 (98.70 to 99.70): BF0

Ion 42.00 (41.70 to 42.70): BF0

Ion 71.00 (70.70 to 71.70): BF0



#21

Naphthalene-d8

Concen: 20.00 ng

RT: 8.14 min Scan# 512

Delta R.T. 0.00 min

Lab File: BF085610.D

Acq: 20 Mar 2016 10:49

Tgt Ion: 136 Resp: 491222


Ion Ratio Lower Upper

136 100

137 12.7 9.1 13.7

54 9.9 7.4 11.0

68 9.9 5.8 8.8#






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF0

Ion 68.00 (67.70 to 68.70): BF0





#22

Acetophenone

Concen: 2.08 ng

RT: 7.25 min Scan# 434

Delta R.T. -0.02 min

Lab File: BF085610.D

Acq: 20 Mar 2016 10:49

Instrument :

BNA_F

ClientSampleId :

SB-01-COMP

Tgt Ion: 105 Resp: 24223


Ion Ratio Lower Upper

105 100

71 449.4 3.6 5.4#

51 20.4 25.4 38.0#

120 0.0 16.9 25.3#






Abundance Ion 105.00 (104.70 to 105.70): E

Ion 71.00 (70.70 to 71.70): BF0

Ion 51.00 (50.70 to 51.70): BF0

Ion 120.00 (119.70 to 120.70): E



#23

Nitrobenzene-d5

Concen: 78.14 ng

RT: 7.42 min Scan# 449

Delta R.T. -0.01 min

Lab File: BF085610.D

Acq: 20 Mar 2016 10:49


Tgt Ion: 82 Resp: 661300

Ion Ratio Lower Upper

82 100

128 44.5 41.8 62.8



54 44.9 33.4 50.0




Abundance Ion 82.00 (81.70 to 82.70): BF0

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BF0





#25

Isophorone

Concen: 4.41 ng

RT: 7.67 min Scan# 471

Delta R.T. -0.01 min

Lab File: BF085610.D

Acq: 20 Mar 2016 10:49

Instrument :

BNA_F

ClientSampleId :

SB-01-COMP





Tgt Ion: 82 Resp: 74746

Ion Ratio Lower Upper

82 100

95 184.6 5.4 8.2#

138 9.3 12.9 19.3#



#31

Naphthalene

Concen: 3.32 ng

RT: 8.16 min Scan# 514

Delta R.T. 0.00 min

Lab File: BF085610.D

Acq: 20 Mar 2016 10:49




Tgt Ion: 128 Resp: 82387


Ion Ratio Lower Upper

128 100

129 33.6 9.4 14.0#

127 22.8 11.0 16.6#





#35

Caprolactam

Concen: 21.48 ng

RT: 8.57 min Scan# 550

Delta R.T. -0.02 min

Lab File: BF085610.D

Acq: 20 Mar 2016 10:49

Instrument :

BNA_F

ClientSampleId :

SB-01-COMP


Tgt Ion:113 Resp: 49409

Ion Ratio Lower Upper

113 100

55 241.3 139.4 179.4#




56 188.2 100.4 140.4#



Abundance Ion 113.00 (112.70 to 113.70): E

Ion 55.00 (54.70 to 55.70): BF0

Ion 56.00 (55.70 to 56.70): BF0



#38

Acenaphthene-d10

Concen: 20.00 ng

RT: 9.90 min Scan# 666

Delta R.T. 0.00 min

Lab File: BF085610.D

Acq: 20 Mar 2016 10:49


Tgt Ion:164 Resp: 214886

Ion Ratio Lower Upper

164 100

162 101.1 82.1 123.1



160 48.0 37.4 56.2




Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E









#41
 2,4,6-Tribromophenol
 Concen: 113.96 ng
 RT: 10.69 min Scan# 735
 Delta R.T. 0.00 min
 Lab File: BF085610.D
 Acq: 20 Mar 2016 10:49

Instrument :
 BNA_F
 ClientSampleId :
 SB-01-COMP

Tgt Ion:330 Resp: 239878
 Ion Ratio Lower Upper
 330 100
 332 97.4 78.2 117.2
 141 29.3 31.5 47.3#




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E






#44
 2-Fluorobiphenyl
 Concen: 72.13 ng
 RT: 9.22 min Scan# 607
 Delta R.T. 0.00 min
 Lab File: BF085610.D
 Acq: 20 Mar 2016 10:49

Tgt Ion:172 Resp: 943955
 Ion Ratio Lower Upper
 172 100
 171 36.0 29.4 44.2
 170 24.2 19.8 29.6



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E





#49

Dimethylphthalate

Concen: 8.69 ng

RT: 9.61 min Scan# 641

Delta R.T. -0.01 min

Lab File: BF085610.D

Acq: 20 Mar 2016 10:49

Instrument :

BNA_F

ClientSampleId :

SB-01-COMP


Tgt Ion:163 Resp: 140910

Ion Ratio Lower Upper

163 100

194 11.6 2.9 4.3#




164 11.8 8.3 12.5



Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E



#50

2,6-Dinitrotoluene

Concen: 2.15 ng

RT: 9.67 min Scan# 646

Delta R.T. -0.01 min

Lab File: BF085610.D

Acq: 20 Mar 2016 10:49


Tgt Ion:165 Resp: 7356

Ion Ratio Lower Upper

165 100

63 40.6 51.8 77.8#



89 29.3 53.7 80.5#




Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BF0

Ion 89.00 (88.70 to 89.70): BF0





#51

Acenaphthene

Concen: 2.46 ng

RT: 9.93 min Scan# 669

Delta R.T. 0.00 min

Lab File: BF085610.D

Acq: 20 Mar 2016 10:49

Instrument :

BNA_F

ClientSampleId :

SB-01-COMP


Tgt Ion:154 Resp: 33806

Ion Ratio Lower Upper

154 100

153 120.4 90.1 135.1




152 71.6 44.2 66.2#



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#63

Phenanthrene-d10

Concen: 20.00 ng

RT: 11.38 min Scan# 796

Delta R.T. 0.00 min

Lab File: BF085610.D

Acq: 20 Mar 2016 10:49


Tgt Ion:188 Resp: 380452

Ion Ratio Lower Upper

188 100

94 11.0 5.4 8.0#



80 11.1 5.5 8.3#




Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BF0

Ion 80.00 (79.70 to 80.70): BF0









#65
n-Nitrosodiphenylamine
Concen: 4.47 ng
RT: 10.56 min Scan# 724
Delta R.T. 0.00 min
Lab File: BF085610.D
Acq: 20 Mar 2016 10:49




Instrument :
BNA_F
ClientSampleId :
SB-01-COMP


Tgt Ion:169 Resp: 53008
Ion Ratio Lower Upper
169 100
168 23.8 54.4 81.6#
167 13.4 29.4 44.0#



#68
Atrazine
Concen: 4.93 ng
RT: 11.26 min Scan# 785
Delta R.T. 0.17 min
Lab File: BF085610.D
Acq: 20 Mar 2016 10:49

Tgt Ion:200 Resp: 17788
Ion Ratio Lower Upper
200 100
173 42.7 7.7 47.7
215 29.9 25.4 65.4





#70

Phenanthrene

Concen: 7.91 ng

RT: 11.41 min Scan# 798

Delta R.T. 0.00 min

Lab File: BF085610.D

Acq: 20 Mar 2016 10:49

Instrument :

BNA_F

ClientSampleId :

SB-01-COMP


Tgt Ion:178 Resp: 158725

Ion Ratio Lower Upper

178 100

176 19.4 16.3 24.5




179 18.5 12.6 18.8



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#71

Anthracene

Concen: 7.55 ng

RT: 11.41 min Scan# 798

Delta R.T. -0.05 min

Lab File: BF085610.D

Acq: 20 Mar 2016 10:49


Tgt Ion:178 Resp: 158725

Ion Ratio Lower Upper

178 100

176 19.4 15.9 23.9



179 18.5 13.0 19.4




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E





#74

Fluoranthene

Concen: 12.60 ng

RT: 12.60 min Scan# 902

Delta R.T. 0.01 min

Lab File: BF085610.D

Acq: 20 Mar 2016 10:49

Instrument :

BNA_F

ClientSampleId :

SB-01-COMP


Tgt Ion:202 Resp: 264631

Ion Ratio Lower Upper

202 100

101 10.0 0.0 36.6




203 17.7 0.0 38.1



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E



#75

Chrysene-d12

Concen: 20.00 ng

RT: 14.01 min Scan# 1026

Delta R.T. 0.00 min

Lab File: BF085610.D

Acq: 20 Mar 2016 10:49


Tgt Ion:240 Resp: 273387

Ion Ratio Lower Upper

240 100

120 10.9 13.8 20.8#



236 26.5 20.6 30.8




Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E





#77

Pyrene

Concen: 12.18 ng

RT: 12.82 min Scan# 922

Delta R.T. 0.00 min

Lab File: BF085610.D

Acq: 20 Mar 2016 10:49

Instrument :

BNA_F

ClientSampleId :

SB-01-COMP


Tgt Ion:202 Resp: 239143

Ion Ratio Lower Upper

202 100

200 21.8 17.6 26.4




203 18.2 14.0 21.0



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E



#78

Terphenyl-d14

Concen: 57.06 ng

RT: 12.96 min Scan# 934

Delta R.T. 0.00 min

Lab File: BF085610.D

Acq: 20 Mar 2016 10:49


Tgt Ion:244 Resp: 648295

Ion Ratio Lower Upper

244 100

212 7.5 6.1 9.1



122 13.7 10.2 15.4




Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E





#80

Benzo(a)anthracene

Concen: 7.06 ng

RT: 14.00 min Scan# 1025

Delta R.T. 0.00 min

Lab File: BF085610.D

Acq: 20 Mar 2016 10:49

Instrument :

BNA_F

ClientSampleId :

SB-01-COMP


Tgt Ion:228 Resp: 112122

Ion Ratio Lower Upper

228 100

226 31.3 22.1 33.1




229 21.2 16.2 24.4



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#82

Chrysene

Concen: 4.77 ng

RT: 14.04 min Scan# 1028

Delta R.T. -0.01 min

Lab File: BF085610.D

Acq: 20 Mar 2016 10:49


Tgt Ion:228 Resp: 72811

Ion Ratio Lower Upper

228 100

226 31.1 25.0 37.6



229 23.4 15.7 23.5




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E









#86
Perylene-d12
Concen: 20.00 ng
RT: 15.44 min Scan# 1151
Delta R.T. 0.00 min
Lab File: BF085610.D
Acq: 20 Mar 2016 10:49




Instrument :
BNA_F
ClientSampleId :
SB-01-COMP


Tgt Ion:264 Resp: 237102
Ion Ratio Lower Upper
264 100
260 24.2 19.4 29.2
265 22.2 17.2 25.8



#87
Benzo(b)fluoranthene
Concen: 9.10 ng
RT: 15.03 min Scan# 1115
Delta R.T. 0.00 min
Lab File: BF085610.D
Acq: 20 Mar 2016 10:49

Tgt Ion:252 Resp: 140836
Ion Ratio Lower Upper
252 100
253 23.6 17.6 26.4
125 15.8 10.3 15.5#





#88

Benzo(k)fluoranthene

Concen: 11.05 ng

RT: 15.03 min Scan# 1115

Delta R.T. -0.03 min

Lab File: BF085610.D

Acq: 20 Mar 2016 10:49

Instrument :

BNA_F

ClientSampleId :

SB-01-COMP


Tgt Ion:252 Resp: 140836

Ion Ratio Lower Upper

252 100

253 23.6 17.8 26.6




125 15.8 7.7 11.5#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#89

Benzo(a)pyrene

Concen: 5.99 ng

RT: 15.37 min Scan# 1145

Delta R.T. -0.01 min

Lab File: BF085610.D

Acq: 20 Mar 2016 10:49


Tgt Ion:252 Resp: 78579

Ion Ratio Lower Upper

252 100

253 24.7 17.4 26.0



125 20.7 10.3 15.5#




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E





#91

Benzo(g,h,i)perylene

Concen: 3.50 ng

RT: 17.26 min Scan# 1310

Delta R.T. 0.00 min

Lab File: BF085610.D

Acq: 20 Mar 2016 10:49

Instrument :

BNA_F

ClientSampleId :

SB-01-COMP

Tgt Ion: 276 Resp: 37080

Ion Ratio Lower Upper

276 100

277 24.7 18.8 28.2

138 25.9 22.6 34.0

