

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF031816\
 Data File : BF085552.D
 Acq On : 19 Mar 2016 4:17
 Operator : UM/SJ
 Sample : PB88981BSD
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB88981BSD

Manual Integrations
 APPROVED

apatel
 3/21/2016 5:08:02 PM

Quant Time: Mar 19 04:41:44 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF031816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 18 23:31:39 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.86	152	127623	20.00	ng	0.00
21) Naphthalene-d8	8.14	136	523071	20.00	ng	0.00
38) Acenaphthene-d10	9.90	164	241797	20.00	ng	0.00
63) Phenanthrene-d10	11.37	188	443843	20.00	ng	-0.01
75) Chrysene-d12	14.01	240	391736	20.00	ng	0.00
86) Perylene-d12	15.43	264	288932	20.00	ng	-0.01

System Monitoring Compounds

5) 2-Fluorophenol	5.46	112	605467	79.69	ng	0.01
7) Phenol-d6	6.48	99	737894	75.72	ng	-0.01
23) Nitrobenzene-d5	7.42	82	625701	69.44	ng	-0.01
41) 2,4,6-Tribromophenol	10.69	330	187509	79.17	ng	0.00
44) 2-Fluorobiphenyl	9.21	172	1073730	73.13	ng	-0.01
78) Terphenyl-d14	12.96	244	1136183	69.79	ng	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	2.55	88	112967	30.59	ng	99
3) Pyridine	3.27	79	334530	36.88	ng	99
4) n-Nitrosodimethylamine	3.21	42	134464	35.51	ng	99
6) Aniline	6.52	93	219253	16.75	ng	96
8) 2-Chlorophenol	6.64	128	324875	36.96	ng	92
9) Benzaldehyde	6.40	77	56337	9.33	ng	89
10) Phenol	6.49	94	375096	33.79	ng	99
11) bis(2-Chloroethyl)ether	6.58	93	290257	34.69	ng	92
12) 1,3-Dichlorobenzene	6.79	146	321646m	33.58	ng	
13) 1,4-Dichlorobenzene	6.87	146	337504	34.54	ng	98
14) 1,2-Dichlorobenzene	7.03	146	316157	36.45	ng	96
15) Benzyl Alcohol	7.00	79	252553	37.84	ng	100
16) 2,2'-oxybis(1-Chloropropan	7.13	45	379584	34.49	ng	99
17) 2-Methylphenol	7.11	107	264510	36.30	ng	99
18) Hexachloroethane	7.36	117	125494	35.57	ng	# 82
19) n-Nitroso-di-n-propylamine	7.27	70	204892	35.20	ng	94
20) 3+4-Methylphenols	7.27	107	321143	38.05	ng	94
22) Acetophenone	7.27	105	414120	33.36	ng	# 93
24) Nitrobenzene	7.44	77	349667	35.45	ng	# 88
25) Isophorone	7.68	82	631239	34.99	ng	# 94
26) 2-Nitrophenol	7.75	139	167158	34.29	ng	# 84
27) 2,4-Dimethylphenol	7.80	122	313421	36.36	ng	97
28) bis(2-Chloroethoxy)methane	7.89	93	364048	35.47	ng	98
29) 2,4-Dichlorophenol	8.00	162	265699	37.32	ng	94
30) 1,2,4-Trichlorobenzene	8.08	180	277400	34.79	ng	98
31) Naphthalene	8.16	128	920018	34.85	ng	99
32) Benzoic acid	7.91	122	203046	28.17	ng	99
33) 4-Chloroaniline	8.21	127	89901	8.31	ng	# 95
34) Hexachlorobutadiene	8.28	225	141891	33.39	ng	98
35) Caprolactam	8.57	113	85210m	34.79	ng	
36) 4-Chloro-3-methylphenol	8.70	107	300485	36.14	ng	96
37) 2-Methylnaphthalene	8.85	142	615844	38.27	ng	98
39) 1,2,4,5-Tetrachlorobenzene	9.02	216	238478	32.51	ng	98
40) Hexachlorocyclopentadiene	9.01	237	271092	82.01	ng	99

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF031816\
 Data File : BF085552.D
 Acq On : 19 Mar 2016 4:17
 Operator : UM/SJ
 Sample : PB88981BSD
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB88981BSD


Manual Integrations
 APPROVED

apatel
 3/21/2016 5:08:02 PM

Quant Time: Mar 19 04:41:44 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF031816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 18 23:31:39 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
42) 2,4,6-Trichlorophenol	9.13	196	189917	37.83	ng	99
43) 2,4,5-Trichlorophenol	9.18	196	188352	36.86	ng	99
45) 1,1'-Biphenyl	9.32	154	684255	32.80	ng	98
46) 2-Chloronaphthalene	9.34	162	518480	34.03	ng	93
47) 2-Nitroaniline	9.44	65	167967	36.44	ng	# 73
48) Acenaphthylene	9.75	152	877061	35.57	ng	99
49) Dimethylphthalate	9.62	163	658676	36.10	ng	98
50) 2,6-Dinitrotoluene	9.68	165	153291	39.79	ng	# 79
51) Acenaphthene	9.92	154	613572	39.65	ng	100
52) 3-Nitroaniline	9.84	138	67805	14.05	ng	92
53) 2,4-Dinitrophenol	9.96	184	153855	65.20	ng	90
54) Dibenzofuran	10.10	168	808575	42.91	ng	96
55) 4-Nitrophenol	10.01	139	239601	64.26	ng	97
56) 2,4-Dinitrotoluene	10.08	165	173088	34.32	ng	# 54
57) Fluorene	10.45	166	576610	36.72	ng	99
58) 2,3,4,6-Tetrachlorophenol	10.22	232	154985	42.13	ng	97
59) Diethylphthalate	10.32	149	666288	36.73	ng	98
60) 4-Chlorophenyl-phenylether	10.44	204	288869	37.67	ng	90
61) 4-Nitroaniline	10.46	138	144689	30.43	ng	90
62) Azobenzene	10.60	77	707120	40.62	ng	95
64) 4,6-Dinitro-2-methylphenol	10.49	198	109072	38.40	ng	# 71
65) n-Nitrosodiphenylamine	10.55	169	481707	34.84	ng	99
66) 4-Bromophenyl-phenylether	10.93	248	169678	38.27	ng	# 87
67) Hexachlorobenzene	10.98	284	171961	36.51	ng	# 67
68) Atrazine	11.09	200	173226	41.15	ng	94
69) Pentachlorophenol	11.18	266	196638	68.76	ng	98
70) Phenanthrene	11.41	178	893116	38.16	ng	99
71) Anthracene	11.45	178	888742	36.22	ng	100
72) Carbazole	11.61	167	825126	38.98	ng	97
73) Di-n-butylphthalate	11.94	149	1001868	37.79	ng	98
74) Fluoranthene	12.59	202	846281	34.54	ng	97
76) Benzidine	12.71	184	236096	17.92	ng	98
77) Pyrene	12.81	202	873779	31.07	ng	99
79) Butylbenzylphthalate	13.43	149	411516	30.69	ng	# 75
80) Benzo(a)anthracene	14.00	228	799565	35.16	ng	99
81) 3,3'-Dichlorobenzidine	13.97	252	69299	8.32	ng	# 95
82) Chrysene	14.04	228	615899	28.15	ng	100
83) Bis(2-ethylhexyl)phthalate	13.99	149	506825	30.43	ng	# 98
84) Di-n-octyl phthalate	14.61	149	940967	34.67	ng	98
85) Indeno(1,2,3-cd)pyrene	16.84	276	549852	30.08	ng	100
87) Benzo(b)fluoranthene	15.03	252	694857m	36.86	ng	
88) Benzo(k)fluoranthene	15.05	252	582993m	37.53	ng	
89) Benzo(a)pyrene	15.37	252	594974	37.23	ng	97
90) Dibenzo(a,h)anthracene	16.85	278	462652	34.71	ng	99
91) Benzo(g,h,i)perylene	17.25	276	448160	34.73	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 6.86 min Scan# 400
Delta R.T. 0.00 min
Lab File: BF085552.D
Acq: 19 Mar 2016 4:17




Instrument :
BNA_F
ClientSampleId :
PB88981BSD



Tgt Ion	Ratio	Resp	Lower	Upper
152	100	127623		
150	156.6	124.2	186.2	
115	52.5	47.0	70.6	


Manual Integrations
APPROVED

apatel
3/21/2016 5:08:02 PM








#2

1,4-Dioxane
Concen: 30.59 ng
RT: 2.55 min Scan# 23
Delta R.T. 0.09 min
Lab File: BF085552.D
Acq: 19 Mar 2016 4:17



Tgt Ion	Ratio	Resp	Lower	Upper
88	100	112967		
58	68.5	54.2	81.2	
43	24.6	19.3	28.9	









#3
Pyridine
Concen: 36.88 ng
RT: 3.27 min Scan# 86
Delta R.T. 0.08 min
Lab File: BF085552.D
Acq: 19 Mar 2016 4:17

Tgt Ion	Ratio	Lower	Upper
79	100		
52	62.9	49.8	74.6
51	31.3	24.7	37.1

Instrument :
BNA_F
ClientSampleId :
PB88981BSD


Manual Integrations
APPROVED


apatel
3/21/2016 5:08:02 PM



#4
n-Nitrosodimethylamine
Concen: 35.51 ng
RT: 3.21 min Scan# 81
Delta R.T. 0.08 min
Lab File: BF085552.D
Acq: 19 Mar 2016 4:17

Tgt Ion	Ratio	Lower	Upper
42	100		
74	153.3	121.4	182.2
44	6.5	5.6	8.4





#5

2-Fluorophenol

Concen: 79.69 ng

RT: 5.46 min Scan# 278

Delta R.T. 0.01 min

Lab File: BF085552.D


Acq: 19 Mar 2016 4:17

Instrument :

BNA_F

ClientSampleId :




PB88981BSD



Tgt Ion:	112	Resp:	605467
Ion Ratio	Lower	Upper	
112	100		
64	45.0	39.9	59.9
63	22.7	20.2	30.2

Manual Integrations
APPROVED

apatel
3/21/2016 5:08:02 PM



#6

Aniline

Concen: 16.75 ng




RT: 6.52 min Scan# 370




Delta R.T. 0.00 min

Lab File: BF085552.D

Acq: 19 Mar 2016 4:17

Tgt Ion:	93	Resp:	219253
Ion Ratio	Lower	Upper	
93	100		
66	36.5	31.4	47.2
65	18.0	15.7	23.5





#7

Phenol-d6

Concen: 75.72 ng

RT: 6.48 min Scan# 367

Delta R.T. -0.01 min

Lab File: BF085552.D

Acq: 19 Mar 2016 4:17

Tgt Ion:	99	Resp:	737894
Ion Ratio	Lower	Upper	
99	100		
42	15.4	11.1	16.7
71	32.2	24.9	37.3

Instrument :





BNA_F

ClientSampleId :

PB88981BSD

Manual Integrations
APPROVED

apatel
3/21/2016 5:08:02 PM



#8

2-Chlorophenol

Concen: 36.96 ng


RT: 6.64 min Scan# 381


Delta R.T. 0.00 min

Lab File: BF085552.D

Acq: 19 Mar 2016 4:17

Tgt Ion:	128	Resp:	324875
Ion Ratio	Lower	Upper	
128	100		
130	33.6	12.9	52.9
64	32.2	20.2	60.2





#9

Benzaldehyde

Concen: 9.33 ng

RT: 6.40 min Scan# 360

Delta R.T. 0.00 min

Lab File: BF085552.D

Acq: 19 Mar 2016 4:17

Instrument :

BNA_F





ClientSampleId :

PB88981BSD

Tgt Ion:	77	Resp:	56337
Ion Ratio	Lower	Upper	
77	100		
105	110.4	80.1	120.1
106	106.0	75.0	115.0

Manual Integrations
APPROVED

apatel
3/21/2016 5:08:02 PM



#10

Phenol

Concen: 33.79 ng




RT: 6.49 min Scan# 368


Delta R.T. -0.01 min

Lab File: BF085552.D

Acq: 19 Mar 2016 4:17

Tgt Ion:	94	Resp:	375096
Ion Ratio	Lower	Upper	
94	100		
65	29.1	8.6	48.6
66	40.8	20.0	60.0









#11
bis(2-Chloroethyl)ether
Concen: 34.69 ng
RT: 6.58 min Scan# 376
Delta R.T. -0.01 min
Lab File: BF085552.D
Acq: 19 Mar 2016 4:17

Instrument :
BNA_F
ClientSampleId :
PB88981BSD

Tgt Ion	Ratio	Resp Lower	Resp Upper
93	100		
63	79.4	50.9	90.9
95	32.1	13.3	53.3




Manual Integrations
APPROVED


apatel
3/21/2016 5:08:02 PM



#12
1,3-Dichlorobenzene
Concen: 33.58 ng m
RT: 6.79 min Scan# 394
Delta R.T. -0.01 min
Lab File: BF085552.D
Acq: 19 Mar 2016 4:17

Tgt Ion	Ratio	Resp Lower	Resp Upper
146	100		
148	63.5	52.3	78.5
75	35.6	17.4	26.2









#13
 1,4-Dichlorobenzene
 Concen: 34.54 ng
 RT: 6.87 min Scan# 401
 Delta R.T. 0.00 min
 Lab File: BF085552.D
 Acq: 19 Mar 2016 4:17

Instrument :
 BNA_F
 ClientSampleId :
 PB88981BSD

Tgt Ion	Ratio	Lower	Upper
146	100		
148	63.8	50.6	76.0
111	44.3	37.8	56.6




Manual Integrations
 APPROVED


apatel
 3/21/2016 5:08:02 PM



#14
 1,2-Dichlorobenzene
 Concen: 36.45 ng
 RT: 7.03 min Scan# 415
 Delta R.T. 0.00 min
 Lab File: BF085552.D
 Acq: 19 Mar 2016 4:17

Tgt Ion	Ratio	Lower	Upper
146	100		
148	65.1	51.4	77.2
111	37.8	34.6	52.0





#15

Benzyl Alcohol

Concen: 37.84 ng

RT: 7.00 min Scan# 412

Delta R.T. 0.00 min

Lab File: BF085552.D

Acq: 19 Mar 2016 4:17

Instrument :

BNA_F





ClientSampleId :

PB88981BSD

Tgt Ion:	79	Resp:	252553
Ion Ratio	Lower	Upper	
79	100		
108	77.2	61.8	92.6
77	61.8	49.6	74.4

Manual Integrations
APPROVED

apatel
3/21/2016 5:08:02 PM



#16

2,2'-oxybis(1-Chloropropane)

Concen: 34.49 ng




RT: 7.13 min Scan# 424


Delta R.T. 0.00 min

Lab File: BF085552.D

Acq: 19 Mar 2016 4:17

Tgt Ion:	45	Resp:	379584
Ion Ratio	Lower	Upper	
45	100		
77	15.9	0.0	35.7
79	12.5	0.0	32.2






#17
2-Methylphenol
Concen: 36.30 ng
RT: 7.11 min Scan# 422
Delta R.T. 0.00 min
Lab File: BF085552.D
Acq: 19 Mar 2016 4:17

Instrument :
BNA_F
ClientSampleId :
PB88981BSD




Tgt Ion	Ratio	Resp Lower	Resp Upper
107	100		
108	113.8	91.4	137.0
77	44.6	36.2	54.2
79	43.6	35.8	53.8

Manual Integrations
APPROVED

apatel
3/21/2016 5:08:02 PM




Abundance Ion 107.00 (106.70 to 107.70): E
Ion 108.00 (107.70 to 108.70): E
Ion 77.00 (76.70 to 77.70): BF0
Ion 79.00 (78.70 to 79.70): BF0











#18
Hexachloroethane
Concen: 35.57 ng
RT: 7.36 min Scan# 444
Delta R.T. -0.01 min
Lab File: BF085552.D
Acq: 19 Mar 2016 4:17

Tgt Ion	Ratio	Resp Lower	Resp Upper
117	100		
119	95.9	76.7	115.1
201	69.0	83.3	124.9#




Abundance Ion 117.00 (116.70 to 117.70): E
Ion 119.00 (118.70 to 119.70): E
Ion 201.00 (200.70 to 201.70): E





#19
n-Nitroso-di-n-propylamine
Concen: 35.20 ng
RT: 7.27 min Scan# 436
Delta R.T. -0.01 min
Lab File: BF085552.D
Acq: 19 Mar 2016 4:17

Tgt Ion	Ratio	Lower	Upper
70	100		
42	49.9	37.1	55.7
101	9.7	8.6	12.8
130	20.5	19.6	29.4




Instrument :
BNA_F
ClientSampleId :
PB88981BSD


Manual Integrations
APPROVED

apatel
3/21/2016 5:08:02 PM

#20
3+4-Methylphenols
Concen: 38.05 ng
RT: 7.27 min Scan# 436
Delta R.T. 0.00 min
Lab File: BF085552.D
Acq: 19 Mar 2016 4:17

Tgt Ion	Ratio	Lower	Upper
107	100		
108	93.3	69.3	109.3
77	111.9	101.7	141.7
79	29.1	8.3	48.3









#21
Naphthalene-d8
Concen: 20.00 ng
RT: 8.14 min Scan# 512
Delta R.T. 0.00 min
Lab File: BF085552.D
Acq: 19 Mar 2016 4:17

Instrument :
BNA_F
ClientSampleId :
PB88981BSD

Tgt	Ion	Ratio	Resp	Lower	Upper
136	100		523071		
137	11.1	9.1	13.7		
54	7.1	7.4	11.0#		
68	6.3	5.8	8.8		




Manual Integrations
APPROVED


apatel
3/21/2016 5:08:02 PM



#22
Acetophenone
Concen: 33.36 ng
RT: 7.27 min Scan# 436
Delta R.T. 0.00 min
Lab File: BF085552.D
Acq: 19 Mar 2016 4:17

Tgt	Ion	Ratio	Resp	Lower	Upper
105	100		414120		
71	10.5	3.6	5.4#		
51	27.1	25.4	38.0		
120	22.2	16.9	25.3		





#23

Nitrobenzene-d5

Concen: 69.44 ng

RT: 7.42 min Scan# 449

Delta R.T. -0.01 min

Lab File: BF085552.D

Acq: 19 Mar 2016 4:17

Instrument :

BNA_F





ClientSampleId :

PB88981BSD

Tgt Ion:	82	Resp:	625701
Ion Ratio	Lower	Upper	
82	100		
128	45.8	41.8	62.8
54	45.0	33.4	50.0

Manual Integrations
APPROVED

apatel
3/21/2016 5:08:02 PM



#24

Nitrobenzene

Concen: 35.45 ng




RT: 7.44 min Scan# 451


Delta R.T. 0.00 min

Lab File: BF085552.D

Acq: 19 Mar 2016 4:17

Tgt Ion:	77	Resp:	349667
Ion Ratio	Lower	Upper	
77	100		
123	53.5	35.0	52.4#
65	12.2	10.6	16.0





#25

Isophorone

Concen: 34.99 ng

RT: 7.68 min Scan# 472

Delta R.T. 0.00 min

Lab File: BF085552.D

Acq: 19 Mar 2016 4:17

Instrument :

BNA_F





ClientSampleId :

PB88981BSD

Tgt Ion	Ratio	Resp	Lower	Upper
82	100	631239		
95	7.4	5.4	8.2	
138	19.7	12.9	19.3#	

Manual Integrations
APPROVED

apatel
3/21/2016 5:08:02 PM



#26

2-Nitrophenol

Concen: 34.29 ng




RT: 7.75 min Scan# 478




Delta R.T. -0.01 min

Lab File: BF085552.D

Acq: 19 Mar 2016 4:17

Tgt Ion	Ratio	Resp	Lower	Upper
139	100	167158		
109	30.8	22.1	33.1	
65	56.9	33.9	50.9#	









#27
2,4-Dimethylphenol
Concen: 36.36 ng
RT: 7.80 min Scan# 482
Delta R.T. 0.00 min
Lab File: BF085552.D
Acq: 19 Mar 2016 4:17

Tgt Ion	Ratio	Resp	Lower	Upper
122	100	313421		
107	111.7	93.0	139.6	
121	58.4	45.9	68.9	

Instrument :
BNA_F
ClientSampleID :
PB88981BSD


Manual Integrations
APPROVED


apatel
3/21/2016 5:08:02 PM



#28
bis(2-Chloroethoxy)methane
Concen: 35.47 ng
RT: 7.89 min Scan# 490
Delta R.T. -0.01 min
Lab File: BF085552.D
Acq: 19 Mar 2016 4:17

Tgt Ion	Ratio	Resp	Lower	Upper
93	100	364048		
95	32.1	26.4	39.6	
123	10.8	9.8	14.6	









#29
2,4-Dichlorophenol
Concen: 37.32 ng
RT: 8.00 min Scan# 500
Delta R.T. 0.00 min
Lab File: BF085552.D
Acq: 19 Mar 2016 4:17

Instrument :
BNA_F
ClientSampleId :
PB88981BSD

Tgt Ion	Ratio	Resp	Lower	Upper
162	100	265699		
164	64.9	43.7	83.7	
98	32.7	20.4	60.4	




Manual Integrations
APPROVED


apatel
3/21/2016 5:08:02 PM



#30
1,2,4-Trichlorobenzene
Concen: 34.79 ng
RT: 8.08 min Scan# 507
Delta R.T. 0.00 min
Lab File: BF085552.D
Acq: 19 Mar 2016 4:17

Tgt Ion	Ratio	Resp	Lower	Upper
180	100	277400		
182	93.7	73.8	110.8	
145	32.8	28.4	42.6	









#31
Naphthalene
Concen: 34.85 ng
RT: 8.16 min Scan# 514
Delta R.T. 0.00 min
Lab File: BF085552.D
Acq: 19 Mar 2016 4:17

Instrument :
BNA_F
ClientSampleId :
PB88981BSD

Tgt Ion	Ratio	Resp Lower	Resp Upper
128	100		
129	11.4	9.4	14.0
127	13.3	11.0	16.6




Manual Integrations
APPROVED


apatel
3/21/2016 5:08:02 PM



#32
Benzoic acid
Concen: 28.17 ng
RT: 7.91 min Scan# 492
Delta R.T. -0.02 min
Lab File: BF085552.D
Acq: 19 Mar 2016 4:17

Tgt Ion	Ratio	Resp Lower	Resp Upper
122	100		
105	121.3	102.9	142.9
77	91.6	71.8	111.8









#33
4-Chloroaniline
Concen: 8.31 ng
RT: 8.21 min Scan# 518
Delta R.T. -0.01 min
Lab File: BF085552.D
Acq: 19 Mar 2016 4:17

Instrument :
BNA_F
ClientSampleId :
PB88981BSD

Tgt Ion	Ratio	Resp Lower	Resp Upper
127	100		
129	31.9	26.1	39.1
65	30.5	19.5	29.3
92	19.9	14.6	22.0




Manual Integrations
APPROVED


apatel
3/21/2016 5:08:02 PM



#34
Hexachlorobutadiene
Concen: 33.39 ng
RT: 8.28 min Scan# 524
Delta R.T. -0.01 min
Lab File: BF085552.D
Acq: 19 Mar 2016 4:17

Tgt Ion	Ratio	Resp Lower	Resp Upper
225	100		
223	63.1	50.4	75.6
227	61.3	51.4	77.2









#35
 Caprolactam
 Concen: 34.79 ng m
 RT: 8.57 min Scan# 550
 Delta R.T. -0.02 min
 Lab File: BF085552.D
 Acq: 19 Mar 2016 4:17

Instrument :
 BNA_F
 ClientSampleId :
 PB88981BSD

Tgt Ion:113 Resp: 85210
 Ion Ratio Lower Upper
 113 100
 55 169.4 139.4 179.4
 56 126.4 100.4 140.4




Manual Integrations
 APPROVED


apatel
 3/21/2016 5:08:02 PM



#36
 4-Chloro-3-methylphenol
 Concen: 36.14 ng
 RT: 8.70 min Scan# 561
 Delta R.T. 0.00 min
 Lab File: BF085552.D
 Acq: 19 Mar 2016 4:17

Tgt Ion:107 Resp: 300485
 Ion Ratio Lower Upper
 107 100
 144 26.2 19.3 28.9
 142 80.5 61.4 92.0









#37
 2-Methylnaphthalene
 Concen: 38.27 ng
 RT: 8.85 min Scan# 574
 Delta R.T. -0.01 min
 Lab File: BF085552.D
 Acq: 19 Mar 2016 4:17

Instrument :
 BNA_F
 ClientSampleId :
 PB88981BSD

Tgt Ion	Ratio	Lower	Upper
142	100		
141	89.8	71.6	107.4
115	37.0	26.8	40.2




Manual Integrations
 APPROVED


apatel
 3/21/2016 5:08:02 PM



#38
 Acenaphthene-d10
 Concen: 20.00 ng
 RT: 9.90 min Scan# 666
 Delta R.T. 0.00 min
 Lab File: BF085552.D
 Acq: 19 Mar 2016 4:17

Tgt Ion	Ratio	Lower	Upper
164	100		
162	103.6	82.1	123.1
160	47.0	37.4	56.2









#39
 1,2,4,5-Tetrachlorobenzene
 Concen: 32.51 ng
 RT: 9.02 min Scan# 589
 Delta R.T. 0.00 min
 Lab File: BF085552.D
 Acq: 19 Mar 2016 4:17

Instrument :
 BNA_F
 ClientSampleId :
 PB88981BSD

Tgt Ion	Ratio	Resp Lower	Upper
216	100		
214	79.5	63.1	94.7
179	23.9	18.2	27.2
108	24.8	17.3	25.9




Manual Integrations
 APPROVED


apatel
 3/21/2016 5:08:02 PM



#40
 Hexachlorocyclopentadiene
 Concen: 82.01 ng
 RT: 9.01 min Scan# 588
 Delta R.T. 0.00 min
 Lab File: BF085552.D
 Acq: 19 Mar 2016 4:17

Tgt Ion	Ratio	Resp Lower	Upper
237	100		
235	63.0	43.7	83.7
272	12.7	0.0	31.9






#41
 2,4,6-Tribromophenol
 Concen: 79.17 ng
 RT: 10.69 min Scan# 735
 Delta R.T. 0.00 min
 Lab File: BF085552.D
 Acq: 19 Mar 2016 4:17

Instrument :
 BNA_F
 ClientSampleId :
 PB88981BSD




Tgt Ion	Ratio	Resp	Lower	Upper
330	100	187509		
332	97.1	78.2	117.2	
141	35.4	31.5	47.3	

Manual Integrations
 APPROVED

apatel
 3/21/2016 5:08:02 PM






Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E




#42
 2,4,6-Trichlorophenol
 Concen: 37.83 ng
 RT: 9.13 min Scan# 599
 Delta R.T. 0.00 min
 Lab File: BF085552.D
 Acq: 19 Mar 2016 4:17

Tgt Ion	Ratio	Resp	Lower	Upper
196	100	189917		
198	95.7	75.3	112.9	
200	30.0	23.7	35.5	


Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 200.00 (199.70 to 200.70): E





#43
 2,4,5-Trichlorophenol
 Concen: 36.86 ng
 RT: 9.18 min Scan# 603
 Delta R.T. 0.00 min
 Lab File: BF085552.D
 Acq: 19 Mar 2016 4:17




Instrument :
 BNA_F
 ClientSampleId :
 PB88981BSD



Tgt Ion	Ratio	Resp	Lower	Upper
196	100	188352		
198	94.1	76.6	114.8	
97	37.7	30.5	45.7	
132	24.4	20.2	30.2	




Manual Integrations
 APPROVED


apatel
 3/21/2016 5:08:02 PM



#44
 2-Fluorobiphenyl
 Concen: 73.13 ng
 RT: 9.21 min Scan# 606
 Delta R.T. -0.01 min
 Lab File: BF085552.D
 Acq: 19 Mar 2016 4:17

Tgt Ion	Ratio	Resp	Lower	Upper
172	100	1073730		
171	36.3	29.4	44.2	
170	24.4	19.8	29.6	





#45

1,1'-Biphenyl

Concen: 32.80 ng

RT: 9.32 min Scan# 615

Delta R.T. 0.00 min

Lab File: BF085552.D

Acq: 19 Mar 2016 4:17

Instrument :

BNA_F





ClientSampleId :

PB88981BSD

Tgt Ion:	154	Resp:	684255
Ion Ratio	Lower	Upper	
154	100		
153	40.8	20.8	60.8
76	14.1	0.0	36.7

Manual Integrations
APPROVED

apatel
3/21/2016 5:08:02 PM



#46

2-Chloronaphthalene

Concen: 34.03 ng




RT: 9.34 min Scan# 617




Delta R.T. -0.01 min

Lab File: BF085552.D

Acq: 19 Mar 2016 4:17

Tgt Ion:	162	Resp:	518480
Ion Ratio	Lower	Upper	
162	100		
127	46.9	31.7	47.5
164	32.8	27.3	40.9









#47
2-Nitroaniline
Concen: 36.44 ng
RT: 9.44 min Scan# 626
Delta R.T. 0.00 min
Lab File: BF085552.D
Acq: 19 Mar 2016 4:17

Instrument :
BNA_F
ClientSampleId :
PB88981BSD

Tgt Ion	Ratio	Lower	Upper
65	100		
92	84.3	59.7	89.5
138	155.3	91.4	137.0


Manual Integrations
APPROVED


apatel
3/21/2016 5:08:02 PM



#48
Acenaphthylene
Concen: 35.57 ng
RT: 9.75 min Scan# 653
Delta R.T. -0.01 min
Lab File: BF085552.D
Acq: 19 Mar 2016 4:17

Tgt Ion	Ratio	Lower	Upper
152	100		
151	20.4	16.8	25.2
153	13.8	10.9	16.3






#49
Dimethylphthalate
Concen: 36.10 ng
RT: 9.62 min Scan# 642
Delta R.T. 0.00 min
Lab File: BF085552.D
Acq: 19 Mar 2016 4:17

Instrument :
BNA_F
ClientSampleId :
PB88981BSD




Tgt Ion	Ratio	Resp Lower	Resp Upper
163	100		
194	3.9	2.9	4.3
164	9.7	8.3	12.5

Manual Integrations
APPROVED

apatel
3/21/2016 5:08:02 PM






Abundance Ion 163.00 (162.70 to 163.70): E
Ion 194.00 (193.70 to 194.70): E
Ion 164.00 (163.70 to 164.70): E




#50
2,6-Dinitrotoluene
Concen: 39.79 ng
RT: 9.68 min Scan# 647
Delta R.T. 0.00 min
Lab File: BF085552.D
Acq: 19 Mar 2016 4:17

Tgt Ion	Ratio	Resp Lower	Resp Upper
165	100		
63	45.6	51.8	77.8#
89	53.1	53.7	80.5#

Abundance Ion 165.00 (164.70 to 165.70): E
Ion 63.00 (62.70 to 63.70): BF0
Ion 89.00 (88.70 to 89.70): BF0





#51

Acenaphthene

Concen: 39.65 ng

RT: 9.92 min Scan# 668

Delta R.T. -0.01 min

Lab File: BF085552.D


Acq: 19 Mar 2016 4:17

Instrument :

BNA_F

ClientSampleId :




PB88981BSD



Tgt Ion:	154	Resp:	613572
Ion Ratio	Lower	Upper	
154	100		
153	113.1	90.1	135.1
152	55.2	44.2	66.2

Manual Integrations
APPROVED

apatel
3/21/2016 5:08:02 PM



#52

3-Nitroaniline


Concen: 14.05 ng

RT: 9.84 min Scan# 661



Delta R.T. -0.01 min


Lab File: BF085552.D

Acq: 19 Mar 2016 4:17



Tgt Ion:	138	Resp:	67805
Ion Ratio	Lower	Upper	
138	100		
108	11.6	8.7	13.1
92	125.7	93.3	139.9









#53
 2,4-Dinitrophenol
 Concen: 65.20 ng
 RT: 9.96 min Scan# 671
 Delta R.T. 0.00 min
 Lab File: BF085552.D
 Acq: 19 Mar 2016 4:17

Instrument :
 BNA_F
 ClientSampleId :
 PB88981BSD

Tgt Ion	Ratio	Resp	Lower	Upper
184	100	153855		
63	73.9	69.5	104.3	
154	69.0	58.3	87.5	




Manual Integrations
 APPROVED


apatel
 3/21/2016 5:08:02 PM



#54
 Dibenzofuran
 Concen: 42.91 ng
 RT: 10.10 min Scan# 684
 Delta R.T. 0.00 min
 Lab File: BF085552.D
 Acq: 19 Mar 2016 4:17

Tgt Ion	Ratio	Resp	Lower	Upper
168	100	808575		
139	37.3	31.9	47.9	
169	12.9	11.0	16.6	









#55
4-Nitrophenol
Concen: 64.26 ng
RT: 10.01 min Scan# 676
Delta R.T. 0.00 min
Lab File: BF085552.D
Acq: 19 Mar 2016 4:17

Instrument :
BNA_F
ClientSampleId :
PB88981BSD

Tgt Ion	Ratio	Resp	Lower	Upper
139	100	239601		
109	70.2	49.2	89.2	
65	91.1	66.9	106.9	




Manual Integrations
APPROVED


apatel
3/21/2016 5:08:02 PM



#56
2,4-Dinitrotoluene
Concen: 34.32 ng
RT: 10.08 min Scan# 682
Delta R.T. -0.01 min
Lab File: BF085552.D
Acq: 19 Mar 2016 4:17

Tgt Ion	Ratio	Resp	Lower	Upper
165	100	173088		
63	51.6	24.9	37.3#	
89	88.5	41.0	61.6#	
182	2.3	2.1	3.1	





#57

Fluorene

Concen: 36.72 ng

RT: 10.45 min Scan# 714

Delta R.T. 0.00 min

Lab File: BF085552.D


Acq: 19 Mar 2016 4:17

Instrument :

BNA_F

ClientSampleId :




PB88981BSD



Tgt Ion	Ratio	Resp	Lower	Upper
166	100	576610		
165	98.4	79.4	119.0	
167	12.6	11.0	16.4	

Manual Integrations
APPROVED

apatel
3/21/2016 5:08:02 PM



#58

2,3,4,6-Tetrachlorophenol

Concen: 42.13 ng




RT: 10.22 min Scan# 694


Delta R.T. 0.00 min

Lab File: BF085552.D

Acq: 19 Mar 2016 4:17

Tgt Ion	Ratio	Resp	Lower	Upper
232	100	154985		
131	53.4	44.9	67.3	
130	2.7	2.3	3.5	
166	33.5	28.1	42.1	









#59
Diethylphthalate
Concen: 36.73 ng
RT: 10.32 min Scan# 703
Delta R.T. 0.00 min
Lab File: BF085552.D
Acq: 19 Mar 2016 4:17

Instrument :
BNA_F
ClientSampleId :
PB88981BSD

Tgt Ion	Ratio	Resp	Lower	Upper
149	100	666288		
177	23.7	17.9	26.9	
150	12.6	10.2	15.2	




Manual Integrations
APPROVED


apatel
3/21/2016 5:08:02 PM



#60
4-Chlorophenyl-phenylether
Concen: 37.67 ng
RT: 10.44 min Scan# 713
Delta R.T. 0.00 min
Lab File: BF085552.D
Acq: 19 Mar 2016 4:17

Tgt Ion	Ratio	Resp	Lower	Upper
204	100	288869		
206	33.1	26.6	39.8	
141	59.6	57.4	86.0	









#61
4-Nitroaniline
Concen: 30.43 ng
RT: 10.46 min Scan# 715
Delta R.T. -0.01 min
Lab File: BF085552.D
Acq: 19 Mar 2016 4:17

Instrument :
BNA_F
ClientSampleId :
PB88981BSD

Tgt Ion	Ratio	Resp	Lower	Upper
138	100	144689		
92	54.0	26.4	66.4	
108	76.7	49.0	89.0	




Manual Integrations
APPROVED


apatel
3/21/2016 5:08:02 PM



#62
Azobenzene
Concen: 40.62 ng
RT: 10.60 min Scan# 727
Delta R.T. 0.00 min
Lab File: BF085552.D
Acq: 19 Mar 2016 4:17

Tgt Ion	Ratio	Resp	Lower	Upper
77	100	707120		
182	27.0	1.9	41.9	
105	24.5	3.5	43.5	
51	25.1	6.3	46.3	





#63

Phenanthrene-d10

Concen: 20.00 ng

RT: 11.37 min Scan# 795

Delta R.T. -0.01 min

Lab File: BF085552.D

Acq: 19 Mar 2016 4:17

Instrument :

BNA_F

ClientSampleId :

PB88981BSD

Tgt Ion:188 Resp: 443843
Ion Ratio Lower Upper


188 100

94 12.3 5.4 8.0#




80 13.4 5.5 8.3#

Manual Integrations
APPROVED

apatel
3/21/2016 5:08:02 PM



Abundance Ion 188.00 (187.70 to 188.70): E
Ion 94.00 (93.70 to 94.70): BF0
Ion 80.00 (79.70 to 80.70): BF0



#64

4,6-Dinitro-2-methylphenol

Concen: 38.40 ng

RT: 10.49 min Scan# 718

Delta R.T. 0.00 min

Lab File: BF085552.D


Acq: 19 Mar 2016 4:17

Tgt Ion:198 Resp: 109072
Ion Ratio Lower Upper



198 100


51 38.2 45.4 85.4#

105 46.8 45.9 85.9



Abundance Ion 198.00 (197.70 to 198.70): E
Ion 51.00 (50.70 to 51.70): BF0
Ion 105.00 (104.70 to 105.70): E





#65

n-Nitrosodiphenylamine

Concen: 34.84 ng

RT: 10.55 min Scan# 723

Delta R.T. -0.01 min

Lab File: BF085552.D

Acq: 19 Mar 2016 4:17

Instrument :

BNA_F


ClientSampleId :

PB88981BSD




Tgt Ion:	169	Resp:	481707
Ion Ratio	Lower	Upper	
169	100		
168	66.7	54.4	81.6
167	36.8	29.4	44.0

Manual Integrations
APPROVED

apatel
3/21/2016 5:08:02 PM



Abundance Ion 169.00 (168.70 to 169.70): E
Ion 168.00 (167.70 to 168.70): E
Ion 167.00 (166.70 to 167.70): E



#66

4-Bromophenyl-phenylether

Concen: 38.27 ng

RT: 10.93 min Scan# 756




Delta R.T. 0.00 min


Lab File: BF085552.D

Acq: 19 Mar 2016 4:17

Tgt Ion:	248	Resp:	169678
Ion Ratio	Lower	Upper	
248	100		
250	97.9	77.4	116.0
141	55.7	63.6	95.4#

Abundance Ion 248.00 (247.70 to 248.70): E
Ion 250.00 (249.70 to 250.70): E
Ion 141.00 (140.70 to 141.70): E





#67

Hexachlorobenzene

Concen: 36.51 ng

RT: 10.98 min Scan# 761

Delta R.T. -0.01 min

Lab File: BF085552.D


Acq: 19 Mar 2016 4:17

Instrument :

BNA_F

ClientSampleId :




PB88981BSD



Tgt Ion:	284	Resp:	171961
Ion Ratio	Lower	Upper	
284	100		
142	59.9	23.1	34.7#
249	35.5	24.2	36.2

Manual Integrations
APPROVED

apatel
3/21/2016 5:08:02 PM



#68

Atrazine

Concen: 41.15 ng




RT: 11.09 min Scan# 770

Delta R.T. 0.00 min

Lab File: BF085552.D

Acq: 19 Mar 2016 4:17

Tgt Ion:	200	Resp:	173226
Ion Ratio	Lower	Upper	
200	100		
173	22.5	7.7	47.7
215	48.1	25.4	65.4






#69
 Pentachlorophenol
 Concen: 68.76 ng
 RT: 11.18 min Scan# 778
 Delta R.T. -0.01 min
 Lab File: BF085552.D
 Acq: 19 Mar 2016 4:17

Instrument :
 BNA_F
 ClientSampleId :
 PB88981BSD




Tgt Ion	Ratio	Resp	Lower	Upper
266	100	196638		
268	61.6	51.5	77.3	
264	64.2	50.6	76.0	

Manual Integrations
 APPROVED

apatel
 3/21/2016 5:08:02 PM






Abundance Ion 265.70 (265.40 to 266.40): E
 Ion 268.00 (267.70 to 268.70): E
 Ion 264.00 (263.70 to 264.70): E




#70
 Phenanthrene
 Concen: 38.16 ng
 RT: 11.41 min Scan# 798
 Delta R.T. 0.00 min
 Lab File: BF085552.D
 Acq: 19 Mar 2016 4:17

Tgt Ion	Ratio	Resp	Lower	Upper
178	100	893116		
176	19.7	16.3	24.5	
179	15.4	12.6	18.8	

Abundance Ion 178.00 (177.70 to 178.70): E
 Ion 176.00 (175.70 to 176.70): E
 Ion 179.00 (178.70 to 179.70): E





#71

Anthracene

Concen: 36.22 ng

RT: 11.45 min Scan# 802

Delta R.T. 0.00 min

Lab File: BF085552.D

Acq: 19 Mar 2016 4:17

Instrument :

BNA_F


ClientSampleId :

PB88981BSD

Tgt Ion	Ratio	Lower	Upper
178	100		
176	19.7	15.9	23.9
179	16.1	13.0	19.4

Manual Integrations
APPROVED




apatel
3/21/2016 5:08:02 PM



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#72

Carbazole

Concen: 38.98 ng

RT: 11.61 min Scan# 816

Delta R.T. 0.00 min

Lab File: BF085552.D




Acq: 19 Mar 2016 4:17


Tgt Ion	Ratio	Lower	Upper
167	100		
166	21.5	17.8	26.8
139	12.2	11.6	17.4

Abundance Ion 167.00 (166.70 to 167.70): E

Ion 166.00 (165.70 to 166.70): E

Ion 139.00 (138.70 to 139.70): E





#73

Di-n-butylphthalate

Concen: 37.79 ng

RT: 11.94 min Scan# 845

Delta R.T. 0.00 min

Lab File: BF085552.D

Acq: 19 Mar 2016 4:17

Instrument :

BNA_F

ClientSampleId :

PB88981BSD

Tgt Ion:149 Resp: 1001868

Ion Ratio Lower Upper

149 100


150 9.0 7.6 11.4

104 4.0 3.8 5.8

Manual Integrations
APPROVED

apatel




3/21/2016 5:08:02 PM



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E



#74

Fluoranthene

Concen: 34.54 ng

RT: 12.59 min Scan# 901

Delta R.T. 0.00 min

Lab File: BF085552.D

Acq: 19 Mar 2016 4:17


Tgt Ion:202 Resp: 846281

Ion Ratio Lower Upper

202 100

101 14.4 0.0 36.6



203 17.8 0.0 38.1




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E





#75

Chrysene-d12

Concen: 20.00 ng

RT: 14.01 min Scan# 1026

Delta R.T. 0.00 min

Lab File: BF085552.D

Acq: 19 Mar 2016 4:17

Instrument :

BNA_F





ClientSampleId :

PB88981BSD

Tgt Ion	Ratio	Lower	Upper
240	100		
120	10.7	13.8	20.8#
236	26.5	20.6	30.8

Manual Integrations
APPROVED

apatel
3/21/2016 5:08:02 PM



#76

Benzidine

Concen: 17.92 ng




RT: 12.71 min Scan# 912


Delta R.T. 0.00 min

Lab File: BF085552.D

Acq: 19 Mar 2016 4:17

Tgt Ion	Ratio	Lower	Upper
184	100		
185	14.5	12.4	18.6
183	11.2	9.4	14.2









#77
 Pyrene
 Concen: 31.07 ng
 RT: 12.81 min Scan# 921
 Delta R.T. -0.01 min
 Lab File: BF085552.D
 Acq: 19 Mar 2016 4:17

Instrument :
 BNA_F
 ClientSampleId :
 PB88981BSD

Tgt Ion	Ratio	Resp	Lower	Upper
202	100	873779		
200	22.0		17.6	26.4
203	18.0		14.0	21.0




Manual Integrations
 APPROVED


apatel
 3/21/2016 5:08:02 PM



#78
 Terphenyl-d14
 Concen: 69.79 ng
 RT: 12.96 min Scan# 934
 Delta R.T. 0.00 min
 Lab File: BF085552.D
 Acq: 19 Mar 2016 4:17

Tgt Ion	Ratio	Resp	Lower	Upper
244	100	1136183		
212	7.2		6.1	9.1
122	10.6		10.2	15.4









#79
Butylbenzylphthalate
Concen: 30.69 ng
RT: 13.43 min Scan# 975
Delta R.T. -0.01 min
Lab File: BF085552.D
Acq: 19 Mar 2016 4:17

Instrument :
BNA_F
ClientSampleId :
PB88981BSD

Tgt Ion	Ratio	Resp Lower	Resp Upper
149	100		
91	79.3	45.8	68.8#
206	15.8	17.8	26.8#




Manual Integrations
APPROVED


apatel
3/21/2016 5:08:02 PM



#80
Benzo(a)anthracene
Concen: 35.16 ng
RT: 14.00 min Scan# 1025
Delta R.T. 0.00 min
Lab File: BF085552.D
Acq: 19 Mar 2016 4:17

Tgt Ion	Ratio	Resp Lower	Resp Upper
228	100		
226	28.0	22.1	33.1
229	20.0	16.2	24.4






#81
 3,3'-Dichlorobenzidine
 Concen: 8.32 ng
 RT: 13.97 min Scan# 1022
 Delta R.T. 0.00 min
 Lab File: BF085552.D
 Acq: 19 Mar 2016 4:17

Instrument :
 BNA_F
 ClientSampleId :
 PB88981BSD




Tgt Ion	Ratio	Resp Lower	Resp Upper
252	100		
254	65.3	50.7	76.1
126	7.6	11.8	17.6

Manual Integrations
 APPROVED

apatel
 3/21/2016 5:08:02 PM




Abundance Ion 252.00 (251.70 to 252.70): E
 Ion 254.00 (253.70 to 254.70): E
 Ion 126.00 (125.70 to 126.70): E






#82
 Chrysene
 Concen: 28.15 ng
 RT: 14.04 min Scan# 1028
 Delta R.T. -0.01 min
 Lab File: BF085552.D
 Acq: 19 Mar 2016 4:17

Tgt Ion	Ratio	Resp Lower	Resp Upper
228	100		
226	31.3	25.0	37.6
229	20.0	15.7	23.5



Abundance Ion 228.00 (227.70 to 228.70): E
 Ion 226.00 (225.70 to 226.70): E
 Ion 229.00 (228.70 to 229.70): E









#83
 Bis(2-ethylhexyl)phthalate
 Concen: 30.43 ng
 RT: 13.99 min Scan# 1024
 Delta R.T. 0.00 min
 Lab File: BF085552.D
 Acq: 19 Mar 2016 4:17

Instrument :
 BNA_F
 ClientSampleId :
 PB88981BSD

Tgt Ion	Ratio	Resp	Lower	Upper
149	100	506825		
167	25.3	19.4	29.2	
279	2.4	1.5	2.3#	




Manual Integrations
 APPROVED


apatel
 3/21/2016 5:08:02 PM



#84
 Di-n-octyl phthalate
 Concen: 34.67 ng
 RT: 14.61 min Scan# 1078
 Delta R.T. 0.00 min
 Lab File: BF085552.D
 Acq: 19 Mar 2016 4:17

Tgt Ion	Ratio	Resp	Lower	Upper
149	100	940967		
167	1.5	1.2	1.8	
43	8.1	7.1	10.7	









#85
 Indeno(1,2,3-cd)pyrene
 Concen: 30.08 ng
 RT: 16.84 min Scan# 1273
 Delta R.T. -0.01 min
 Lab File: BF085552.D
 Acq: 19 Mar 2016 4:17

Instrument :
 BNA_F
 ClientSampleId :
 PB88981BSD

Tgt Ion	Ratio	Resp	Lower	Upper
276	100	549852		
138	30.1		24.2	36.4
277	24.9		20.0	30.0




Manual Integrations
 APPROVED


apatel
 3/21/2016 5:08:02 PM



#86
 Perylene-d12
 Concen: 20.00 ng
 RT: 15.43 min Scan# 1150
 Delta R.T. -0.01 min
 Lab File: BF085552.D
 Acq: 19 Mar 2016 4:17

Tgt Ion	Ratio	Resp	Lower	Upper
264	100	288932		
260	24.7		19.4	29.2
265	21.6		17.2	25.8





#87

Benzo(b)fluoranthene

Concen: 36.86 ng m

RT: 15.03 min Scan# 1115

Delta R.T. 0.00 min

Lab File: BF085552.D

Acq: 19 Mar 2016 4:17

Instrument :

BNA_F

ClientSampleId :

PB88981BSD

Tgt Ion: 252 Resp: 694857
Ion Ratio Lower Upper

252 100


253 21.8 17.6 26.4

125 9.7 10.3 15.5#




Manual Integrations
APPROVED

apatel

3/21/2016 5:08:02 PM



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E



#88

Benzo(k)fluoranthene

Concen: 37.53 ng m

RT: 15.05 min Scan# 1117

Delta R.T. -0.01 min

Lab File: BF085552.D


Acq: 19 Mar 2016 4:17

Tgt Ion: 252 Resp: 582993
Ion Ratio Lower Upper



252 100


253 22.5 17.8 26.6

125 14.4 7.7 11.5#



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E









#89
Benzo(a)pyrene
Concen: 37.23 ng
RT: 15.37 min Scan# 1145
Delta R.T. -0.01 min
Lab File: BF085552.D
Acq: 19 Mar 2016 4:17

Instrument :
BNA_F
ClientSampleId :
PB88981BSD

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.2	17.4	26.0
125	15.3	10.3	15.5




Manual Integrations
APPROVED


apatel
3/21/2016 5:08:02 PM



#90
Dibenzo(a,h)anthracene
Concen: 34.71 ng
RT: 16.85 min Scan# 1274
Delta R.T. -0.01 min
Lab File: BF085552.D
Acq: 19 Mar 2016 4:17

Tgt Ion	Ratio	Lower	Upper
278	100		
139	18.3	14.4	21.6
279	24.1	19.1	28.7





#91
 Benzo(g,h,i)perylene
 Concen: 34.73 ng
 RT: 17.25 min Scan# 1309
 Delta R.T. -0.01 min
 Lab File: BF085552.D
 Acq: 19 Mar 2016 4:17

Instrument :
 BNA_F
 ClientSampleId :
 PB88981BSD

Tgt Ion:	276	Resp:	448160
Ion Ratio		Lower	Upper
276	100		
277	24.2	18.8	28.2
138	28.9	22.6	34.0

Manual Integrations
 APPROVED

apatel
 3/21/2016 5:08:02 PM

