

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF031817\
 Data File : BF093756.D
 Acq On : 18 Mar 2017 16:17
 Operator : SJ/MA
 Sample : I2133-10
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 HY-SB415A

Manual Integrations
 APPROVED

mohammad
 3/20/2017 1:00:54 PM

Quant Time: Mar 20 07:50:33 2017
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF031817.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 18 05:21:29 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.85	152	216246	20.00	ng	0.00
21) Naphthalene-d8	8.14	136	782973	20.00	ng	0.00
38) Acenaphthene-d10	9.89	164	266011	20.00	ng	0.00
63) Phenanthrene-d10	11.37	188	481194	20.00	ng	0.01
75) Chrysene-d12	14.00	240	314208	20.00	ng	0.00
86) Perylene-d12	15.43	264	309776	20.00	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.45	112	1650401	127.24	ng	0.01
7) Phenol-d6	6.48	99	1978185	123.05	ng	0.00
23) Nitrobenzene-d5	7.41	82	1261608	96.71	ng	0.00
41) 2,4,6-Tribromophenol	10.68	330	327268	187.98	ng	0.00
44) 2-Fluorobiphenyl	9.21	172	1552231	110.11	ng	0.00
78) Terphenyl-d14	12.96	244	1123173	76.94	ng	0.01
Target Compounds						
10) Phenol	6.49	94	75735	4.03	ng	Qvalue 99
48) Acenaphthylene	9.75	152	86169	3.53	ng	# 66
49) Dimethylphthalate	9.60	163	344437	19.82	ng	# 95
51) Acenaphthene	9.92	154	88244	6.13	ng	95
70) Phenanthrene	11.40	178	204558	7.52	ng	# 87
71) Anthracene	11.44	178	90597	3.40	ng	# 69
74) Fluoranthene	12.58	202	429810	16.30	ng	94
77) Pyrene	12.81	202	618100	23.13	ng	98
80) Benzo(a)anthracene	13.99	228	434059	21.52	ng	91
82) Chrysene	14.02	228	314551m	16.14	ng	
85) Indeno(1,2,3-cd)pyrene	16.83	276	502961	30.47	ng	93
87) Benzo(b)fluoranthene	15.03	252	663956m	33.81	ng	
88) Benzo(k)fluoranthene	15.04	252	207295m	11.94	ng	
89) Benzo(a)pyrene	15.37	252	525883	30.30	ng	# 88
90) Dibenzo(a,h)anthracene	16.83	278	110345m	7.51	ng	
91) Benzo(g,h,i)perylene	17.24	276	554419	37.13	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF031817\
Data File : BF093756.D
Acq On : 18 Mar 2017 16:17
Operator : SJ/MA
Sample : I2133-10
Misc :
ALS Vial : 39 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleID :
HY-SB415A

Manual Integrations
APPROVED

mohammad
3/20/2017 1:00:54 PM

Quant Time: Mar 20 07:50:33 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF031817.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Mar 18 05:21:29 2017
Response via : Initial Calibration

