


Data Path : Z:\HPCHEM1\BNA_F\Data\BF031817\
 Data File : BF093732.D
 Acq On : 18 Mar 2017 5:05
 Operator : SJ/MA
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF031817.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Sat Mar 18 05:21:29 2017



AutoFind: Scans 852, 853, 854; Background Corrected with Scan 849

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	41.7	40663	PASS
68	69	0.00	2	1.9	736	PASS
69	198	0.00	100	38.9	37973	PASS
70	69	0.00	2	0.3	127	PASS
127	198	10	80	51.5	50194	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	97530	PASS
199	198	5	9	6.9	6692	PASS
275	198	10	60	21.9	21347	PASS
365	198	1	100	2.4	2362	PASS
441	442	0.01	24	16.2	13814	PASS
442	198	50	100	87.2	85090	PASS
443	442	15	24	18.6	15796	PASS