

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF031820\
 Data File : BF119415.D
 Acq On : 18 Mar 2020 15:02
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF031820

Quant Time: Mar 18 17:28:29 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF031820.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 18 15:11:18 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
2	1,4-Dioxane	0.621	0.640	-3.1	102	0.00
3	Pyridine	1.709	1.745	-2.1	101	0.00
4	n-Nitrosodimethylamine	0.834	0.845	-1.3	100	0.00
5 S	2-Fluorophenol	1.256	1.305	-3.9	102	0.00
6	Aniline	2.215	2.259	-2.0	99	0.00
7 S	Phenol-d6	1.605	1.656	-3.2	100	0.00
8	2-Chlorophenol	1.320	1.408	-6.7	103	0.00
9	Benzaldehyde	1.018	1.098	-7.9	107	0.00
10 C	Phenol	1.712	1.787	-4.4	102	0.00
11	bis(2-Chloroethyl)ether	1.370	1.434	-4.7	103	0.00
12	1,3-Dichlorobenzene	1.428	1.516	-6.2	105	0.00
13 C	1,4-Dichlorobenzene	1.428	1.514	-6.0	105	0.00
14	1,2-Dichlorobenzene	1.335	1.423	-6.6	103	0.00
15	Benzyl Alcohol	1.207	1.260	-4.4	100	0.00
16	2,2'-oxybis(1-Chloropropane	2.763	2.965	-7.3	106	0.00
17	2-Methylphenol	1.209	1.248	-3.2	101	0.00
18	Hexachloroethane	0.523	0.567	-8.4	107	0.00
19 P	n-Nitroso-di-n-propylamine	0.967	1.007	-4.1	103	0.00
20	3+4-Methylphenols	1.482	1.564	-5.5	102	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	103	0.00
22	Acetophenone	0.478	0.486	-1.7	102	0.00
23 S	Nitrobenzene-d5	0.368	0.391	-6.3	105	0.00
24	Nitrobenzene	0.393	0.411	-4.6	104	0.00
25	Isophorone	0.746	0.755	-1.2	102	0.00
26 C	2-Nitrophenol	0.155	0.174	-12.3	111	0.00
27	2,4-Dimethylphenol	0.320	0.326	-1.9	102	0.00
28	bis(2-Chloroethoxy)methane	0.441	0.454	-2.9	104	0.00
29 C	2,4-Dichlorophenol	0.294	0.304	-3.4	104	0.00
30	1,2,4-Trichlorobenzene	0.325	0.338	-4.0	105	0.00
31	Naphthalene	1.029	1.067	-3.7	103	0.00
32	Benzoic acid	0.206	0.237	-15.0	116	0.00
33	4-Chloroaniline	0.472	0.465	1.5	98	0.00
34 C	Hexachlorobutadiene	0.182	0.193	-6.0	107	0.00
35	Caprolactam	0.096	0.096	0.0	97	0.00
36 C	4-Chloro-3-methylphenol	0.322	0.326	-1.2	101	0.00
37	2-Methylnaphthalene	0.675	0.695	-3.0	103	0.00
38	1-Methylnaphthalene	0.639	0.659	-3.1	102	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
40	1,2,4,5-Tetrachlorobenzene	0.586	0.611	-4.3	104	0.00
41 P	Hexachlorocyclopentadiene	0.333	0.359	-7.8	107	0.00
42 S	2,4,6-Tribromophenol	0.206	0.224	-8.7	107	0.00
43 C	2,4,6-Trichlorophenol	0.420	0.446	-6.2	106	0.00
44	2,4,5-Trichlorophenol	0.470	0.490	-4.3	103	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF031820\
 Data File : BF119415.D
 Acq On : 18 Mar 2020 15:02
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF031820

Quant Time: Mar 18 17:28:29 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF031820.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 18 15:11:18 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.350	1.348	0.1	103	0.00
46	1,1'-Biphenyl	1.629	1.720	-5.6	104	0.00
47	2-Chloronaphthalene	1.276	1.336	-4.7	104	0.00
48	2-Nitroaniline	0.440	0.478	-8.6	104	0.00
49	Acenaphthylene	2.004	2.077	-3.6	102	0.00
50	Dimethylphthalate	1.421	1.491	-4.9	104	0.00
51	2,6-Dinitrotoluene	0.304	0.334	-9.9	107	0.00
52 C	Acenaphthene	1.249	1.289	-3.2	103	0.00
53	3-Nitroaniline	0.384	0.410	-6.8	104	0.00
54 P	2,4-Dinitrophenol	0.112	0.129	-15.2	119	0.00
55	Dibenzofuran	1.791	1.860	-3.9	103	0.00
56 P	4-Nitrophenol	0.303	0.326	-7.6	102	0.00
57	2,4-Dinitrotoluene	0.384	0.434	-13.0	110	0.00
58	Fluorene	1.324	1.379	-4.2	102	0.00
59	2,3,4,6-Tetrachlorophenol	0.351	0.378	-7.7	104	0.00
60	Diethylphthalate	1.442	1.540	-6.8	106	0.00
61	4-Chlorophenyl-phenylether	0.648	0.686	-5.9	105	0.00
62	4-Nitroaniline	0.369	0.387	-4.9	102	0.00
63	Azobenzene	1.452	1.511	-4.1	103	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	103	0.00
65	4,6-Dinitro-2-methylphenol	0.084	0.096	-14.3	114	0.00
66 c	n-Nitrosodiphenylamine	0.657	0.674	-2.6	102	0.00
67	4-Bromophenyl-phenylether	0.228	0.234	-2.6	102	0.00
68	Hexachlorobenzene	0.233	0.245	-5.2	105	0.00
69	Atrazine	0.180	0.186	-3.3	103	0.00
70 C	Pentachlorophenol	0.137	0.146	-6.6	107	0.00
71	Phenanthrene	1.037	1.076	-3.8	104	0.00
72	Anthracene	1.054	1.096	-4.0	102	0.00
73	Carbazole	1.043	1.066	-2.2	101	0.00
74	Di-n-butylphthalate	1.116	1.230	-10.2	110	0.00
75 C	Fluoranthene	1.122	1.168	-4.1	103	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	104	0.00
77	Benzidine	0.846	0.841	0.6	103	0.00
78	Pyrene	1.641	1.616	1.5	102	0.00
79 S	Terphenyl-d14	1.021	1.030	-0.9	106	0.00
80	Butylbenzylphthalate	0.664	0.705	-6.2	111	0.00
81	Benzo(a)anthracene	1.301	1.304	-0.2	101	0.00
82	3,3'-Dichlorobenzidine	0.513	0.533	-3.9	103	0.00
83	Chrysene	1.342	1.376	-2.5	104	0.00
84	Bis(2-ethylhexyl)phthalate	0.791	0.880	-11.3	117	0.00
85 c	Di-n-octyl phthalate	1.392	1.571	-12.9	113	0.00
86	Indeno(1,2,3-cd)pyrene	1.264	1.248	1.3	107	0.00
87 I	Perylene-d12	1.000	1.000	0.0	104	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF031820\
Data File : BF119415.D
Acq On : 18 Mar 2020 15:02
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF031820

Quant Time: Mar 18 17:28:29 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF031820.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 18 15:11:18 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.336	1.430	-7.0	111	0.00
89	Benzo(k)fluoranthene	1.272	1.251	1.7	98	0.00
90 C	Benzo(a)pyrene	1.214	1.257	-3.5	106	0.00
91	Dibenzo(a,h)anthracene	1.062	1.055	0.7	108	0.00
92	Benzo(g,h,i)perylene	1.067	1.002	6.1	104	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0