

Data Path : Z:\HPCHEM1\BNA F\DATA\BF031918\
 Data File : BF103791.D
 Acq On : 20 Mar 2018 8:09
 Operator : JU/SJ
 Sample : J1954-05
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 031518-JETT-F-U-M

Quant Time: Mar 20 10:42:44 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF031518.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 15 18:31:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.97	152	145301	20.00	ng	0.00
21) Naphthalene-d8	8.25	136	580320	20.00	ng	0.00
38) Acenaphthene-d10	10.01	164	266927	20.00	ng	0.00
63) Phenanthrene-d10	11.50	188	431514	20.00	ng	0.00
75) Chrysene-d12	14.16	240	306248	20.00	ng	0.00
86) Perylene-d12	15.70	264	255948	20.00	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.59	112	378269	39.60	ng	0.00
7) Phenol-d6	6.58	99	285497	24.43	ng	0.00
23) Nitrobenzene-d5	7.53	82	880472	105.80	ng	0.00
41) 2,4,6-Tribromophenol	10.80	330	231253	100.76	ng	0.00
44) 2-Fluorobiphenyl	9.33	172	1449440	86.62	ng	0.00
78) Terphenyl-d14	13.09	244	1197103	90.73	ng	0.00
Target Compounds						Qvalue
49) Dimethylphthalate	9.72	163	96646	4.746	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF031918\
Data File : BF103791.D
Acq On : 20 Mar 2018 8:09
Operator : JU/SJ
Sample : J1954-05
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
031518-JETT-F-U-M

Quant Time: Mar 20 10:42:44 2018
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF031518.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 15 18:31:53 2018
Response via : Initial Calibration

