

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF031919\
 Data File : BF113134.D
 Acq On : 19 Mar 2019 19:15
 Operator : JU/SJ
 Sample : K1935-10
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-7(6-7)

Quant Time: Mar 20 01:48:27 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF022719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 13 03:42:55 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.72	152	165742	20.00	ng	-0.01
21) Naphthalene-d8	8.00	136	450422	20.00	ng	-0.01
39) Acenaphthene-d10	9.76	164	180311	20.00	ng	0.00
64) Phenanthrene-d10	11.24	188	296021	20.00	ng	0.00
76) Chrysene-d12	13.87	240	242712	20.00	ng	0.00
87) Perylene-d12	15.29	264	206814	20.00	ng	0.02

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
5) 2-Fluorophenol	5.35	112	992076	93.11	ng	0.00
7) Phenol-d6	6.37	99	1287440	96.19	ng	0.00
23) Nitrobenzene-d5	7.29	82	739395	98.23	ng	-0.01
42) 2,4,6-Tribromophenol	10.55	330	220036	114.95	ng	0.00
45) 2-Fluorobiphenyl	9.09	172	1017680	97.16	ng	0.00
79) Terphenyl-d14	12.83	244	767667	61.31	ng	0.00


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
25) Isophorone	7.63	82	96352	5.942	ng	# 71
35) Caprolactam	8.45	113	50656	22.858	ng	# 37
50) Dimethylphthalate	9.47	163	79880	6.008	ng	# 92
52) Acenaphthene	9.79	154	53539	4.696	ng	# 96
54) 2,4-Dinitrophenol	9.82	184	699	7.021	ng	# 1
56) 4-Nitrophenol	9.90	139	7767	2.833	ng	# 1
65) 4,6-Dinitro-2-methylphenol	10.30	198	1103	5.432	ng	# 1
66) n-Nitrosodiphenylamine	10.43	169	44295	4.666	ng	# 52
71) Phenanthrene	11.27	178	102451	6.956	ng	# 92
72) Anthracene	11.32	178	46087	2.989	ng	# 77
75) Fluoranthene	12.46	202	159571	10.113	ng	# 95
78) Pyrene	12.69	202	214219	10.470	ng	# 97
81) Benzo(a)anthracene	13.87	228	134707	8.008	ng	# 89
83) Chrysene	13.90	228	154852	9.398	ng	# 89
86) Indeno(1,2,3-cd)pyrene	16.65	276	58850	4.189	ng	# 92
88) Benzo(b)fluoranthene	14.89	252	176349	13.183	ng	# 83
89) Benzo(k)fluoranthene	14.89	252	176349	14.554	ng	# 83
90) Benzo(a)pyrene	15.23	252	112463	9.505	ng	# 86
91) Dibenzo(a,h)anthracene	16.66	278	31557	3.125	ng	# 65
92) Benzo(g,h,i)perylene	17.06	276	92394	9.113	ng	# 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF031919\
Data File : BF113134.D
Acq On : 19 Mar 2019 19:15
Operator : JU/SJ
Sample : K1935-10
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-7(6-7)

Quant Time: Mar 20 01:48:27 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF022719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 13 03:42:55 2019
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 6.72 min Scan# 788
 Delta R.T. -0.01 min
 Lab File: BF113134.D
 Acq: 19 Mar 2019 19:15

Instrument :
 BNA_F
 ClientSampleId :
 SB-7(6-7)

Tgt Ion:152 Resp: 165742
 Ion Ratio Lower Upper
 152 100
 150 158.5 124.2 186.2
 115 58.6 48.2 72.4




Abundance Ion 152.00 (151.70 to 152.70): E
 Ion 150.00 (149.70 to 150.70): E
 Ion 115.00 (114.70 to 115.70): E






#5
 2-Fluorophenol
 Concen: 93.110 ng
 RT: 5.35 min Scan# 554
 Delta R.T. -0.00 min
 Lab File: BF113134.D
 Acq: 19 Mar 2019 19:15

Tgt Ion:112 Resp: 992076
 Ion Ratio Lower Upper
 112 100
 64 56.8 47.6 71.4
 63 28.8 24.4 36.6



Abundance Ion 112.00 (111.70 to 112.70): E
 Ion 64.00 (63.70 to 64.70): BF1
 Ion 63.00 (62.70 to 63.70): BF1





#7

Phenol-d6

Concen: 96.190 ng

RT: 6.37 min Scan# 729

Delta R.T. -0.00 min

Lab File: BF113134.D

Acq: 19 Mar 2019 19:15

Instrument :

BNA_F

ClientSampleId :

SB-7(6-7)


Tgt Ion: 99 Resp: 1287440

Ion Ratio Lower Upper

99 100

42 17.7 16.3 24.5




71 31.3 26.2 39.4



Abundance Ion 99.00 (98.70 to 99.70): BF1

Ion 42.00 (41.70 to 42.70): BF1

Ion 71.00 (70.70 to 71.70): BF1



#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.00 min Scan# 1006

Delta R.T. -0.01 min

Lab File: BF113134.D

Acq: 19 Mar 2019 19:15

Tgt Ion: 136 Resp: 450422


Ion Ratio Lower Upper

136 100

137 11.1 8.9 13.3

54 9.6 7.3 10.9

68 7.3 5.4 8.2






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF1

Ion 68.00 (67.70 to 68.70): BF1









#23
Nitrobenzene-d5
Concen: 98.228 ng
RT: 7.29 min Scan# 884
Delta R.T. -0.01 min
Lab File: BF113134.D
Acq: 19 Mar 2019 19:15

Instrument :
BNA_F
ClientSampleId :
SB-7(6-7)

Tgt Ion	Ratio	Lower	Upper
82	100		
128	46.9	36.3	54.5
54	52.0	43.7	65.5




Abundance Ion 82.00 (81.70 to 82.70): BF1
Ion 128.00 (127.70 to 128.70): BF1
Ion 54.00 (53.70 to 54.70): BF1






#25
Isophorone
Concen: 5.942 ng
RT: 7.63 min Scan# 943
Delta R.T. 0.08 min
Lab File: BF113134.D
Acq: 19 Mar 2019 19:15

Tgt Ion	Ratio	Lower	Upper
82	100		
95	7.1	5.2	7.8
138	0.5	13.8	20.8#



Abundance Ion 82.00 (81.70 to 82.70): BF1
Ion 95.00 (94.70 to 95.70): BF1
Ion 138.00 (137.70 to 138.70): BF1





#35

Caprolactam

Concen: 22.858 ng

RT: 8.45 min Scan# 1081

Delta R.T. -0.02 min

Lab File: BF113134.D

Acq: 19 Mar 2019 19:15

Instrument :

BNA_F

ClientSampled :

SB-7(6-7)


Tgt Ion:113 Resp: 50656

Ion Ratio Lower Upper

113 100

55 255.4 163.5 203.5#




56 230.7 115.3 155.3#



Abundance Ion 113.00 (112.70 to 113.70): E

Ion 55.00 (54.70 to 55.70): BF1

Ion 56.00 (55.70 to 56.70): BF1



#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.76 min Scan# 1305

Delta R.T. -0.01 min

Lab File: BF113134.D

Acq: 19 Mar 2019 19:15


Tgt Ion:164 Resp: 180311

Ion Ratio Lower Upper

164 100

162 99.4 80.6 120.8



160 45.1 36.0 54.0



Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E









#42
 2,4,6-Tribromophenol
 Concen: 114.948 ng
 RT: 10.55 min Scan# 1439
 Delta R.T. -0.00 min
 Lab File: BF113134.D
 Acq: 19 Mar 2019 19:15

Instrument :
 BNA_F
 ClientSampleId :
 SB-7(6-7)

Tgt Ion:330 Resp: 220036
 Ion Ratio Lower Upper
 330 100
 332 96.5 76.2 114.4
 141 26.7 27.0 40.6#




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E






#45
 2-Fluorobiphenyl
 Concen: 97.162 ng
 RT: 9.09 min Scan# 1190
 Delta R.T. -0.01 min
 Lab File: BF113134.D
 Acq: 19 Mar 2019 19:15

Tgt Ion:172 Resp: 1017680
 Ion Ratio Lower Upper
 172 100
 171 36.0 29.0 43.4
 170 24.5 20.0 30.0



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E









#50
Dimethylphthalate
Concen: 6.008 ng
RT: 9.47 min Scan# 1256
Delta R.T. -0.02 min
Lab File: BF113134.D
Acq: 19 Mar 2019 19:15

Instrument :
BNA_F
ClientSampled :
SB-7(6-7)

Tgt Ion:163 Resp: 79880
Ion Ratio Lower Upper
163 100
194 11.2 3.0 4.4#
164 11.7 8.4 12.6




Abundance Ion 163.00 (162.70 to 163.70): E
Ion 194.00 (193.70 to 194.70): E
Ion 164.00 (163.70 to 164.70): E






#52
Acenaphthene
Concen: 4.696 ng
RT: 9.79 min Scan# 1310
Delta R.T. -0.01 min
Lab File: BF113134.D
Acq: 19 Mar 2019 19:15

Tgt Ion:154 Resp: 53539
Ion Ratio Lower Upper
154 100
153 115.0 87.8 131.8
152 55.4 43.0 64.6



Abundance Ion 154.00 (153.70 to 154.70): E
Ion 153.00 (152.70 to 153.70): E
Ion 152.00 (151.70 to 152.70): E









#54
 2,4-Dinitrophenol
 Concen: 7.021 ng
 RT: 9.82 min Scan# 1314
 Delta R.T. -0.01 min
 Lab File: BF113134.D
 Acq: 19 Mar 2019 19:15




Instrument :
 BNA_F
 ClientSampleId :
 SB-7(6-7)


Tgt Ion:184 Resp: 699
 Ion Ratio Lower Upper
 184 100
 63 253.8 45.3 67.9#
 154 390.6 46.0 69.0#



#56
 4-Nitrophenol
 Concen: 2.833 ng
 RT: 9.90 min Scan# 1329
 Delta R.T. 0.02 min
 Lab File: BF113134.D
 Acq: 19 Mar 2019 19:15

Tgt Ion:139 Resp: 7767
 Ion Ratio Lower Upper
 139 100
 109 303.4 46.8 86.8#
 65 74.7 80.7 120.7#





#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.24 min Scan# 1557

Delta R.T. -0.00 min

Lab File: BF113134.D

Acq: 19 Mar 2019 19:15

Instrument :

BNA_F

ClientSampleId :

SB-7(6-7)


Tgt Ion:188 Resp: 296021

Ion Ratio Lower Upper

188 100

94 10.1 9.0 13.4




80 9.9 9.2 13.8



Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BF1

Ion 80.00 (79.70 to 80.70): BF1



#65

4,6-Dinitro-2-methylphenol

Concen: 5.432 ng

RT: 10.30 min Scan# 1397

Delta R.T. -0.06 min

Lab File: BF113134.D

Acq: 19 Mar 2019 19:15


Tgt Ion:198 Resp: 1103

Ion Ratio Lower Upper

198 100

51 292.1 43.8 83.8#



105 1326.6 36.5 76.5#




Abundance Ion 198.00 (197.70 to 198.70): E

Ion 51.00 (50.70 to 51.70): BF1

Ion 105.00 (104.70 to 105.70): E









#66
n-Nitrosodiphenylamine
Concen: 4.666 ng
RT: 10.43 min Scan# 1418
Delta R.T. 0.01 min
Lab File: BF113134.D
Acq: 19 Mar 2019 19:15




Instrument :
BNA_F
ClientSampled :
SB-7(6-7)


Tgt Ion:169 Resp: 44295
Ion Ratio Lower Upper
169 100
168 24.9 53.0 79.6#
167 12.0 29.3 43.9#



#71
Phenanthrene
Concen: 6.956 ng
RT: 11.27 min Scan# 1561
Delta R.T. -0.00 min
Lab File: BF113134.D
Acq: 19 Mar 2019 19:15

Tgt Ion:178 Resp: 102451
Ion Ratio Lower Upper
178 100
176 19.9 16.3 24.5
179 23.3 12.7 19.1#









#72
 Anthracene
 Concen: 2.989 ng
 RT: 11.32 min Scan# 1569
 Delta R.T. -0.01 min
 Lab File: BF113134.D
 Acq: 19 Mar 2019 19:15




Instrument :
 BNA_F
 ClientSampleId :
 SB-7(6-7)


Tgt Ion:178 Resp: 46087
 Ion Ratio Lower Upper
 178 100
 176 22.3 15.8 23.6
 179 34.8 13.0 19.6#



#75
 Fluoranthene
 Concen: 10.113 ng
 RT: 12.46 min Scan# 1763
 Delta R.T. 0.01 min
 Lab File: BF113134.D
 Acq: 19 Mar 2019 19:15

Tgt Ion:202 Resp: 159571
 Ion Ratio Lower Upper
 202 100
 101 11.6 0.0 33.8
 203 20.4 0.0 38.4









#76
Chrysene-d12
Concen: 20.000 ng
RT: 13.87 min Scan# 2004
Delta R.T. 0.01 min
Lab File: BF113134.D
Acq: 19 Mar 2019 19:15




Instrument :
BNA_F
ClientSampleId :
SB-7(6-7)


Tgt Ion:240 Resp: 242712
Ion Ratio Lower Upper
240 100
120 13.1 8.9 13.3
236 26.4 20.7 31.1



#78
Pyrene
Concen: 10.470 ng
RT: 12.69 min Scan# 1802
Delta R.T. 0.01 min
Lab File: BF113134.D
Acq: 19 Mar 2019 19:15

Tgt Ion:202 Resp: 214219
Ion Ratio Lower Upper
202 100
200 19.7 17.0 25.4
203 19.7 14.4 21.6









#79
 Terphenyl-d14
 Concen: 61.312 ng
 RT: 12.83 min Scan# 1826
 Delta R.T. 0.01 min
 Lab File: BF113134.D
 Acq: 19 Mar 2019 19:15




Instrument :
 BNA_F
 ClientSampleId :
 SB-7(6-7)


Tgt Ion:244 Resp: 767667
 Ion Ratio Lower Upper
 244 100
 212 7.0 5.6 8.4
 122 11.7 10.5 15.7



#81
 Benzo(a)anthracene
 Concen: 8.008 ng
 RT: 13.87 min Scan# 2003
 Delta R.T. 0.01 min
 Lab File: BF113134.D
 Acq: 19 Mar 2019 19:15

Tgt Ion:228 Resp: 134707
 Ion Ratio Lower Upper
 228 100
 226 29.0 22.2 33.4
 229 30.2 16.0 24.0#





#83

Chrysene

Concen: 9.398 ng

RT: 13.90 min Scan# 2008

Delta R.T. -0.00 min

Lab File: BF113134.D

Acq: 19 Mar 2019 19:15

Instrument :

BNA_F

ClientSampleId :

SB-7(6-7)


Tgt Ion:228 Resp: 154852

Ion Ratio Lower Upper

228 100

226 31.4 24.9 37.3




229 32.5 16.3 24.5#



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#86

Indeno(1,2,3-cd)pyrene

Concen: 4.189 ng

RT: 16.65 min Scan# 2476

Delta R.T. 0.01 min

Lab File: BF113134.D

Acq: 19 Mar 2019 19:15


Tgt Ion:276 Resp: 58850

Ion Ratio Lower Upper

276 100

138 22.5 23.2 34.8#



277 26.2 19.9 29.9




Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E









#87
Perylene-d12
Concen: 20.000 ng
RT: 15.29 min Scan# 2245
Delta R.T. 0.02 min
Lab File: BF113134.D
Acq: 19 Mar 2019 19:15




Instrument :
BNA_F
ClientSampleId :
SB-7(6-7)


Tgt Ion:264 Resp: 206814
Ion Ratio Lower Upper
264 100
260 24.5 18.7 28.1
265 24.5 17.2 25.8



#88
Benzo(b)fluoranthene
Concen: 13.183 ng
RT: 14.89 min Scan# 2176
Delta R.T. 0.01 min
Lab File: BF113134.D
Acq: 19 Mar 2019 19:15

Tgt Ion:252 Resp: 176349
Ion Ratio Lower Upper
252 100
253 26.0 18.0 27.0
125 26.3 10.2 15.2#









#89
Benzo(k)fluoranthene
Concen: 14.554 ng
RT: 14.89 min Scan# 2176
Delta R.T. -0.02 min
Lab File: BF113134.D
Acq: 19 Mar 2019 19:15

Instrument :
BNA_F
ClientSampleId :
SB-7(6-7)

Tgt Ion	Ratio	Lower	Upper
252	100		
253	26.0	18.2	27.2
125	26.3	9.8	14.6#




Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E






#90
Benzo(a)pyrene
Concen: 9.505 ng
RT: 15.23 min Scan# 2234
Delta R.T. 0.01 min
Lab File: BF113134.D
Acq: 19 Mar 2019 19:15

Tgt Ion	Ratio	Lower	Upper
252	100		
253	25.8	18.0	27.0
125	24.0	10.7	16.1#



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E





#91

Dibenzo(a,h)anthracene

Concen: 3.125 ng

RT: 16.66 min Scan# 2477

Delta R.T. 0.01 min

Lab File: BF113134.D

Acq: 19 Mar 2019 19:15

Instrument :

BNA_F

ClientSampleId :

SB-7(6-7)


Tgt Ion: 278 Resp: 31557

Ion Ratio Lower Upper

278 100

139 35.5 14.3 21.5#

279 39.7 19.4 29.0#






Abundance

Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E



#92

Benzo(g,h,i)perylene

Concen: 9.113 ng

RT: 17.06 min Scan# 2545

Delta R.T. 0.01 min

Lab File: BF113134.D

Acq: 19 Mar 2019 19:15


Tgt Ion: 276 Resp: 92394

Ion Ratio Lower Upper

276 100

277 26.5 18.6 27.8

138 24.4 19.4 29.0



Abundance

Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

