




Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF031919\  
Data File : BF113122.D  
Acq On : 19 Mar 2019 10:56  
Operator : JU/SJ  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
DFTPP

Quant Time: Mar 20 01:44:10 2019  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_F\METHODS\8270-BF022719.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Mar 13 03:42:55 2019  
Response via : Initial Calibration



TIC: BF113122.D

(70) Pentachlorophenol (C)

11.045min (-0.012) 33225.98ng




response 126779

| Ion    | Exp%  | Act%  |
|--------|-------|-------|
| 265.70 | 100   | 100   |
| 268.00 | 66.60 | 66.36 |
| 264.00 | 61.10 | 61.87 |
| 0.00   | 0.00  | 0.00  |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF031919\  
 Data File : BF113122.D  
 Acq On : 19 Mar 2019 10:56  
 Operator : JU/SJ  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 DFTPP

Quant Time: Mar 20 01:44:10 2019  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_F\METHODS\8270-BF022719.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed Mar 13 03:42:55 2019  
 Response via : Initial Calibration



TIC: BF113122.D

(77) Benzidine

12.568min (-0.000) 14231.74ng

response 812204

| Ion    | Exp%  | Act%  |
|--------|-------|-------|
| 184.00 | 100   | 100   |
| 185.00 | 15.70 | 14.21 |
| 183.00 | 10.70 | 10.72 |
| 0.00   | 0.00  | 0.00  |

Data Path : Z:\SVOASRV\HPCHEM1\BNA\_F\DATA\BF031919\  
 Data File : BF113129.D  
 Acq On : 19 Mar 2019 16:55  
 Operator : JU/SJ  
 Sample : K1940-01  
 Misc :  
 ALS Vial : 8 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 DFTPP

Quant Time: Mar 20 01:47:11 2019  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_F\METHODS\8270-BF022719.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed Mar 13 03:42:55 2019  
 Response via : Initial Calibration

| Internal Standards        | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|---------------------------|-------|------|----------|-------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4 | 6.72  | 152  | 183286   | 20.00 | ng    | -0.01    |
| 21) Naphthalene-d8        | 8.00  | 136  | 579118   | 20.00 | ng    | -0.01    |
| 39) Acenaphthene-d10      | 9.76  | 164  | 304529   | 20.00 | ng    | -0.01    |
| 64) Phenanthrene-d10      | 11.24 | 188  | 565888   | 20.00 | ng    | 0.00     |
| 76) Chrysene-d12          | 13.87 | 240  | 460167   | 20.00 | ng    | 0.00     |
| 87) Perylene-d12          | 15.28 | 264  | 446664   | 20.00 | ng    | 0.00     |

| System Monitoring Compounds | R.T.  | QIon | Response | Conc   | Units | Dev(Min) |
|-----------------------------|-------|------|----------|--------|-------|----------|
| 5) 2-Fluorophenol           | 5.36  | 112  | 1181504  | 100.27 | ng    | 0.01     |
| 7) Phenol-d6                | 6.37  | 99   | 1286209  | 86.90  | ng    | 0.00     |
| 23) Nitrobenzene-d5         | 7.29  | 82   | 902576   | 93.26  | ng    | -0.01    |
| 42) 2,4,6-Tribromophenol    | 10.55 | 330  | 405692   | 125.49 | ng    | 0.00     |
| 45) 2-Fluorobiphenyl        | 9.09  | 172  | 1570626  | 85.62  | ng    | 0.00     |
| 79) Terphenyl-d14           | 12.82 | 244  | 1560859  | 65.75  | ng    | 0.00     |


| Target Compounds               | R.T.  | QIon | Response | Conc    | Units | Qvalue |
|--------------------------------|-------|------|----------|---------|-------|--------|
| 15) Benzyl Alcohol             | 6.88  | 79   | 23411    | 2.074   | ng    | # 56   |
| 20) 3+4-Methylphenols          | 7.14  | 107  | 79642    | 6.630   | ng    | # 56   |
| 25) Isophorone                 | 7.29  | 82   | 890176   | 42.694  | ng    | # 66   |
| 32) Benzoic acid               | 8.05  | 122  | 3655610  | 625.226 | ng    | 95     |
| 50) Dimethylphthalate          | 9.76  | 163  | 85739    | 3.818   | ng    | # 1    |
| 54) 2,4-Dinitrophenol          | 9.77  | 184  | 636      | 6.759   | ng    | # 1    |
| 57) 2,4-Dinitrotoluene         | 9.76  | 165  | 42526    | 7.317   | ng    | # 20   |
| 65) 4,6-Dinitro-2-methylphenol | 10.47 | 198  | 1017     | 5.008   | ng    | # 1    |
| 71) Phenanthrene               | 11.26 | 178  | 77634    | 2.757   | ng    | 96     |
| 72) Anthracene                 | 11.26 | 178  | 77634    | 2.634   | ng    | 97     |
| 75) Fluoranthene               | 12.45 | 202  | 62181    | 2.061   | ng    | 97     |
| 77) Benzidine                  | 12.44 | 184  | 206607   | 8.654   | ng    | # 89   |
| 84) Bis(2-ethylhexyl)phthalate | 13.86 | 149  | 125410   | 6.244   | ng    | 99     |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA\_F\DATA\BF031919\  
Data File : BF113129.D  
Acq On : 19 Mar 2019 16:55  
Operator : JU/SJ  
Sample : K1940-01  
Misc :  
ALS Vial : 8 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
DFTPP

Quant Time: Mar 20 01:47:11 2019  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_F\METHODS\8270-BF022719.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Mar 13 03:42:55 2019  
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4  
 Concen: 20.000 ng  
 RT: 6.72 min Scan# 788  
 Delta R.T. -0.01 min  
 Lab File: BF113129.D  
 Acq: 19 Mar 2019 16:55

Instrument :  
 BNA\_F  
 ClientSampled :  
 DFTPP

Tgt Ion:152 Resp: 183286  
 Ion Ratio Lower Upper  
 152 100  
 150 155.7 124.2 186.2  
 115 59.4 48.2 72.4




Abundance Ion 152.00 (151.70 to 152.70): E  
 Ion 150.00 (149.70 to 150.70): E  
 Ion 115.00 (114.70 to 115.70): E





#5


2-Fluorophenol  
 Concen: 100.274 ng  
 RT: 5.36 min Scan# 556  
 Delta R.T. 0.01 min  
 Lab File: BF113129.D  
 Acq: 19 Mar 2019 16:55

Tgt Ion:112 Resp: 1181504  
 Ion Ratio Lower Upper  
 112 100  
 64 60.3 47.6 71.4  
 63 30.6 24.4 36.6



Abundance Ion 112.00 (111.70 to 112.70): E  
 Ion 64.00 (63.70 to 64.70): BF1  
 Ion 63.00 (62.70 to 63.70): BF1





#7

Phenol-d6

Concen: 86.900 ng

RT: 6.37 min Scan# 729

Delta R.T. 0.00 min

Lab File: BF113129.D

Acq: 19 Mar 2019 16:55

Instrument :

BNA\_F

ClientSampleId :

DFTPP


Tgt Ion: 99 Resp: 1286209

Ion Ratio Lower Upper




99 100

42 19.4 16.3 24.5

71 33.1 26.2 39.4



Abundance Ion 99.00 (98.70 to 99.70): BF1  
Ion 42.00 (41.70 to 42.70): BF1  
Ion 71.00 (70.70 to 71.70): BF1



#15

Benzyl Alcohol

Concen: 2.074 ng

RT: 6.88 min Scan# 815

Delta R.T. 0.00 min

Lab File: BF113129.D

Acq: 19 Mar 2019 16:55


Tgt Ion: 79 Resp: 23411

Ion Ratio Lower Upper



79 100


108 39.2 64.2 96.4#

77 95.3 51.1 76.7#



Abundance Ion 79.00 (78.70 to 79.70): BF1  
Ion 108.00 (107.70 to 108.70): BF1  
Ion 77.00 (76.70 to 77.70): BF1





#20

3+4-Methylphenols

Concen: 6.630 ng

RT: 7.14 min Scan# 859

Delta R.T. -0.01 min

Lab File: BF113129.D

Acq: 19 Mar 2019 16:55





Instrument :

BNA\_F

ClientSampleId :

DFTPP

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 107   | Resp: | 79642  |
| Ion      | Ratio | Lower | Upper  |
| 107      | 100   |       |        |
| 108      | 80.3  | 71.4  | 111.4  |
| 77       | 29.8  | 94.1  | 134.1# |
| 79       | 21.1  | 8.6   | 48.6   |



#21

Naphthalene-d8

Concen: 20.000 ng




RT: 8.00 min Scan# 1006


Delta R.T. -0.01 min

Lab File: BF113129.D

Acq: 19 Mar 2019 16:55

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 136   | Resp: | 579118 |
| Ion      | Ratio | Lower | Upper  |
| 136      | 100   |       |        |
| 137      | 11.1  | 8.9   | 13.3   |
| 54       | 8.0   | 7.3   | 10.9   |
| 68       | 5.6   | 5.4   | 8.2    |









#23  
Nitrobenzene-d5  
Concen: 93.260 ng  
RT: 7.29 min Scan# 884  
Delta R.T. -0.01 min  
Lab File: BF113129.D  
Acq: 19 Mar 2019 16:55

Instrument :  
BNA\_F  
ClientSampleId :  
DFTPP

Tgt Ion: 82 Resp: 902576  
Ion Ratio Lower Upper  
82 100  
128 45.0 36.3 54.5  
54 53.6 43.7 65.5




Abundance Ion 82.00 (81.70 to 82.70): BF1  
Ion 128.00 (127.70 to 128.70): BF1  
Ion 54.00 (53.70 to 54.70): BF1






#25  
Isophorone  
Concen: 42.694 ng  
RT: 7.29 min Scan# 884  
Delta R.T. -0.27 min  
Lab File: BF113129.D  
Acq: 19 Mar 2019 16:55

Tgt Ion: 82 Resp: 890176  
Ion Ratio Lower Upper  
82 100  
95 0.1 5.2 7.8#  
138 0.0 13.8 20.8#



Abundance Ion 82.00 (81.70 to 82.70): BF1  
Ion 95.00 (94.70 to 95.70): BF1  
Ion 138.00 (137.70 to 138.70): BF1





#32

Benzoic acid

Concen: 625.226 ng

RT: 8.05 min Scan# 1013

Delta R.T. 0.24 min

Lab File: BF113129.D

Acq: 19 Mar 2019 16:55

Instrument :

BNA\_F

ClientSampleId :

DFTPP


Tgt Ion:122 Resp: 3655610

Ion Ratio Lower Upper

122 100

105 121.2 106.8 146.8




77 93.4 79.4 119.4



Abundance Ion 122.00 (121.70 to 122.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 77.00 (76.70 to 77.70): BF1



#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.76 min Scan# 1304

Delta R.T. -0.01 min

Lab File: BF113129.D

Acq: 19 Mar 2019 16:55


Tgt Ion:164 Resp: 304529

Ion Ratio Lower Upper

164 100

162 102.9 80.6 120.8



160 46.6 36.0 54.0




Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E





#42

2,4,6-Tribromophenol

Concen: 125.486 ng

RT: 10.55 min Scan# 1438

Delta R.T. -0.01 min

Lab File: BF113129.D

Acq: 19 Mar 2019 16:55

Instrument :

BNA\_F

ClientSampleId :

DFTPP


Tgt Ion:330 Resp: 405692

Ion Ratio Lower Upper

330 100

332 95.8 76.2 114.4




141 28.7 27.0 40.6



Abundance Ion 329.80 (329.50 to 330.50): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E



#45

2-Fluorobiphenyl

Concen: 85.618 ng

RT: 9.09 min Scan# 1190

Delta R.T. -0.01 min

Lab File: BF113129.D

Acq: 19 Mar 2019 16:55


Tgt Ion:172 Resp: 1570626

Ion Ratio Lower Upper

172 100

171 36.2 29.0 43.4



170 24.7 20.0 30.0




Abundance Ion 172.00 (171.70 to 172.70): E

Ion 171.00 (170.70 to 171.70): E

Ion 170.00 (169.70 to 170.70): E









#50  
Dimethylphthalate  
Concen: 3.818 ng  
RT: 9.76 min Scan# 1304  
Delta R.T. 0.26 min  
Lab File: BF113129.D  
Acq: 19 Mar 2019 16:55

Instrument :  
BNA\_F  
ClientSampleId :  
DFTPP

Tgt Ion:163 Resp: 85739  
Ion Ratio Lower Upper  
163 100  
194 0.0 3.0 4.4#  
164 356.0 8.4 12.6#




Abundance Ion 163.00 (162.70 to 163.70): E  
Ion 194.00 (193.70 to 194.70): E  
Ion 164.00 (163.70 to 164.70): E






#54  
2,4-Dinitrophenol  
Concen: 6.759 ng  
RT: 9.77 min Scan# 1307  
Delta R.T. -0.05 min  
Lab File: BF113129.D  
Acq: 19 Mar 2019 16:55

Tgt Ion:184 Resp: 636  
Ion Ratio Lower Upper  
184 100  
63 183.3 45.3 67.9#  
154 200.2 46.0 69.0#



Abundance Ion 184.00 (183.70 to 184.70): E  
Ion 63.00 (62.70 to 63.70): BF1  
Ion 154.00 (153.70 to 154.70): E





#57

2,4-Dinitrotoluene

Concen: 7.317 ng

RT: 9.76 min Scan# 1304

Delta R.T. -0.20 min

Lab File: BF113129.D

Acq: 19 Mar 2019 16:55

Instrument :

BNA\_F

ClientSampleId :

DFTPP

Tgt Ion:165 Resp: 42526



Ion Ratio Lower Upper

165 100

63 2.4 42.2 63.2#

89 3.4 66.2 99.4#

182 2.1 1.8 2.6





Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BF1

Ion 89.00 (88.70 to 89.70): BF1

Ion 182.00 (181.70 to 182.70): E



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.24 min Scan# 1556

Delta R.T. -0.01 min

Lab File: BF113129.D

Acq: 19 Mar 2019 16:55



Tgt Ion:188 Resp: 565888

Ion Ratio Lower Upper

188 100

94 7.8 9.0 13.4#


80 8.2 9.2 13.8#




Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BF1

Ion 80.00 (79.70 to 80.70): BF1





#65

4,6-Dinitro-2-methylphenol

Concen: 5.008 ng

RT: 10.47 min Scan# 1425

Delta R.T. 0.11 min

Lab File: BF113129.D

Acq: 19 Mar 2019 16:55

Instrument :

BNA\_F

ClientSampleId :

DFTPP


Tgt Ion:198 Resp: 1017

Ion Ratio Lower Upper

198 100

51 192.2 43.8 83.8#




105 636.8 36.5 76.5#



Abundance Ion 198.00 (197.70 to 198.70): E

Ion 51.00 (50.70 to 51.70): BF1

Ion 105.00 (104.70 to 105.70): E



#71

Phenanthrene

Concen: 2.757 ng

RT: 11.26 min Scan# 1559

Delta R.T. -0.01 min

Lab File: BF113129.D

Acq: 19 Mar 2019 16:55


Tgt Ion:178 Resp: 77634

Ion Ratio Lower Upper

178 100

176 19.0 16.3 24.5



179 18.3 12.7 19.1




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E









#72  
 Anthracene  
 Concen: 2.634 ng  
 RT: 11.26 min Scan# 1559  
 Delta R.T. -0.06 min  
 Lab File: BF113129.D  
 Acq: 19 Mar 2019 16:55




Instrument :  
 BNA\_F  
 ClientSampleId :  
 DFTPP


Tgt Ion:178 Resp: 77634  
 Ion Ratio Lower Upper  
 178 100  
 176 19.0 15.8 23.6  
 179 18.3 13.0 19.6



#75  
 Fluoranthene  
 Concen: 2.061 ng  
 RT: 12.45 min Scan# 1761  
 Delta R.T. -0.01 min  
 Lab File: BF113129.D  
 Acq: 19 Mar 2019 16:55

Tgt Ion:202 Resp: 62181  
 Ion Ratio Lower Upper  
 202 100  
 101 11.3 0.0 33.8  
 203 17.8 0.0 38.4









#76  
Chrysene-d12  
Concen: 20.000 ng  
RT: 13.87 min Scan# 2003  
Delta R.T. 0.00 min  
Lab File: BF113129.D  
Acq: 19 Mar 2019 16:55

Instrument :  
BNA\_F  
ClientSampleId :  
DFTPP

Tgt Ion:240 Resp: 460167  
Ion Ratio Lower Upper  
240 100  
120 10.6 8.9 13.3  
236 26.2 20.7 31.1




Abundance Ion 240.00 (239.70 to 240.70): E  
Ion 120.00 (119.70 to 120.70): E  
Ion 236.00 (235.70 to 236.70): E






#77  
Benzidine  
Concen: 8.654 ng  
RT: 12.44 min Scan# 1760  
Delta R.T. -0.13 min  
Lab File: BF113129.D  
Acq: 19 Mar 2019 16:55

Tgt Ion:184 Resp: 206607  
Ion Ratio Lower Upper  
184 100  
185 15.0 12.6 18.8  
183 1.7 8.6 12.8#



Abundance Ion 184.00 (183.70 to 184.70): E  
Ion 185.00 (184.70 to 185.70): E  
Ion 183.00 (182.70 to 183.70): E





#79

Terphenyl-d14

Concen: 65.753 ng

RT: 12.82 min Scan# 1825

Delta R.T. 0.00 min

Lab File: BF113129.D

Acq: 19 Mar 2019 16:55

Instrument :

BNA\_F

ClientSampleId :

DFTPP





Tgt Ion:244 Resp: 1560859

Ion Ratio Lower Upper

244 100

212 7.0 5.6 8.4

122 11.7 10.5 15.7



#84

Bis(2-ethylhexyl)phthalate

Concen: 6.244 ng

RT: 13.86 min Scan# 2001

Delta R.T. 0.00 min

Lab File: BF113129.D

Acq: 19 Mar 2019 16:55




Tgt Ion:149 Resp: 125410


Ion Ratio Lower Upper

149 100

167 27.7 22.4 33.6

279 4.6 3.3 4.9





#87  
Perylene-d12  
Concen: 20.000 ng  
RT: 15.28 min Scan# 2243  
Delta R.T. 0.01 min  
Lab File: BF113129.D  
Acq: 19 Mar 2019 16:55

Instrument :  
BNA\_F  
ClientSampleId :  
DFTPP

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 264     | 100   |       |       |
| 260     | 24.0  | 18.7  | 28.1  |
| 265     | 21.6  | 17.2  | 25.8  |

