

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF031919\
 Data File : BF113158.D
 Acq On : 20 Mar 2019 7:20
 Operator : JU/SJ
 Sample : K1948-15
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PST-025-SW080-031319

Quant Time: Mar 20 08:23:47 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF022719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 13 03:42:55 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.72	152	150081	20.00	ng	-0.01
21) Naphthalene-d8	8.00	136	513415	20.00	ng	-0.01
39) Acenaphthene-d10	9.75	164	223501	20.00	ng	-0.02
64) Phenanthrene-d10	11.23	188	373093	20.00	ng	-0.01
76) Chrysene-d12	13.86	240	321333	20.00	ng	0.00
87) Perylene-d12	15.27	264	261755	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.35	112	961845	99.69	ng	0.00
7) Phenol-d6	6.37	99	1264137	104.30	ng	0.00
23) Nitrobenzene-d5	7.28	82	639639	74.55	ng	-0.02
42) 2,4,6-Tribromophenol	10.54	330	242327	102.13	ng	-0.01
45) 2-Fluorobiphenyl	9.08	172	1096894	80.08	ng	-0.01
79) Terphenyl-d14	12.82	244	945623	57.05	ng	0.00
Target Compounds						
50) Dimethylphthalate	9.47	163	98907	6.002	ng	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF031919\
Data File : BF113158.D
Acq On : 20 Mar 2019 7:20
Operator : JU/SJ
Sample : K1948-15
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PST-025-SW080-031319

Quant Time: Mar 20 08:23:47 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF022719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 13 03:42:55 2019
Response via : Initial Calibration

