

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF031920\
 Data File : BF119449.D
 Acq On : 19 Mar 2020 20:26
 Operator : CG/JU
 Sample : L1954-06
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TP-11A

Quant Time: Mar 20 04:27:13 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF031820.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 18 15:11:18 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.85	152	333134	20.00	ng	0.00
21) Naphthalene-d8	8.13	136	1274328	20.00	ng	-0.01
39) Acenaphthene-d10	9.89	164	663662	20.00	ng	-0.01
64) Phenanthrene-d10	11.38	188	1162135	20.00	ng	-0.01
76) Chrysene-d12	14.02	240	656145	20.00	ng	-0.01
87) Perylene-d12	15.49	264	683696	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.46	112	2053868	98.15	ng	0.00
7) Phenol-d6	6.47	99	2627232	98.28	ng	0.00
23) Nitrobenzene-d5	7.41	82	1592063	67.90	ng	-0.01
42) 2,4,6-Tribromophenol	10.68	330	663322	96.88	ng	-0.01
45) 2-Fluorobiphenyl	9.22	172	2858439	63.81	ng	0.00
79) Terphenyl-d14	12.97	244	2681582	80.07	ng	0.00
Target Compounds						
50) Dimethylphthalate	9.60	163	231737	4.916	ng	100
75) Fluoranthene	12.59	202	170543	2.615	ng	95
78) Pyrene	12.82	202	152993	2.841	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF031920\
Data File : BF119449.D
Acq On : 19 Mar 2020 20:26
Operator : CG/JU
Sample : L1954-06
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TP-11A

Quant Time: Mar 20 04:27:13 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF031820.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 18 15:11:18 2020
Response via : Initial Calibration

