


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF032015\
 Data File : BF077880.D
 Acq On : 21 Mar 2015 00:14
 Operator : TP/IZ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF031115.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Mar 20 23:52:05 2015



AutoFind: Scans 1067, 1068, 1069; Background Corrected with Scan 1064

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.6	3853	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	37.3	4037	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	52.7	5703	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	10829	PASS
199	198	5	9	6.6	711	PASS
275	198	10	60	26.5	2865	PASS
365	198	1	100	2.4	256	PASS
441	442	0.01	24	15.6	1408	PASS
442	198	50	100	83.5	9041	PASS
443	442	15	24	19.5	1759	PASS