

Data Path : Z:\HPCHEM1\BNA F\DATA\BF032018\
 Data File : BF103803.D
 Acq On : 20 Mar 2018 15:46
 Operator : JU/SJ
 Sample : J1821-04RE
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :

Quant Time: Mar 20 17:54:06 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF031518.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 15 18:31:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.97	152	179419	20.00	ng	0.00
21) Naphthalene-d8	8.25	136	787399	20.00	ng	0.00
38) Acenaphthene-d10	10.01	164	345442	20.00	ng	0.00
63) Phenanthrene-d10	11.50	188	611077	20.00	ng	0.00
75) Chrysene-d12	14.16	240	476556	20.00	ng	0.00
86) Perylene-d12	15.69	264	377397	20.00	ng	0.01

System Monitoring Compounds

5) 2-Fluorophenol	5.60	112	79430	6.73	ng	0.02
7) Phenol-d6	6.58	99	275227	19.07	ng	0.00
23) Nitrobenzene-d5	7.53	82	507300	44.93	ng	0.00
41) 2,4,6-Tribromophenol	10.80	330	14375	4.84	ng	0.00
44) 2-Fluorobiphenyl	9.33	172	944764	43.63	ng	0.00
78) Terphenyl-d14	13.09	244	1150796	56.05	ng	0.00

Target Compounds


						Qvalue
19) n-Nitroso-di-n-propylamine	7.53	70	66806	7.223	ng	# 75
25) Isophorone	7.62	82	70904	2.713	ng	# 51
26) 2-Nitrophenol	8.08	139	1773	9.728	ng	# 46
32) Benzoic acid	8.02	122	1117	10.017	ng	# 23
47) 2-Nitroaniline	9.55	65	229	6.847	ng	# 2
50) 2,6-Dinitrotoluene	9.75	165	3778	4.663	ng	# 24
52) 3-Nitroaniline	9.96	138	314	3.862	ng	# 2
55) 4-Nitrophenol	10.10	139	170	5.003	ng	# 1
56) 2,4-Dinitrotoluene	10.01	165	45966	12.409	ng	# 23
61) 4-Nitroaniline	10.56	138	291	3.795	ng	# 37
70) Phenanthrene	11.53	178	70972	2.123	ng	# 98
73) Di-n-butylphthalate	12.06	149	1174384	29.660	ng	100
74) Fluoranthene	12.73	202	164353	4.772	ng	97
77) Pyrene	12.95	202	162427	4.641	ng	99
80) Benzo(a)anthracene	14.18	228	76894	2.549	ng	97
82) Chrysene	14.18	228	76894	2.674	ng	97
83) Bis(2-ethylhexyl)phthalate	14.12	149	74456	3.361	ng	99
87) Benzo(b)fluoranthene	15.23	252	115799	4.857	ng	# 94
88) Benzo(k)fluoranthene	15.23	252	115799	5.052	ng	# 97
89) Benzo(a)pyrene	15.62	252	63088	2.917	ng	# 93
91) Benzo(a,h,i)perylene	17.73	276	41050	2.253	ng	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF032018\
Data File : BF103803.D
Acq On : 20 Mar 2018 15:46
Operator : JU/SJ
Sample : J1821-04RE
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :

Quant Time: Mar 20 17:54:06 2018
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF031518.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 15 18:31:53 2018
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.97 min Scan# 830
Delta R.T. 0.00 min
Lab File: BF103803.D
Acq: 20 Mar 2018 15:46

Instrument :
BNA_F
ClientSampleId :

Tot Ion:152 Resp: 179419
Ion Ratio Lower Upper
152 100
150 155.3 124.6 186.8
115 63.8 50.1 75.1




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5




2-Fluorophenol
Concen: 6.734 ng
RT: 5.60 min Scan# 597
Delta R.T. 0.02 min
Lab File: BF103803.D
Acq: 20 Mar 2018 15:46

Tgt Ion:112 Resp: 79430
Ion Ratio Lower Upper
112 100
64 53.8 47.9 71.9
63 26.1 23.7 35.5



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BF1
Ion 63.00 (62.70 to 63.70): BF1





#7

Phenol-d6

Concen: 19.071 ng

RT: 6.58 min Scan# 764

Delta R.T. -0.01 min

Lab File: BF103803.D

Acq: 20 Mar 2018 15:46

Instrument :

BNA_F

ClientSampleId :





Tot Ion: 99 Resp: 275227

Ion Ratio Lower Upper

99 100

42 15.6 14.5 21.7

71 30.8 25.4 38.0



#19

n-Nitroso-di-n-propylamine

Concen: 7.223 ng

RT: 7.53 min Scan# 925

Delta R.T. 0.15 min

Lab File: BF103803.D

Acq: 20 Mar 2018 15:46

Tgt Ion: 70 Resp: 66806


Ion Ratio Lower Upper


70 100

42 44.4 47.5 71.3#

101 0.0 7.4 11.2#

130 2.1 16.6 25.0#









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.25 min Scan# 1048
Delta R.T. 0.00 min
Lab File: BF103803.D
Acq: 20 Mar 2018 15:46

Instrument :
BNA_F
ClientSampleId :

Tot Ion:136	Resp:	787399
Ion Ratio	Lower	Upper
136	100	
137	11.3	8.9 13.3
54	7.7	6.6 9.8
68	6.7	5.0 7.4




Abundance Ion 136.00 (135.70 to 136.70): E
Ion 137.00 (136.70 to 137.70): E
Ion 54.00 (53.70 to 54.70): BF1
Ion 68.00 (67.70 to 68.70): BF1






#23
Nitrobenzene-d5
Concen: 44.929 ng
RT: 7.53 min Scan# 925
Delta R.T. -0.01 min
Lab File: BF103803.D
Acq: 20 Mar 2018 15:46

Tgt Ion: 82	Resp:	507300
Ion Ratio	Lower	Upper
82	100	
128	45.3	36.1 54.1
54	44.8	41.4 62.2



Abundance Ion 82.00 (81.70 to 82.70): BF1
Ion 128.00 (127.70 to 128.70): E
Ion 54.00 (53.70 to 54.70): BF1





#25

Isophorone

Concen: 2.713 ng

RT: 7.62 min Scan# 941

Delta R.T. -0.16 min

Lab File: BF103803.D

Acq: 20 Mar 2018 15:46

Instrument :

BNA_F

ClientSampleId :


Tot Ion: 82 Resp: 70904

Ion Ratio Lower Upper

82 100

95 0.9 5.2 7.8#




138 45.7 14.9 22.3#



Abundance Ion 82.00 (81.70 to 82.70): BF1

Ion 95.00 (94.70 to 95.70): BF1

Ion 138.00 (137.70 to 138.70): BF1



#26

2-Nitrophenol

Concen: 9.728 ng

RT: 8.08 min Scan# 1019

Delta R.T. 0.22 min

Lab File: BF103803.D

Acq: 20 Mar 2018 15:46


Tgt Ion: 139 Resp: 1773

Ion Ratio Lower Upper

139 100

109 17.3 22.7 34.1#



65 0.0 40.5 60.7#




Abundance Ion 139.00 (138.70 to 139.70): BF1

Ion 109.00 (108.70 to 109.70): BF1

Ion 65.00 (64.70 to 65.70): BF1









#32
Benzoic acid
Concen: 10.017 ng
RT: 8.02 min Scan# 1009
Delta R.T. 0.02 min
Lab File: BF103803.D
Acq: 20 Mar 2018 15:46




Instrument :
BNA_F
ClientSampleId :


Tot Ion:122 Resp: 1117
Ion Ratio Lower Upper
122 100
105 0.0 101.1 141.1#
77 57.3 70.7 110.7#



#38
Acenaphthene-d10
Concen: 20.000 ng
RT: 10.01 min Scan# 1347
Delta R.T. 0.00 min
Lab File: BF103803.D
Acq: 20 Mar 2018 15:46

Tgt Ion:164 Resp: 345442
Ion Ratio Lower Upper
164 100
162 104.1 86.1 129.1
160 45.6 38.3 57.5









#41
 2,4,6-Tribromophenol
 Concen: 4.840 ng
 RT: 10.80 min Scan# 1481
 Delta R.T. 0.00 min
 Lab File: BF103803.D
 Acq: 20 Mar 2018 15:46

Instrument :
 BNA_F
 ClientSampled :

Tot Ion	330	Resp	14375
Ion Ratio	100	Lower	Upper
332	98.7	77.0	115.6
141	34.8	29.9	44.9




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E






#44
 2-Fluorobiphenyl
 Concen: 43.626 ng
 RT: 9.33 min Scan# 1231
 Delta R.T. 0.00 min
 Lab File: BF103803.D
 Acq: 20 Mar 2018 15:46

Tgt Ion	172	Resp	944764
Ion Ratio	100	Lower	Upper
171	36.4	29.7	44.5
170	24.0	19.6	29.4



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E





#47

2-Nitroaniline

Concen: 6.847 ng

RT: 9.55 min Scan# 1269

Delta R.T. 0.01 min

Lab File: BF103803.D

Acq: 20 Mar 2018 15:46

Instrument :

BNA_F

ClientSampled :





Tot Ion: 65 Resp: 229

Ion Ratio Lower Upper

65 100

92 0.0 55.8 83.6#

138 0.0 91.2 136.8#



#50

2,6-Dinitrotoluene

Concen: 4.663 ng

RT: 9.75 min Scan# 1303

Delta R.T. -0.03 min

Lab File: BF103803.D

Acq: 20 Mar 2018 15:46




Tgt Ion: 165 Resp: 3778


Ion Ratio Lower Upper

165 100

63 0.0 44.7 67.1#

89 0.0 44.0 66.0#









#52
3-Nitroaniline
Concen: 3.862 ng
RT: 9.96 min Scan# 1338
Delta R.T. -0.01 min
Lab File: BF103803.D
Acq: 20 Mar 2018 15:46




Instrument :
BNA_F
ClientSampleId :


Tot Ion:138 Resp: 314
Ion Ratio Lower Upper
138 100
108 0.0 9.4 14.0#
92 0.0 89.4 134.2#



#55
4-Nitrophenol
Concen: 5.003 ng
RT: 10.10 min Scan# 1362
Delta R.T. 0.00 min
Lab File: BF103803.D
Acq: 20 Mar 2018 15:46

Tgt Ion:139 Resp: 170
Ion Ratio Lower Upper
139 100
109 369.3 48.4 88.4#
65 136.5 72.6 112.6#






#56
 2,4-Dinitrotoluene
 Concen: 12.409 ng
 RT: 10.01 min Scan# 1347
 Delta R.T. -0.18 min
 Lab File: BF103803.D
 Acq: 20 Mar 2018 15:46

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion	Ratio	Lower	Upper
165	100		
63	1.4	36.4	54.6#
89	1.6	57.8	86.6#
182	0.0	1.6	2.4#






Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BF1


Ion 89.00 (88.70 to 89.70): BF1

Ion 182.00 (181.70 to 182.70): E



#61
 4-Nitroaniline
 Concen: 3.795 ng
 RT: 10.56 min Scan# 1440
 Delta R.T. -0.02 min
 Lab File: BF103803.D
 Acq: 20 Mar 2018 15:46



Tgt Ion	Ratio	Lower	Upper
138	100		
92	70.1	30.2	70.2
108	0.0	52.5	92.5#




Abundance Ion 138.00 (137.70 to 138.70): E

Ion 92.00 (91.70 to 92.70): BF1

Ion 108.00 (107.70 to 108.70): E









#63
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.50 min Scan# 1601
Delta R.T. 0.01 min
Lab File: BF103803.D
Acq: 20 Mar 2018 15:46

Instrument :
BNA_F
ClientSampleId :

Tot Ion:188 Resp: 611077
Ion Ratio Lower Upper
188 100
94 11.4 8.5 12.7
80 12.0 9.2 13.8




Abundance Ion 188.00 (187.70 to 188.70): E
Ion 94.00 (93.70 to 94.70): BF1
Ion 80.00 (79.70 to 80.70): BF1






#70
Phenanthrene
Concen: 2.123 ng
RT: 11.53 min Scan# 1605
Delta R.T. 0.00 min
Lab File: BF103803.D
Acq: 20 Mar 2018 15:46

Tgt Ion:178 Resp: 70972
Ion Ratio Lower Upper
178 100
176 18.9 15.8 23.8
179 15.7 13.0 19.6



Abundance Ion 178.00 (177.70 to 178.70): E
Ion 176.00 (175.70 to 176.70): E
Ion 179.00 (178.70 to 179.70): E





#73

Di-n-butylphthalate

Concen: 29.660 ng

RT: 12.06 min Scan# 1695

Delta R.T. 0.01 min

Lab File: BF103803.D

Acq: 20 Mar 2018 15:46

Instrument :

BNA_F

ClientSampleId :


Tot Ion:149 Resp: 1174384

Ion Ratio Lower Upper

149 100

150 9.4 7.8 11.6




104 4.8 3.8 5.8



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E



#74

Fluoranthene

Concen: 4.772 ng

RT: 12.73 min Scan# 1809

Delta R.T. 0.01 min

Lab File: BF103803.D

Acq: 20 Mar 2018 15:46


Tgt Ion:202 Resp: 164353

Ion Ratio Lower Upper

202 100

101 12.7 0.0 34.1



203 16.8 0.0 38.1




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E





#75

Chrysene-d12

Concen: 20.000 ng

RT: 14.16 min Scan# 2052

Delta R.T. 0.01 min

Lab File: BF103803.D

Acq: 20 Mar 2018 15:46

Instrument :

BNA_F

ClientSampleId :


Tot Ion:240 Resp: 476556

Ion Ratio Lower Upper

240 100

120 13.6 9.5 14.3




236 26.2 20.7 31.1



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#77

Pyrene

Concen: 4.641 ng

RT: 12.95 min Scan# 1847

Delta R.T. 0.00 min

Lab File: BF103803.D

Acq: 20 Mar 2018 15:46


Tgt Ion:202 Resp: 162427

Ion Ratio Lower Upper

202 100

200 20.7 17.4 26.2



203 17.8 14.3 21.5




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E









#78
 Terphenyl-d14
 Concen: 56.051 ng
 RT: 13.09 min Scan# 1871
 Delta R.T. 0.01 min
 Lab File: BF103803.D
 Acq: 20 Mar 2018 15:46




Instrument :
 BNA_F
 ClientSampled :


Tot Ion:244 Resp: 1150796
 Ion Ratio Lower Upper
 244 100
 212 6.9 5.4 8.0
 122 13.2 11.0 16.4



#80
 Benzo(a)anthracene
 Concen: 2.549 ng
 RT: 14.18 min Scan# 2056
 Delta R.T. 0.04 min
 Lab File: BF103803.D
 Acq: 20 Mar 2018 15:46

Tgt Ion:228 Resp: 76894
 Ion Ratio Lower Upper
 228 100
 226 29.5 22.6 33.8
 229 21.8 15.8 23.6





#82

Chrysene

Concen: 2.674 ng

RT: 14.18 min Scan# 2056

Delta R.T. 0.00 min

Lab File: BF103803.D

Acq: 20 Mar 2018 15:46

Instrument :

BNA_F

ClientSampled :


Tot Ion:228 Resp: 76894

Ion Ratio Lower Upper

228 100

226 29.5 25.0 37.6




229 21.8 16.2 24.4



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#83

Bis(2-ethylhexyl)phthalate

Concen: 3.361 ng

RT: 14.12 min Scan# 2045

Delta R.T. 0.00 min

Lab File: BF103803.D

Acq: 20 Mar 2018 15:46


Tgt Ion:149 Resp: 74456

Ion Ratio Lower Upper

149 100

167 26.2 21.3 31.9



279 3.5 3.0 4.4




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E









#86
Pervlene-d12
Concen: 20.000 ng
RT: 15.69 min Scan# 2313
Delta R.T. 0.01 min
Lab File: BF103803.D
Acq: 20 Mar 2018 15:46




Instrument :
BNA_F
ClientSampleId :


Tot Ion:264 Resp: 377397
Ion Ratio Lower Upper
264 100
260 23.8 19.2 28.8
265 21.6 17.5 26.3



#87
Benzo(b)fluoranthene
Concen: 4.857 ng
RT: 15.23 min Scan# 2235
Delta R.T. 0.01 min
Lab File: BF103803.D
Acq: 20 Mar 2018 15:46

Tgt Ion:252 Resp: 115799
Ion Ratio Lower Upper
252 100
253 22.3 17.0 25.6
125 16.1 9.0 13.6#









#88
Benzo(k)fluoranthene
Concen: 5.052 ng
RT: 15.23 min Scan# 2235
Delta R.T. -0.02 min
Lab File: BF103803.D
Acq: 20 Mar 2018 15:46




Instrument :
BNA_F
ClientSampleId :


Tot Ion:252 Resp: 115799
Ion Ratio Lower Upper
252 100
253 22.3 17.9 26.9
125 16.1 10.2 15.2#



#89
Benzo(a)pyrene
Concen: 2.917 ng
RT: 15.62 min Scan# 2301
Delta R.T. 0.01 min
Lab File: BF103803.D
Acq: 20 Mar 2018 15:46

Tgt Ion:252 Resp: 63088
Ion Ratio Lower Upper
252 100
253 23.0 17.1 25.7
125 17.1 9.8 14.6#





#91

Benzo(a,h,i)perylene

Concen: 2.253 ng

RT: 17.73 min Scan# 2659

Delta R.T. 0.01 min

Lab File: BF103803.D

Acq: 20 Mar 2018 15:46

Instrument :

BNA_F

ClientSampleId :


Tot Ion: 276 Resp: 41050

Ion Ratio Lower Upper

276 100

277 24.3 17.9 26.9

138 30.6 21.8 32.8



Abundance

Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

