

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA F\DATA\BF032018\
 Data File : BF103813.D
 Acq On : 20 Mar 2018 20:10
 Operator : JU/SJ
 Sample : J1977-02
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SP

Manual Integrations
 APPROVED

Sohil
 3/22/2018 12:57:43 PM

Quant Time: Mar 21 01:49:00 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF031518.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 15 18:31:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.97	152	172059	20.00	ng	0.00
21) Naphthalene-d8	8.25	136	667541	20.00	ng	0.00
38) Acenaphthene-d10	10.02	164	294255	20.00	ng	0.00
63) Phenanthrene-d10	11.50	188	467707	20.00	ng	0.00
75) Chrysene-d12	14.16	240	291726	20.00	ng	0.00
86) Perylene-d12	15.70	264	254548	20.00	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.60	112	1045250	92.40	ng	0.02
7) Phenol-d6	6.59	99	1280729	92.54	ng	0.00
23) Nitrobenzene-d5	7.53	82	779698	81.45	ng	0.00
41) 2,4,6-Tribromophenol	10.80	330	239444	94.64	ng	0.00
44) 2-Fluorobiphenyl	9.33	172	1215619	65.90	ng	0.00
78) Terphenyl-d14	13.09	244	876572	69.75	ng	0.00
Target Compounds						Qvalue
49) Dimethylphthalate	9.72	163	158803	7.074	ng	100
70) Phenanthrene	11.53	178	64240	2.511	ng	97
74) Fluoranthene	12.72	202	91822	3.484	ng	98
77) Pyrene	12.96	202	78359	3.658	ng	98
80) Benzo(a)anthracene	14.14	228	38596	2.090	ng	99
87) Benzo(b)fluoranthene	15.24	252	39674m	2.467	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SVOASRV\HPCHEM1\BNA F\DATA\BF032018\
Data File : BF103813.D
Acq On : 20 Mar 2018 20:10
Operator : JU/SJ
Sample : J1977-02
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SP

Manual Integrations
APPROVED

Sohil
3/22/2018 12:57:43 PM

Quant Time: Mar 21 01:49:00 2018
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF031518.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 15 18:31:53 2018
Response via : Initial Calibration

