

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032024\
 Data File : BF137651.D
 Acq On : 20 Mar 2024 13:11
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 20 13:45:47 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 14 14:45:20 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	84	0.00
2	1,4-Dioxane	40.000	36.206	9.5	74	0.00
3	Pyridine	40.000	38.012	5.0	73	0.00
4	n-Nitrosodimethylamine	40.000	34.799	13.0	71	0.01
5 S	2-Fluorophenol	80.000	77.571	3.0	78	0.00
6	Aniline	40.000	36.640	8.4	73	0.00
7 S	Phenol-d6	80.000	72.846	8.9	76	0.00
8	2-Chlorophenol	40.000	37.554	6.1	78	0.00
9	Benzaldehyde	40.000	36.881	7.8	72	0.00
10 C	Phenol	40.000	36.706	8.2	75	0.00
11	bis(2-Chloroethyl)ether	40.000	35.575	11.1	71	0.00
12	1,3-Dichlorobenzene	40.000	37.676	5.8	78	0.00
13 C	1,4-Dichlorobenzene	40.000	37.675	5.8	78	0.00
14	1,2-Dichlorobenzene	40.000	37.752	5.6	79	0.00
15	Benzyl Alcohol	40.000	36.557	8.6	72	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	32.384	19.0	68	0.00
17	2-Methylphenol	40.000	37.603	6.0	78	0.00
18	Hexachloroethane	40.000	38.753	3.1	80	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	33.784	15.5	72	0.00
20	3+4-Methylphenols	40.000	36.433	8.9	76	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	80	0.00
22	Acetophenone	40.000	37.726	5.7	76	0.00
23 S	Nitrobenzene-d5	80.000	75.664	5.4	74	0.00
24	Nitrobenzene	40.000	38.133	4.7	75	0.00
25	Isophorone	40.000	36.235	9.4	72	0.00
26 C	2-Nitrophenol	40.000	40.128	-0.3	77	0.00
27	2,4-Dimethylphenol	40.000	38.968	2.6	76	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.088	4.8	75	0.00
29 C	2,4-Dichlorophenol	40.000	39.768	0.6	77	0.00
30	1,2,4-Trichlorobenzene	40.000	39.156	2.1	78	0.00
31	Naphthalene	40.000	37.982	5.0	77	0.00
32	Benzoic acid	40.000	33.879	15.3	71	0.00
33	4-Chloroaniline	40.000	38.961	2.6	74	0.00
34 C	Hexachlorobutadiene	40.000	39.384	1.5	78	0.00
35	Caprolactam	40.000	38.615	3.5	75	0.00
36 C	4-Chloro-3-methylphenol	40.000	38.053	4.9	74	0.00
37	2-Methylnaphthalene	40.000	37.706	5.7	75	0.00
38	1-Methylnaphthalene	40.000	36.842	7.9	74	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	78	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.418	1.5	76	0.00
41 P	Hexachlorocyclopentadiene	40.000	42.005	-5.0	88	0.00
42 S	2,4,6-Tribromophenol	80.000	80.805	-1.0	79	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.905	2.7	73	0.00
44	2,4,5-Trichlorophenol	40.000	40.428	-1.1	77	0.00
45 S	2-Fluorobiphenyl	80.000	75.255	5.9	75	0.00
46	1,1'-Biphenyl	40.000	38.315	4.2	74	0.00
47	2-Chloronaphthalene	40.000	38.282	4.3	74	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032024\
 Data File : BF137651.D
 Acq On : 20 Mar 2024 13:11
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 20 13:45:47 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 14 14:45:20 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.540	1.2	72	0.00
49	Acenaphthylene	40.000	37.722	5.7	73	0.00
50	Dimethylphthalate	40.000	37.907	5.2	74	0.00
51	2,6-Dinitrotoluene	40.000	39.693	0.8	75	0.00
52 C	Acenaphthene	40.000	37.145	7.1	73	0.00
53	3-Nitroaniline	40.000	40.066	-0.2	73	0.00
54 P	2,4-Dinitrophenol	40.000	36.260	9.4	71	0.00
55	Dibenzofuran	40.000	37.106	7.2	72	0.00
56 P	4-Nitrophenol	40.000	37.360	6.6	72	0.00
57	2,4-Dinitrotoluene	40.000	40.220	-0.5	74	0.00
58	Fluorene	40.000	36.729	8.2	73	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.708	3.2	74	0.00
60	Diethylphthalate	40.000	37.125	7.2	72	0.00
61	4-Chlorophenyl-phenylether	40.000	37.120	7.2	75	0.00
62	4-Nitroaniline	40.000	36.896	7.8	69	0.00
63	Azobenzene	40.000	35.442	11.4	68	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	74	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.402	-3.5	73	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.845	2.9	73	0.00
67	4-Bromophenyl-phenylether	40.000	40.285	-0.7	74	0.00
68	Hexachlorobenzene	40.000	40.977	-2.4	77	0.00
69	Atrazine	40.000	41.307	-3.3	76	0.00
70 C	Pentachlorophenol	40.000	35.488	11.3	63	0.00
71	Phenanthrene	40.000	37.974	5.1	73	0.00
72	Anthracene	40.000	38.317	4.2	73	0.00
73	Carbazole	40.000	38.006	5.0	71	0.00
74	Di-n-butylphthalate	40.000	39.009	2.5	73	0.00
75 C	Fluoranthene	40.000	37.857	5.4	73	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	89	0.00
77	Benzydine	40.000	32.155	19.6	62	0.00
78	Pyrene	40.000	34.407	14.0	73	0.00
79 S	Terphenyl-d14	80.000	69.564	13.0	76	0.00
80	Butylbenzylphthalate	40.000	49.070	-22.7	104	0.00
81	Benzo(a)anthracene	40.000	35.662	10.8	79	0.00
82	3,3'-Dichlorobenzidine	40.000	43.095	-7.7	86	0.00
83	Chrysene	40.000	36.262	9.3	79	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	48.795	-22.0	102	0.00
85 c	Di-n-octyl phthalate	40.000	40.733	-1.8	91	0.00
86 I	Perylene-d12	20.000	20.000	0.0	86	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	34.658	13.4	71	0.00
88	Benzo(b)fluoranthene	40.000	39.364	1.6	84	0.00
89	Benzo(k)fluoranthene	40.000	35.584	11.0	79	0.00
90 C	Benzo(a)pyrene	40.000	37.698	5.8	78	0.00
91	Dibenzo(a,h)anthracene	40.000	34.762	13.1	70	0.00
92	Benzo(g,h,i)perylene	40.000	33.939	15.2	69	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032024\
Data File : BF137651.D
Acq On : 20 Mar 2024 13:11
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Mar 20 13:45:47 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 14 14:45:20 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0