




Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032025\
Data File : BF142008.D
Acq On : 20 Mar 2025 12:04
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 20 12:27:18 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:46:22 2025
Response via : Initial Calibration



TIC: BF142008.D\data.ms

(70) Pentachlorophenol (C)




11.198min (-0.006) 32857.24 ng

response 267137

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.80	63.17
264.00	62.70	62.81
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032025\
Data File : BF142008.D
Acq On : 20 Mar 2025 12:04
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 20 12:27:18 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:46:22 2025
Response via : Initial Calibration



TIC: BF142008.D\data.ms

(77) Benzidine

12.733min (+ 0.000) 132001.90 ng

response 812179

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.40	13.98
183.00	12.10	12.02
0.00	0.00	0.00