



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032025\
Data File : BF142008.D
Acq On : 20 Mar 2025 12:04
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Mon Mar 10 15:46:22 2025



AutoFind: Scans 1624, 1625, 1626; Background Corrected with Scan 1617

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	27.6	61292	PASS
68	69	0.00	2	1.9	1250	PASS
69	198	0.00	100	30.0	66624	PASS
70	69	0.00	2	0.5	332	PASS
127	198	10	80	44.0	97883	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	222379	PASS
199	198	5	9	6.6	14698	PASS
275	198	10	60	31.5	70011	PASS
365	198	1	100	4.5	9948	PASS
441	198	0.01	100	20.0	44565	PASS
442	442	50	100	100.0	288896	PASS
443	442	15	24	19.5	56251	PASS

DDT Breakdown




Date	Instrument Name	DFTPP Data File
3/20/2025	BNA_F	<u>BF142008.D</u>
Compound Name	Response	Retention Time
DDT	601484	13.557
DDD	14312	13.257
DDE	1358	12.922
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
15670	617154	2.54

Instrument :
BNA_F
ClientSampleId :
DFTPP

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032025\
Data File : BF142008.D
Acq On : 20 Mar 2025 12:04
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Mar 20 12:27:18 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:46:22 2025
Response via : Initial Calibration



TIC: BF142008.D\data.ms

(70) Pentachlorophenol (C)

11.198min (-0.006) 32857.24 ng




response 267137

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.80	63.17
264.00	62.70	62.81
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032025\
Data File : BF142008.D
Acq On : 20 Mar 2025 12:04
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Mar 20 12:27:18 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:46:22 2025
Response via : Initial Calibration



TIC: BF142008.D\data.ms

(77) Benzidine

12.733min (+ 0.000) 132001.90 ng

response 812179

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.40	13.98
183.00	12.10	12.02
0.00	0.00	0.00