

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF032117\
 Data File : BF093793.D
 Acq On : 21 Mar 2017 12:57
 Operator : SJ/MA
 Sample : I2280-03DL 250X
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 CTMIX-2DL

Quant Time: Mar 21 14:43:52 2017
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF032017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 21 14:38:20 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.82	152	195468	20.00	ng	0.00
21) Naphthalene-d8	8.12	136	791734	20.00	ng	0.00
38) Acenaphthene-d10	9.86	164	376772	20.00	ng	0.00
63) Phenanthrene-d10	11.34	188	651055	20.00	ng	0.00
75) Chrysene-d12	13.97	240	487046	20.00	ng	0.00
86) Perylene-d12	15.37	264	333476	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.41	112	5862	0.52	ng	-0.01
7) Phenol-d6	6.45	99	7342	0.53	ng	-0.01
23) Nitrobenzene-d5	7.37	82	4054	0.29	ng	-0.02
41) 2,4,6-Tribromophenol	10.64	330	959	0.33	ng	-0.01
44) 2-Fluorobiphenyl	9.19	172	8544	0.38	ng	0.00
78) Terphenyl-d14	12.92	244	8959	0.38	ng	-0.01
Target Compounds						Qvalue
31) Naphthalene	8.14	128	1416321	38.54	ng	99
37) 2-Methylnaphthalene	8.82	142	159361	6.64	ng	97
48) Acenaphthylene	9.72	152	141755	3.71	ng	98
54) Dibenzofuran	10.06	168	141523	4.61	ng	96
57) Fluorene	10.40	166	161221	6.85	ng	99
70) Phenanthrene	11.36	178	639504	18.44	ng	99
71) Anthracene	11.41	178	178432	5.00	ng	98
72) Carbazole	11.57	167	72409	2.05	ng	# 95
74) Fluoranthene	12.54	202	404301	11.34	ng	94
77) Pyrene	12.77	202	332851	8.21	ng	99
80) Benzo(a)anthracene	13.96	228	120695m	3.96	ng	
82) Chrysene	13.99	228	92623m	3.04	ng	
87) Benzo(b)fluoranthene	14.97	252	82451m	3.67	ng	
89) Benzo(a)pyrene	15.31	252	65734	3.61	ng	96
91) Benzo(g,h,i)perylene	17.12	276	31923	2.12	ng	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF032117\
Data File : BF093793.D
Acq On : 21 Mar 2017 12:57
Operator : SJ/MA
Sample : I2280-03DL 250X
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
CTMIX-2DL

Quant Time: Mar 21 14:43:52 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF032017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 21 14:38:20 2017
Response via : Initial Calibration

