

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF032117\
 Data File : BF093814.D
 Acq On : 21 Mar 2017 22:35
 Operator : SJ/MA
 Sample : I2304-09
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SPA-B5(0-5)

Quant Time: Mar 22 03:21:30 2017
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF032017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 21 14:38:20 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.82	152	201094	20.00	ng	0.00
21) Naphthalene-d8	8.12	136	806875	20.00	ng	0.00
38) Acenaphthene-d10	9.86	164	351993	20.00	ng	0.00
63) Phenanthrene-d10	11.34	188	493087	20.00	ng	0.00
75) Chrysene-d12	13.97	240	323453	20.00	ng	0.00
86) Perylene-d12	15.38	264	309070	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.43	112	1501928	130.61	ng	0.01
7) Phenol-d6	6.47	99	2087064	146.88	ng	0.01
23) Nitrobenzene-d5	7.40	82	1257454	88.41	ng	0.00
41) 2,4,6-Tribromophenol	10.65	330	359870	132.00	ng	0.00
44) 2-Fluorobiphenyl	9.19	172	1711254	81.75	ng	0.00
78) Terphenyl-d14	12.93	244	998197	62.92	ng	0.00
Target Compounds						
10) Phenol	6.47	94	89971	5.63	ng	# 20
49) Dimethylphthalate	9.58	163	267242	10.67	ng	# 99
79) Butylbenzylphthalate	13.41	149	31749	2.64	ng	# 77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF032117\
Data File : BF093814.D
Acq On : 21 Mar 2017 22:35
Operator : SJ/MA
Sample : I2304-09
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SPA-B5(0-5)

Quant Time: Mar 22 03:21:30 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF032017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 21 14:38:20 2017
Response via : Initial Calibration

