

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF032117\
 Data File : BF093821.D
 Acq On : 22 Mar 2017 7:19
 Operator : SJ/MA
 Sample : I2323-02
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 031717-PRRT-E-U-M

Quant Time: Mar 22 07:42:58 2017
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF032017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 21 14:38:20 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.82	152	201981	20.00	ng	0.00
21) Naphthalene-d8	8.12	136	807639	20.00	ng	0.00
38) Acenaphthene-d10	9.86	164	380619	20.00	ng	0.00
63) Phenanthrene-d10	11.34	188	605886	20.00	ng	0.00
75) Chrysene-d12	13.97	240	392035	20.00	ng	0.00
86) Perylene-d12	15.37	264	334433	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.42	112	517289	44.79	ng	0.00
7) Phenol-d6	6.45	99	406480	28.48	ng	-0.01
23) Nitrobenzene-d5	7.38	82	1155745	81.18	ng	-0.01
41) 2,4,6-Tribromophenol	10.65	330	249354	84.58	ng	0.00
44) 2-Fluorobiphenyl	9.19	172	1848821	81.68	ng	0.00
78) Terphenyl-d14	12.93	244	1358512	70.65	ng	0.00
Target Compounds						
10) Phenol	6.46	94	54484	3.40	ng	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF032117\
Data File : BF093821.D
Acq On : 22 Mar 2017 7:19
Operator : SJ/MA
Sample : I2323-02
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
031717-PRRT-E-U-M

Quant Time: Mar 22 07:42:58 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF032017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 21 14:38:20 2017
Response via : Initial Calibration

