

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF032117\
 Data File : BF093824.D
 Acq On : 22 Mar 2017 8:41
 Operator : SJ/MA
 Sample : I2318-01DL 100X
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 HP-CT-00DL

Quant Time: Mar 22 14:35:47 2017
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF032017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 21 14:38:20 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.82	152	209478	20.00	ng	0.00
21) Naphthalene-d8	8.12	136	848119	20.00	ng	0.00
38) Acenaphthene-d10	9.86	164	387459	20.00	ng	0.00
63) Phenanthrene-d10	11.34	188	595590	20.00	ng	0.00
75) Chrysene-d12	13.97	240	378519	20.00	ng	0.00
86) Perylene-d12	15.37	264	367886	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.41	112	12474	1.04	ng	-0.01
7) Phenol-d6	6.45	99	15790	1.07	ng	-0.01
23) Nitrobenzene-d5	7.38	82	8883	0.59	ng	-0.01
41) 2,4,6-Tribromophenol	10.64	330	2124	0.71	ng	-0.01
44) 2-Fluorobiphenyl	9.19	172	22483	0.98	ng	0.00
78) Terphenyl-d14	12.92	244	15027	0.81	ng	-0.01
Target Compounds						Qvalue
31) Naphthalene	8.14	128	2806057	71.28	ng	98
37) 2-Methylnaphthalene	8.82	142	691759	26.92	ng	99
45) 1,1'-Biphenyl	9.28	154	178409	5.67	ng	95
48) Acenaphthylene	9.73	152	614054	15.63	ng	98
51) Acenaphthene	9.90	154	141339	6.02	ng	97
54) Dibenzofuran	10.07	168	235960	7.47	ng	# 90
57) Fluorene	10.41	166	609924	25.22	ng	99
70) Phenanthrene	11.37	178	2075065	65.42	ng	100
71) Anthracene	11.42	178	619607	18.97	ng	99
72) Carbazole	11.57	167	127980	3.96	ng	97
74) Fluoranthene	12.55	202	895250	27.46	ng	100
77) Pyrene	12.78	202	1200964	38.13	ng	100
80) Benzo(a)anthracene	13.96	228	379044	15.99	ng	# 74
82) Chrysene	13.99	228	261493m	11.04	ng	
85) Indeno(1,2,3-cd)pyrene	16.72	276	93337	4.81	ng	94
87) Benzo(b)fluoranthene	14.97	252	228663m	9.22	ng	
88) Benzo(k)fluoranthene	14.99	252	114891m	6.27	ng	
89) Benzo(a)pyrene	15.32	252	267455	13.30	ng	97
91) Benzo(g,h,i)perylene	17.13	276	110091	6.64	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF032117\
Data File : BF093824.D
Acq On : 22 Mar 2017 8:41
Operator : SJ/MA
Sample : I2318-01DL 100X
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
HP-CT-00DL

Quant Time: Mar 22 14:35:47 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF032017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 21 14:38:20 2017
Response via : Initial Calibration

