

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF032117\
 Data File : BF093826.D
 Acq On : 22 Mar 2017 9:35
 Operator : SJ/MA
 Sample : I2318-05DL 100X
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 HP-CT-02DL

Quant Time: Mar 22 14:47:07 2017
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF032017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 21 14:38:20 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.82	152	201061	20.00	ng	0.00
21) Naphthalene-d8	8.12	136	828846	20.00	ng	0.00
38) Acenaphthene-d10	9.86	164	371209	20.00	ng	0.00
63) Phenanthrene-d10	11.34	188	577257	20.00	ng	0.00
75) Chrysene-d12	13.97	240	383662	20.00	ng	0.00
86) Perylene-d12	15.37	264	354149	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.41	112	13187	1.15	ng	-0.01
7) Phenol-d6	6.45	99	17334	1.22	ng	-0.01
23) Nitrobenzene-d5	7.38	82	9832	0.67	ng	-0.01
41) 2,4,6-Tribromophenol	10.64	330	1846	0.64	ng	-0.01
44) 2-Fluorobiphenyl	9.19	172	22046	1.00	ng	0.00
78) Terphenyl-d14	12.92	244	14553	0.77	ng	-0.01
Target Compounds						Qvalue
31) Naphthalene	8.14	128	1526878	39.69	ng	99
37) 2-Methylnaphthalene	8.82	142	431704	17.19	ng	98
45) 1,1'-Biphenyl	9.28	154	93894	3.11	ng	96
48) Acenaphthylene	9.72	152	296955	7.89	ng	98
51) Acenaphthene	9.90	154	91244	4.06	ng	96
54) Dibenzofuran	10.06	168	108273	3.58	ng	# 91
57) Fluorene	10.40	166	325741	14.06	ng	100
70) Phenanthrene	11.36	178	1148010	37.34	ng	99
71) Anthracene	11.42	178	326215	10.31	ng	97
74) Fluoranthene	12.55	202	655222	20.74	ng	95
77) Pyrene	12.78	202	782053	24.49	ng	99
80) Benzo(a)anthracene	13.96	228	232568m	9.68	ng	
82) Chrysene	13.99	228	229031m	9.54	ng	
85) Indeno(1,2,3-cd)pyrene	16.72	276	73287	3.73	ng	95
87) Benzo(b)fluoranthene	14.97	252	188891m	7.91	ng	
88) Benzo(k)fluoranthene	14.99	252	84718m	4.81	ng	
89) Benzo(a)pyrene	15.32	252	186024	9.61	ng	97
91) Benzo(g,h,i)perylene	17.12	276	81498	5.11	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF032117\
Data File : BF093826.D
Acq On : 22 Mar 2017 9:35
Operator : SJ/MA
Sample : I2318-05DL 100X
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
HP-CT-02DL

Quant Time: Mar 22 14:47:07 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF032017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 21 14:38:20 2017
Response via : Initial Calibration

