

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032125\
 Data File : BF142023.D
 Acq On : 21 Mar 2025 11:23
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 21 20:15:03 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 10 15:46:22 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.00
2	1,4-Dioxane	0.502	0.476	5.2	83	-0.05
3	Pyridine	1.232	1.117	9.3	80	-0.04
4	n-Nitrosodimethylamine	0.587	0.528	10.1	79	-0.04
5 S	2-Fluorophenol	1.198	1.142	4.7	87	0.00
6	Aniline	1.505	1.312	12.8	78	0.00
7 S	Phenol-d6	1.526	1.389	9.0	84	0.00
8	2-Chlorophenol	1.329	1.242	6.5	86	0.00
9	Benzaldehyde	0.853	0.928	-8.8	105	0.00
10 C	Phenol	1.605	1.463	8.8	83	0.00
11	bis(2-Chloroethyl)ether	1.205	1.083	10.1	83	0.00
12	1,3-Dichlorobenzene	1.435	1.368	4.7	87	0.00
13 C	1,4-Dichlorobenzene	1.452	1.373	5.4	86	0.00
14	1,2-Dichlorobenzene	1.367	1.291	5.6	87	0.00
15	Benzyl Alcohol	1.233	1.120	9.2	83	0.00
16	2,2'-oxybis(1-Chloropropane	1.455	1.187	18.4	76	0.00
17	2-Methylphenol	1.057	0.948	10.3	82	0.00
18	Hexachloroethane	0.544	0.510	6.3	86	0.00
19 P	n-Nitroso-di-n-propylamine	0.980	0.807	17.7	77	0.00
20	3+4-Methylphenols	1.354	1.195	11.7	81	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	82	0.00
22	Acetophenone	0.479	0.451	5.8	81	0.00
23 S	Nitrobenzene-d5	0.355	0.346	2.5	82	0.00
24	Nitrobenzene	0.353	0.341	3.4	81	0.00
25	Isophorone	0.628	0.566	9.9	78	0.00
26 C	2-Nitrophenol	0.173	0.182	-5.2	86	0.00
27	2,4-Dimethylphenol	0.239	0.221	7.5	79	0.00
28	bis(2-Chloroethoxy)methane	0.394	0.360	8.6	80	0.00
29 C	2,4-Dichlorophenol	0.296	0.287	3.0	83	0.00
30	1,2,4-Trichlorobenzene	0.327	0.320	2.1	85	0.00
31	Naphthalene	1.027	0.972	5.4	81	0.00
32	Benzoic acid	0.210	0.218	-3.8	86	0.00
33	4-Chloroaniline	0.363	0.326	10.2	77	0.00
34 C	Hexachlorobutadiene	0.213	0.214	-0.5	87	0.00
35	Caprolactam	0.090	0.078	13.3	76	0.00
36 C	4-Chloro-3-methylphenol	0.331	0.308	6.9	80	0.00
37	2-Methylnaphthalene	0.687	0.632	8.0	80	0.00
38	1-Methylnaphthalene	0.663	0.612	7.7	81	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	79	0.00
40	1,2,4,5-Tetrachlorobenzene	0.609	0.617	-1.3	82	0.00
41 P	Hexachlorocyclopentadiene	0.238	0.232	2.5	73	0.00
42 S	2,4,6-Tribromophenol	0.254	0.240	5.5	76	0.00
43 C	2,4,6-Trichlorophenol	0.398	0.384	3.5	76	0.00
44	2,4,5-Trichlorophenol	0.403	0.402	0.2	81	0.00
45 S	2-Fluorobiphenyl	1.315	1.274	3.1	80	0.00
46	1,1'-Biphenyl	1.520	1.466	3.6	79	0.00
47	2-Chloronaphthalene	1.133	1.096	3.3	79	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032125\
 Data File : BF142023.D
 Acq On : 21 Mar 2025 11:23
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 21 20:15:03 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 10 15:46:22 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.328	0.307	6.4	75	0.00
49	Acenaphthylene	1.681	1.585	5.7	76	0.00
50	Dimethylphthalate	1.401	1.271	9.3	75	0.00
51	2,6-Dinitrotoluene	0.292	0.277	5.1	76	0.00
52 C	Acenaphthene	1.181	1.149	2.7	79	0.00
53	3-Nitroaniline	0.296	0.260	12.2	70	0.00
54 P	2,4-Dinitrophenol	0.139	0.148	-6.5	82	0.00
55	Dibenzofuran	1.688	1.580	6.4	77	0.00
56 P	4-Nitrophenol	0.223	0.197	11.7	68	0.00
57	2,4-Dinitrotoluene	0.381	0.361	5.2	76	0.00
58	Fluorene	1.302	1.197	8.1	77	0.00
59	2,3,4,6-Tetrachlorophenol	0.367	0.351	4.4	77	0.00
60	Diethylphthalate	1.397	1.244	11.0	73	0.00
61	4-Chlorophenyl-phenylether	0.679	0.632	6.9	77	0.00
62	4-Nitroaniline	0.288	0.239	17.0	65	0.00
63	Azobenzene	1.298	1.128	13.1	71	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	70	0.00
65	4,6-Dinitro-2-methylphenol	0.119	0.132	-10.9	76	0.00
66 c	n-Nitrosodiphenylamine	0.647	0.656	-1.4	74	0.00
67	4-Bromophenyl-phenylether	0.251	0.258	-2.8	74	0.00
68	Hexachlorobenzene	0.275	0.289	-5.1	75	0.00
69	Atrazine	0.198	0.199	-0.5	85	0.00
70 C	Pentachlorophenol	0.170	0.178	-4.7	70	0.00
71	Phenanthrene	1.080	1.039	3.8	70	0.00
72	Anthracene	1.084	1.041	4.0	70	0.00
73	Carbazole	0.935	0.837	10.5	65	0.00
74	Di-n-butylphthalate	1.240	1.097	11.5	65	0.00
75 C	Fluoranthene	1.146	0.973	15.1	62	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	66	0.00
77	Benzydine	0.279	0.339	-21.5	62	0.00
78	Pyrene	1.730	1.511	12.7	62	0.00
79 S	Terphenyl-d14	1.353	1.155	14.6	60	0.00
80	Butylbenzylphthalate	0.677	0.571	15.7	58	0.00
81	Benzo(a)anthracene	1.312	1.222	6.9	63	0.00
82	3,3'-Dichlorobenzidine	0.379	0.397	-4.7	70	0.00
83	Chrysene	1.190	1.160	2.5	67	0.00
84	Bis(2-ethylhexyl)phthalate	0.934	0.796	14.8	58	0.00
85 c	Di-n-octyl phthalate	1.299	1.362	-4.8	71	0.00
86 I	Perylene-d12	1.000	1.000	0.0	92	0.00
87	Indeno(1,2,3-cd)pyrene	1.294	1.154	10.8	83	0.00
88	Benzo(b)fluoranthene	1.371	1.129	17.7	86	0.00
89	Benzo(k)fluoranthene	1.173	0.935	20.3	72	0.00
90 C	Benzo(a)pyrene	1.073	0.998	7.0	89	0.00
91	Dibenzo(a,h)anthracene	1.067	0.928	13.0	80	0.00
92	Benzo(g,h,i)perylene	1.055	0.928	12.0	81	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032125\
Data File : BF142023.D
Acq On : 21 Mar 2025 11:23
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Mar 21 20:15:03 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:46:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0