

Data Path : Z:\HPCHEM1\BNA F\DATA\BF032217\
 Data File : BF093849.D
 Acq On : 22 Mar 2017 23:02
 Operator : SJ/MA
 Sample : I2296-06
 Misc : L00 20 PPB
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 LOQ-WATER-06-QT2-2017

Quant Time: Mar 23 01:23:48 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF032217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 22 17:41:33 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	5.99	152	195392	20.00	ng	0.00
21) Naphthalene-d8	7.49	136	790664	20.00	ng	0.00
38) Acenaphthene-d10	9.63	164	356263	20.00	ng	0.00
63) Phenanthrene-d10	11.46	188	613127	20.00	ng	0.00
75) Chrysene-d12	14.71	240	504924	20.00	ng	0.00
86) Perylene-d12	16.35	264	399468	20.00	ng	0.00

System Monitoring Compounds						
5) 2-Fluorophenol	4.51	112	1408295	121.74	ng	0.00
7) Phenol-d6	5.57	99	1724880	120.53	ng	0.00
23) Nitrobenzene-d5	6.64	82	1340641	96.77	ng	0.00
41) 2,4,6-Tribromophenol	10.61	330	371292	131.89	ng	0.00
44) 2-Fluorobiphenyl	8.82	172	2248044	96.25	ng	0.00
78) Terphenyl-d14	13.44	244	1957358	86.89	ng	0.00

Target Compounds						Qvalue
2) 1,4-Dioxane	2.37	88	113104	20.35	ng	98
3) Pyridine	2.86	79	314668	20.77	ng	99
4) n-Nitrosodimethylamine	2.80	42	129844	20.83	ng	92
6) Aniline	5.61	93	403583	20.27	ng	98
8) 2-Chlorophenol	5.75	128	258173	20.30	ng	99
9) Benzaldehyde	5.48	77	203888	21.10	ng	99
10) Phenol	5.59	94	356488	21.89	ng	90
11) bis(2-Chloroethyl)ether	5.69	93	270669	21.27	ng	99
12) 1,3-Dichlorobenzene	5.92	146	305253	21.40	ng	99
13) 1,4-Dichlorobenzene	6.00	146	310753	21.51	ng	97
14) 1,2-Dichlorobenzene	6.19	146	284761	21.40	ng	95
15) Benzyl Alcohol	6.15	79	228206	21.79	ng	96
16) 2,2'-oxybis(1-Chloropropan	6.31	45	414286	21.12	ng	99
17) 2-Methylphenol	6.27	107	223903	20.56	ng	97
18) Hexachloroethane	6.58	117	109646	21.59	ng	# 84
19) n-Nitroso-di-n-propylamine	6.47	70	197809	21.51	ng	97
20) 3+4-Methylphenols	6.46	107	279881	21.22	ng	91
22) Acetophenone	6.46	105	382495	21.71	ng	# 94
24) Nitrobenzene	6.66	77	290922	21.29	ng	95
25) Isophorone	6.94	82	509823	20.94	ng	97
26) 2-Nitrophenol	7.03	139	137693	21.13	ng	95
27) 2,4-Dimethylphenol	7.09	122	235816	21.00	ng	97
28) bis(2-Chloroethoxy)methane	7.20	93	342693	21.35	ng	98
29) 2,4-Dichlorophenol	7.32	162	221271	21.08	ng	98
30) 1,2,4-Trichlorobenzene	7.42	180	233537	21.60	ng	98
31) Naphthalene	7.52	128	825152	21.70	ng	100
32) Benzoic acid	7.20	122	137883	18.45	ng	97
33) 4-Chloroaniline	7.57	127	319029	20.70	ng	95
34) Hexachlorobutadiene	7.67	225	120030	21.65	ng	97
35) Caprolactam	8.00	113	68652	20.51	ng	99
36) 4-Chloro-3-methylphenol	8.17	107	229061	20.88	ng	87
37) 2-Methylnaphthalene	8.36	142	553443	21.84	ng	98
39) 1,2,4,5-Tetrachlorobenzene	8.56	216	215463	22.11	ng	100
40) Hexachlorocyclopentadiene	8.56	237	99584	21.84	ng	100

Data Path : Z:\HPCHEM1\BNA F\DATA\BF032217\
 Data File : BF093849.D
 Acq On : 22 Mar 2017 23:02
 Operator : SJ/MA
 Sample : I2296-06
 Misc : L00 20 PPB
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 LOQ-WATER-06-QT2-2017

Quant Time: Mar 23 01:23:48 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF032217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 22 17:41:33 2017
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
42) 2,4,6-Trichlorophenol	8.70	196	146041	21.18	nq	97
43) 2,4,5-Trichlorophenol	8.75	196	157206	21.07	nq	94
45) 1,1'-Biphenyl	8.93	154	639006	22.24	nq	97
46) 2-Chloronaphthalene	8.96	162	494420	21.98	nq	96
47) 2-Nitroaniline	9.07	65	137360	20.58	nq	91
48) Acenaphthylene	9.46	152	826293	22.10	nq	99
49) Dimethylphthalate	9.32	163	548312	21.65	nq	99
50) 2,6-Dinitrotoluene	9.37	165	124818	21.50	nq	# 83
51) Acenaphthene	9.67	154	472275	21.65	nq	99
52) 3-Nitroaniline	9.57	138	139369	20.44	nq	93
53) 2,4-Dinitrophenol	9.70	184	48098	19.24	nq	# 72
54) Dibenzofuran	9.89	168	660924	22.26	nq	99
55) 4-Nitrophenol	9.78	139	109232	20.46	nq	# 70
56) 2,4-Dinitrotoluene	9.87	165	162439	22.27	nq	# 85
57) Fluorene	10.31	166	535398	21.74	nq	98
58) 2,3,4,6-Tetrachlorophenol	10.03	232	109893	21.47	nq	99
59) Diethylphthalate	10.19	149	546940	21.85	nq	99
60) 4-Chlorophenyl-phenylether	10.32	204	233507	22.26	nq	90
61) 4-Nitroaniline	10.33	138	130268	19.91	nq	96
62) Azobenzene	10.51	77	514454	21.73	nq	97
64) 4,6-Dinitro-2-methylphenol	10.37	198	71961	19.16	nq	81
65) n-Nitrosodiphenylamine	10.46	169	466501	22.00	nq	98
66) 4-Bromophenyl-phenylether	10.92	248	130747	21.68	nq	99
67) Hexachlorobenzene	10.99	284	134121	21.97	nq	# 87
68) Atrazine	11.13	200	137470	21.65	nq	96
69) Pentachlorophenol	11.22	266	76690	20.18	nq	99
70) Phenanthrene	11.49	178	760455	22.30	nq	99
71) Anthracene	11.55	178	781528	22.25	nq	99
72) Carbazole	11.74	167	776225	21.56	nq	98
73) Di-n-butylphthalate	12.20	149	855592	22.85	nq	99
74) Fluoranthene	12.95	202	763137	21.85	nq	99
76) Benzidine	13.12	184	389560	19.49	nq	# 92
77) Pyrene	13.23	202	793579	21.66	nq	99
79) Butylbenzylphthalate	14.04	149	339745	21.76	nq	# 86
80) Benzo(a)anthracene	14.70	228	629990	21.40	nq	99
81) 3,3'-Dichlorobenzidine	14.67	252	215012	20.43	nq	# 95
82) Chrysene	14.74	228	609409	22.30	nq	99
83) Bis(2-ethylhexyl)phthalate	14.76	149	503091	23.62	nq	# 99
84) Di-n-octyl phthalate	15.52	149	786316	22.10	nq	99
85) Indeno(1,2,3-cd)pyrene	17.73	276	445702	18.83	nq	97
87) Benzo(b)fluoranthene	15.93	252	528406	21.58	nq	100
88) Benzo(k)fluoranthene	15.96	252	531238	22.01	nq	# 95
89) Benzo(a)pyrene	16.29	252	480244	21.30	nq	99
90) Dibenzo(a,h)anthracene	17.76	278	381010	19.95	nq	96
91) Benzo(g,h,i)perylene	18.14	276	370801	19.27	nq	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

```
Data Path : Z:\HPCHEM1\BNA F\DATA\BF032217\  
Data File : BF093849.D  
Acq On    : 22 Mar 2017  23:02  
Operator  : SJ/MA  
Sample    : I2296-06  
Misc      : L00 20 PPB  
ALS Vial  : 21      Sample Multiplier: 1
```

Instrument :
BNA_F
ClientSampleId :
LOO-WATER-06-OT2-2017

Quant Time: Mar 23 01:23:48 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF032217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 22 17:41:33 2017
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 20.00 ng

RT: 5.99 min Scan# 697

Delta R.T. -0.00 min

Lab File: BF093849.D

Acq: 22 Mar 2017 23:02

Instrument :

BNA_F

ClientSampleId :

LOQ-WATER-06-QT2-2017


Tot Ion:152 Resp: 195392

Ion Ratio Lower Upper

152 100

150 153.9 126.2 189.2




115 60.2 52.2 78.2



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane

Concen: 20.35 ng

RT: 2.37 min Scan# 82

Delta R.T. -0.00 min

Lab File: BF093849.D

Acq: 22 Mar 2017 23:02


Tgt Ion: 88 Resp: 113104

Ion Ratio Lower Upper

88 100

58 75.0 61.4 92.0



43 27.7 23.4 35.0




Abundance Ion 88.00 (87.70 to 88.70): BF0

Ion 58.00 (57.70 to 58.70): BF0

Ion 43.00 (42.70 to 43.70): BF0





#3

Pvridine

Concen: 20.77 ng

RT: 2.86 min Scan# 166

Delta R.T. 0.01 min

Lab File: BF093849.D

Acq: 22 Mar 2017 23:02

Instrument :

BNA_F

ClientSampleId :

LOQ-WATER-06-QT2-2017


Tot Ion: 79 Resp: 314668

Ion Ratio Lower Upper

79 100

52 69.5 54.8 82.2




51 34.3 27.0 40.4



Abundance Ion 79.00 (78.70 to 79.70): BF0

Ion 52.00 (51.70 to 52.70): BF0

Ion 51.00 (50.70 to 51.70): BF0



#4

n-Nitrosodimethylamine

Concen: 20.83 ng

RT: 2.80 min Scan# 156

Delta R.T. -0.01 min

Lab File: BF093849.D

Acq: 22 Mar 2017 23:02


Tgt Ion: 42 Resp: 129844

Ion Ratio Lower Upper

42 100

74 139.9 103.9 155.9



44 7.1 5.7 8.5



Abundance Ion 42.00 (41.70 to 42.70): BF0

Ion 74.00 (73.70 to 74.70): BF0

Ion 44.00 (43.70 to 44.70): BF0





#5

2-Fluorophenol

Concen: 121.74 ng

RT: 4.51 min Scan# 446

Delta R.T. 0.01 min

Lab File: BF093849.D

Acq: 22 Mar 2017 23:02

Instrument :

BNA_F

ClientSampleId :

LOQ-WATER-06-QT2-2017


Tot Ion:112 Resp: 1408295

Ion Ratio Lower Upper

112 100

64 56.3 46.0 69.0




63 28.7 23.4 35.0



Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BF0

Ion 63.00 (62.70 to 63.70): BF0



#6

Aniline

Concen: 20.27 ng

RT: 5.61 min Scan# 633

Delta R.T. -0.00 min

Lab File: BF093849.D

Acq: 22 Mar 2017 23:02


Tgt Ion: 93 Resp: 403583

Ion Ratio Lower Upper

93 100

66 39.1 32.9 49.3



65 19.5 16.0 24.0




Abundance Ion 93.00 (92.70 to 93.70): BF0

Ion 66.00 (65.70 to 66.70): BF0

Ion 65.00 (64.70 to 65.70): BF0





#7

Phenol-d6

Concen: 120.53 ng

RT: 5.57 min Scan# 627

Delta R.T. 0.01 min

Lab File: BF093849.D

Acq: 22 Mar 2017 23:02

Instrument :

BNA_F

ClientSampleId :

LOQ-WATER-06-QT2-2017


Tot Ion: 99 Resp: 1724880

Ion	Ratio	Lower	Upper
99	100		
42	18.3	13.5	20.3
71	33.2	24.5	36.7

99 100

42 18.3 13.5 20.3




71 33.2 24.5 36.7



Abundance Ion 99.00 (98.70 to 99.70): BF0

Ion 42.00 (41.70 to 42.70): BF0

Ion 71.00 (70.70 to 71.70): BF0



#8

2-Chlorophenol

Concen: 20.30 ng

RT: 5.75 min Scan# 656

Delta R.T. -0.00 min

Lab File: BF093849.D

Acq: 22 Mar 2017 23:02


Tgt Ion:128 Resp: 258173

Ion	Ratio	Lower	Upper
128	100		
130	33.6	12.9	52.9
64	47.4	28.4	68.4

128 100

130 33.6 12.9 52.9



64 47.4 28.4 68.4




Abundance Ion 128.00 (127.70 to 128.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 64.00 (63.70 to 64.70): BF0





#9

Benzaldehyde

Concen: 21.10 ng

RT: 5.48 min Scan# 611

Delta R.T. -0.00 min

Lab File: BF093849.D

Acq: 22 Mar 2017 23:02

Instrument :

BNA_F

ClientSampleId :

LOQ-WATER-06-QT2-2017


Tot Ion: 77 Resp: 203888

Ion Ratio Lower Upper

77 100

105 101.7 83.8 123.8




106 99.3 79.1 119.1



Abundance Ion 77.00 (76.70 to 77.70): BF0

Ion 105.00 (104.70 to 105.70): BF0

Ion 106.00 (105.70 to 106.70): BF0



#10

Phenol

Concen: 21.89 ng

RT: 5.59 min Scan# 629

Delta R.T. -0.00 min

Lab File: BF093849.D

Acq: 22 Mar 2017 23:02


Tgt Ion: 94 Resp: 356488

Ion Ratio Lower Upper

94 100

65 26.6 9.9 49.9



66 35.1 23.1 63.1




Abundance Ion 94.00 (93.70 to 94.70): BF0

Ion 65.00 (64.70 to 65.70): BF0

Ion 66.00 (65.70 to 66.70): BF0









#11
bis(2-Chloroethyl)ether
Concen: 21.27 ng
RT: 5.69 min Scan# 646
Delta R.T. -0.01 min
Lab File: BF093849.D
Acq: 22 Mar 2017 23:02




Instrument :
BNA_F
ClientSampleId :
LOQ-WATER-06-QT2-2017


Tot Ion: 93 Resp: 270669
Ion Ratio Lower Upper
93 100
63 78.2 59.2 99.2
95 31.9 12.8 52.8



#12
1,3-Dichlorobenzene
Concen: 21.40 ng
RT: 5.92 min Scan# 686
Delta R.T. -0.00 min
Lab File: BF093849.D
Acq: 22 Mar 2017 23:02

Tgt Ion: 146 Resp: 305253
Ion Ratio Lower Upper
146 100
148 64.5 51.8 77.6
75 27.7 23.1 34.7





#13

1,4-Dichlorobenzene

Concen: 21.51 ng

RT: 6.00 min Scan# 700

Delta R.T. -0.01 min

Lab File: BF093849.D

Acq: 22 Mar 2017 23:02

Instrument :

BNA_F

ClientSampleId :

LOQ-WATER-06-QT2-2017


Tot Ion:146 Resp: 310753

Ion Ratio Lower Upper

146 100

148 63.9 52.2 78.2




111 41.1 30.3 45.5



Abundance Ion 146.00 (145.70 to 146.70): E

Ion 148.00 (147.70 to 148.70): E

Ion 111.00 (110.70 to 111.70): E



#14

1,2-Dichlorobenzene

Concen: 21.40 ng

RT: 6.19 min Scan# 731

Delta R.T. -0.00 min

Lab File: BF093849.D

Acq: 22 Mar 2017 23:02


Tgt Ion:146 Resp: 284761

Ion Ratio Lower Upper

146 100

148 64.3 50.4 75.6



111 41.7 38.2 57.4



Abundance Ion 146.00 (145.70 to 146.70): E

Ion 148.00 (147.70 to 148.70): E

Ion 111.00 (110.70 to 111.70): E





#15

Benzyl Alcohol

Concen: 21.79 ng

RT: 6.15 min Scan# 724

Delta R.T. -0.01 min

Lab File: BF093849.D

Acq: 22 Mar 2017 23:02

Instrument :

BNA_F

ClientSampleId :

LOQ-WATER-06-QT2-2017





Tot Ion: 79 Resp: 228206

Ion Ratio Lower Upper

79 100

108 85.6 63.4 95.0

77 62.1 49.2 73.8



#16

2,2'-oxybis(1-chloropropane)

Concen: 21.12 ng

RT: 6.31 min Scan# 752

Delta R.T. -0.00 min

Lab File: BF093849.D

Acq: 22 Mar 2017 23:02




Tgt Ion: 45 Resp: 414286


Ion Ratio Lower Upper

45 100

77 13.7 0.0 33.2

79 10.0 0.0 30.0









#17
2-Methylphenol
Concen: 20.56 ng
RT: 6.27 min Scan# 746
Delta R.T. -0.01 min
Lab File: BF093849.D
Acq: 22 Mar 2017 23:02

Instrument :
BNA_F
ClientSampleId :
LOQ-WATER-06-QT2-2017

Tot Ion:	107	Resp:	223903
Ion	Ratio	Lower	Upper
107	100		
108	118.3	93.4	140.0
77	41.5	35.8	53.6
79	40.5	34.8	52.2




Abundance Ion 107.00 (106.70 to 107.70): E
Ion 108.00 (107.70 to 108.70): E
Ion 77.00 (76.70 to 77.70): BF0
Ion 79.00 (78.70 to 79.70): BF0






#18
Hexachloroethane
Concen: 21.59 ng
RT: 6.58 min Scan# 798
Delta R.T. -0.00 min
Lab File: BF093849.D
Acq: 22 Mar 2017 23:02

Tgt Ion:	117	Resp:	109646
Ion	Ratio	Lower	Upper
117	100		
119	98.3	77.4	116.0
201	88.4	94.6	141.8#



Abundance Ion 117.00 (116.70 to 117.70): E
Ion 119.00 (118.70 to 119.70): E
Ion 201.00 (200.70 to 201.70): E








#19
n-Nitroso-di-n-propylamine
Concen: 21.51 ng
RT: 6.47 min Scan# 779
Delta R.T. -0.01 min
Lab File: BF093849.D
Acq: 22 Mar 2017 23:02





Instrument :
BNA_F
ClientSampleId :
LOQ-WATER-06-QT2-2017


Tot Ion:	70	Resp:	197809
Ion	Ratio	Lower	Upper
70	100		
42	60.5	47.2	70.8
101	10.5	7.2	10.8
130	22.6	15.9	23.9



#20
3+4-Methylphenols
Concen: 21.22 ng
RT: 6.46 min Scan# 777
Delta R.T. -0.01 min
Lab File: BF093849.D
Acq: 22 Mar 2017 23:02

Tgt Ion:	107	Resp:	279881
Ion	Ratio	Lower	Upper
107	100		
108	98.7	71.0	111.0
77	123.4	90.5	130.5
79	29.8	8.4	48.4








#21
Naphthalene-d8
Concen: 20.00 ng
RT: 7.49 min Scan# 953
Delta R.T. -0.00 min
Lab File: BF093849.D
Acq: 22 Mar 2017 23:02




Instrument :
BNA_F
ClientSampleId :
LOQ-WATER-06-QT2-2017


Tot Ion:136	Resp:	790664
Ion Ratio	Lower	Upper
136	100	
137	10.9	8.5 12.7
54	7.4	4.9 7.3#
68	5.9	4.1 6.1



#22
Acetophenone
Concen: 21.71 ng
RT: 6.46 min Scan# 777
Delta R.T. -0.01 min
Lab File: BF093849.D
Acq: 22 Mar 2017 23:02

Tgt Ion:105	Resp:	382495
Ion Ratio	Lower	Upper
105	100	
71	4.4	2.8 4.2#
51	32.1	30.7 46.1
120	21.7	17.4 26.0





#23

Nitrobenzene-d5

Concen: 96.77 ng

RT: 6.64 min Scan# 808

Delta R.T. -0.00 min

Lab File: BF093849.D

Acq: 22 Mar 2017 23:02

Instrument :

BNA_F

ClientSampleId :

LOQ-WATER-06-QT2-2017


Tot Ion: 82 Resp: 1340641

Ion Ratio Lower Upper

82 100

128 46.8 34.0 51.0




54 50.4 42.2 63.2



Abundance Ion 82.00 (81.70 to 82.70): BF0

Ion 128.00 (127.70 to 128.70): BF0

Ion 54.00 (53.70 to 54.70): BF0



#24

Nitrobenzene

Concen: 21.29 ng

RT: 6.66 min Scan# 811

Delta R.T. -0.01 min

Lab File: BF093849.D

Acq: 22 Mar 2017 23:02


Tgt Ion: 77 Resp: 290922

Ion Ratio Lower Upper

77 100

123 48.5 35.8 53.8



65 12.5 10.6 15.8




Abundance Ion 77.00 (76.70 to 77.70): BF0

Ion 123.00 (122.70 to 123.70): BF0

Ion 65.00 (64.70 to 65.70): BF0





#25

Isophorone

Concen: 20.94 ng

RT: 6.94 min Scan# 859

Delta R.T. -0.01 min

Lab File: BF093849.D

Acq: 22 Mar 2017 23:02

Instrument :

BNA_F

ClientSampleId :

LOQ-WATER-06-QT2-2017


Tot Ion: 82 Resp: 509823

Ion Ratio Lower Upper




82 100

95 6.5 5.3 7.9

138 18.0 13.1 19.7



Abundance Ion 82.00 (81.70 to 82.70): BF0
Ion 95.00 (94.70 to 95.70): BF0
Ion 138.00 (137.70 to 138.70): BF0



#26

2-Nitrophenol

Concen: 21.13 ng

RT: 7.03 min Scan# 874

Delta R.T. -0.01 min

Lab File: BF093849.D

Acq: 22 Mar 2017 23:02


Tgt Ion:139 Resp: 137693

Ion Ratio Lower Upper



139 100


109 25.0 21.0 31.6

65 45.7 33.0 49.6



Abundance Ion 139.00 (138.70 to 139.70): E
Ion 109.00 (108.70 to 109.70): E
Ion 65.00 (64.70 to 65.70): BF0









#27
2,4-Dimethylphenol
Concen: 21.00 ng
RT: 7.09 min Scan# 884
Delta R.T. -0.01 min
Lab File: BF093849.D
Acq: 22 Mar 2017 23:02




Instrument :
BNA_F
ClientSampleId :
LOQ-WATER-06-QT2-2017

Tot Ion:122 Resp: 235816
Ion Ratio Lower Upper
122 100
107 106.8 89.2 133.8
121 55.5 45.2 67.8



#28
bis(2-Chloroethoxy)methane
Concen: 21.35 ng
RT: 7.20 min Scan# 904
Delta R.T. -0.01 min
Lab File: BF093849.D
Acq: 22 Mar 2017 23:02

Tgt Ion: 93 Resp: 342693
Ion Ratio Lower Upper
93 100
95 32.3 26.5 39.7
123 12.6 9.0 13.4









#29
 2,4-Dichlorophenol
 Concen: 21.08 ng
 RT: 7.32 min Scan# 923
 Delta R.T. -0.01 min
 Lab File: BF093849.D
 Acq: 22 Mar 2017 23:02

Instrument :
 BNA_F
 ClientSampleId :
 LOQ-WATER-06-QT2-2017

Tot Ion:162 Resp: 221271
 Ion Ratio Lower Upper
 162 100
 164 64.3 44.6 84.6
 98 37.0 14.8 54.8




Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 164.00 (163.70 to 164.70): E
 Ion 98.00 (97.70 to 98.70): BF0






#30
 1,2,4-Trichlorobenzene
 Concen: 21.60 ng
 RT: 7.42 min Scan# 941
 Delta R.T. -0.01 min
 Lab File: BF093849.D
 Acq: 22 Mar 2017 23:02

Tgt Ion:180 Resp: 233537
 Ion Ratio Lower Upper
 180 100
 182 93.3 75.8 113.6
 145 32.5 25.3 37.9



Abundance Ion 180.00 (179.70 to 180.70): E
 Ion 182.00 (181.70 to 182.70): E
 Ion 145.00 (144.70 to 145.70): E





#31

Naphthalene

Concen: 21.70 ng

RT: 7.52 min Scan# 957

Delta R.T. -0.01 min

Lab File: BF093849.D

Acq: 22 Mar 2017 23:02

Instrument :

BNA_F

ClientSampleId :

LOQ-WATER-06-QT2-2017


Tot Ion:128 Resp: 825152

Ion Ratio Lower Upper

128 100

129 11.3 9.0 13.6




127 13.4 10.8 16.2



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#32

Benzoic acid

Concen: 18.45 ng

RT: 7.20 min Scan# 904

Delta R.T. -0.04 min

Lab File: BF093849.D

Acq: 22 Mar 2017 23:02


Tgt Ion:122 Resp: 137883

Ion Ratio Lower Upper

122 100

105 124.6 103.3 143.3



77 88.4 73.7 113.7



Abundance Ion 122.00 (121.70 to 122.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 77.00 (76.70 to 77.70): BF0









#33
4-Chloroaniline
Concen: 20.70 ng
RT: 7.57 min Scan# 967
Delta R.T. -0.01 min
Lab File: BF093849.D
Acq: 22 Mar 2017 23:02




Instrument :
BNA_F
ClientSampleId :
LOQ-WATER-06-QT2-2017


Tot Ion:	127	Resp:	319029
Ion	Ratio	Lower	Upper
127	100		
129	32.8	26.2	39.2
65	29.3	27.8	41.8
92	18.8	16.8	25.2



#34
Hexachlorobutadiene
Concen: 21.65 ng
RT: 7.67 min Scan# 984
Delta R.T. -0.01 min
Lab File: BF093849.D
Acq: 22 Mar 2017 23:02

Tgt Ion:	225	Resp:	120030
Ion	Ratio	Lower	Upper
225	100		
223	63.9	48.8	73.2
227	62.1	51.1	76.7









#35
 Caprolactam
 Concen: 20.51 ng
 RT: 8.00 min Scan# 1039
 Delta R.T. -0.03 min
 Lab File: BF093849.D
 Acq: 22 Mar 2017 23:02

Instrument :
 BNA_F
 ClientSampleId :
 LOQ-WATER-06-QT2-2017

Tot Ion:113 Resp: 68652
 Ion Ratio Lower Upper
 113 100
 55 169.6 150.2 190.2
 56 125.4 107.8 147.8




Abundance Ion 113.00 (112.70 to 113.70): E
 Ion 55.00 (54.70 to 55.70): BF0
 Ion 56.00 (55.70 to 56.70): BF0






#36
 4-Chloro-3-methylphenol
 Concen: 20.88 ng
 RT: 8.17 min Scan# 1068
 Delta R.T. -0.01 min
 Lab File: BF093849.D
 Acq: 22 Mar 2017 23:02

Tgt Ion:107 Resp: 229061
 Ion Ratio Lower Upper
 107 100
 144 24.8 24.2 36.4
 142 78.5 73.4 110.0



Abundance Ion 107.00 (106.70 to 107.70): E
 Ion 144.00 (143.70 to 144.70): E
 Ion 142.00 (141.70 to 142.70): E









#37
2-Methylnaphthalene
Concen: 21.84 ng
RT: 8.36 min Scan# 1100
Delta R.T. -0.01 min
Lab File: BF093849.D
Acq: 22 Mar 2017 23:02

Instrument :
BNA_F
ClientSampleId :
LOQ-WATER-06-QT2-2017

Tot Ion	Ratio	Lower	Upper
142	100		
141	87.7	71.3	106.9
115	32.3	27.1	40.7




Abundance Ion 142.00 (141.70 to 142.70): E
Ion 141.00 (140.70 to 141.70): E
Ion 115.00 (114.70 to 115.70): E





#38
Acenaphthene-d10
Concen: 20.00 ng
RT: 9.63 min Scan# 1317
Delta R.T. -0.01 min
Lab File: BF093849.D
Acq: 22 Mar 2017 23:02

Tgt Ion	Ratio	Lower	Upper
164	100		
162	102.1	82.6	123.8
160	44.3	36.2	54.2



Abundance Ion 164.00 (163.70 to 164.70): E
Ion 162.00 (161.70 to 162.70): E
Ion 160.00 (159.70 to 160.70): E





#39

1,2,4,5-Tetrachlorobenzene

Concen: 22.11 ng

RT: 8.56 min Scan# 1135

Delta R.T. -0.01 min

Lab File: BF093849.D

Acq: 22 Mar 2017 23:02

Instrument :

BNA_F

ClientSampleId :

LOQ-WATER-06-QT2-2017

Tot Ion:216 Resp: 215463


Ion Ratio Lower Upper

216 100

214 79.3 63.4 95.2

179 22.5 17.9 26.9

108 21.4 16.5 24.7






Abundance Ion 216.00 (215.70 to 216.70): E

Ion 214.00 (213.70 to 214.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 108.00 (107.70 to 108.70): E



#40

Hexachlorocyclopentadiene

Concen: 21.84 ng

RT: 8.56 min Scan# 1134

Delta R.T. -0.01 min

Lab File: BF093849.D

Acq: 22 Mar 2017 23:02


Tgt Ion:237 Resp: 99584

Ion Ratio Lower Upper

237 100

235 62.7 42.6 82.6



272 11.5 0.0 31.9




Abundance Ion 236.80 (236.50 to 237.50): E

Ion 234.90 (234.60 to 235.60): E

Ion 271.75 (271.45 to 272.45): E





#41

2,4,6-Tribromophenol

Concen: 131.89 ng

RT: 10.61 min Scan# 1483

Delta R.T. -0.00 min

Lab File: BF093849.D

Acq: 22 Mar 2017 23:02

Instrument :

BNA_F

ClientSampleId :

LOQ-WATER-06-QT2-2017


Tot Ion:330 Resp: 371292

Ion Ratio Lower Upper

330 100

332 97.6 76.6 115.0




141 40.8 31.4 47.2



Abundance Ion 329.80 (329.50 to 330.50): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E



#42

2,4,6-Trichlorophenol

Concen: 21.18 ng

RT: 8.70 min Scan# 1159

Delta R.T. -0.00 min

Lab File: BF093849.D

Acq: 22 Mar 2017 23:02


Tgt Ion:196 Resp: 146041

Ion Ratio Lower Upper

196 100

198 95.8 74.2 111.4



200 31.7 24.0 36.0



Abundance Ion 196.00 (195.70 to 196.70): E

Ion 198.00 (197.70 to 198.70): E

Ion 200.00 (199.70 to 200.70): E









#43
 2,4,5-Trichlorophenol
 Concen: 21.07 ng
 RT: 8.75 min Scan# 1166
 Delta R.T. -0.01 min
 Lab File: BF093849.D
 Acq: 22 Mar 2017 23:02

Instrument :
 BNA_F
 ClientSampleId :
 LOQ-WATER-06-QT2-2017

Tot Ion:196 Resp: 157206
 Ion Ratio Lower Upper
 196 100
 198 97.3 75.7 113.5
 97 52.0 47.7 71.5
 132 30.7 28.0 42.0




Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 97.00 (96.70 to 97.70): BF0
 Ion 132.00 (131.70 to 132.70): E





#44
 2-Fluorobiphenyl
 Concen: 96.25 ng
 RT: 8.82 min Scan# 1179
 Delta R.T. -0.00 min
 Lab File: BF093849.D
 Acq: 22 Mar 2017 23:02

Tgt Ion:172 Resp: 2248044
 Ion Ratio Lower Upper
 172 100
 171 36.5 29.6 44.4
 170 24.4 19.8 29.8



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E









#45
 1,1'-Biphenyl
 Concen: 22.24 ng
 RT: 8.93 min Scan# 1198
 Delta R.T. -0.01 min
 Lab File: BF093849.D
 Acq: 22 Mar 2017 23:02

Instrument :
 BNA_F
 ClientSampleId :
 LOQ-WATER-06-QT2-2017

Tot Ion	Ratio	Lower	Upper
154	100		
153	40.1	21.2	61.2
76	13.5	0.0	29.9




Abundance Ion 154.00 (153.70 to 154.70): E
 Ion 153.00 (152.70 to 153.70): E
 Ion 76.00 (75.70 to 76.70): BFO






#46
 2-Chloronaphthalene
 Concen: 21.98 ng
 RT: 8.96 min Scan# 1202
 Delta R.T. -0.00 min
 Lab File: BF093849.D
 Acq: 22 Mar 2017 23:02

Tgt Ion	Ratio	Lower	Upper
162	100		
127	38.3	28.3	42.5
164	32.3	27.0	40.4



Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 127.00 (126.70 to 127.70): E
 Ion 164.00 (163.70 to 164.70): E









#47
2-Nitroaniline
Concen: 20.58 ng
RT: 9.07 min Scan# 1221
Delta R.T. -0.01 min
Lab File: BF093849.D
Acq: 22 Mar 2017 23:02




Instrument :
BNA_F
ClientSampleId :
LOQ-WATER-06-QT2-2017


Tot Ion: 65 Resp: 137360
Ion Ratio Lower Upper
65 100
92 69.2 53.9 80.9
138 112.8 78.8 118.2



#48
Acenaphthylene
Concen: 22.10 ng
RT: 9.46 min Scan# 1288
Delta R.T. -0.01 min
Lab File: BF093849.D
Acq: 22 Mar 2017 23:02

Tgt Ion: 152 Resp: 826293
Ion Ratio Lower Upper
152 100
151 20.7 16.8 25.2
153 13.1 10.8 16.2









#49
 Dimethylphthalate
 Concen: 21.65 ng
 RT: 9.32 min Scan# 1263
 Delta R.T. -0.01 min
 Lab File: BF093849.D
 Acq: 22 Mar 2017 23:02

Instrument :
 BNA_F
 ClientSampleId :
 LOQ-WATER-06-QT2-2017

Tot Ion:163	Resp:	548312
Ion Ratio	Lower	Upper
163	100	
194	3.4	2.6 4.0
164	10.1	8.4 12.6




Abundance Ion 163.00 (162.70 to 163.70): E
 Ion 194.00 (193.70 to 194.70): E
 Ion 164.00 (163.70 to 164.70): E






#50
 2,6-Dinitrotoluene
 Concen: 21.50 ng
 RT: 9.37 min Scan# 1273
 Delta R.T. -0.01 min
 Lab File: BF093849.D
 Acq: 22 Mar 2017 23:02

Tgt Ion:165	Resp:	124818
Ion Ratio	Lower	Upper
165	100	
63	58.9	34.3 51.5#
89	52.6	37.1 55.7



Abundance Ion 165.00 (164.70 to 165.70): E
 Ion 63.00 (62.70 to 63.70): BF0
 Ion 89.00 (88.70 to 89.70): BF0





#51

Acenaphthene

Concen: 21.65 ng

RT: 9.67 min Scan# 1324

Delta R.T. -0.01 min

Lab File: BF093849.D

Acq: 22 Mar 2017 23:02

Instrument :

BNA_F

ClientSampleId :

LOQ-WATER-06-QT2-2017


Tot Ion:154 Resp: 472275

Ion Ratio Lower Upper

154 100

153 114.0 90.5 135.7




152 54.7 43.6 65.4



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#52

3-Nitroaniline

Concen: 20.44 ng

RT: 9.57 min Scan# 1307

Delta R.T. -0.01 min

Lab File: BF093849.D

Acq: 22 Mar 2017 23:02


Tgt Ion:138 Resp: 139369

Ion Ratio Lower Upper

138 100

108 10.6 9.1 13.7



92 109.3 93.8 140.6




Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BF0









#53
 2,4-Dinitrophenol
 Concen: 19.24 ng
 RT: 9.70 min Scan# 1329
 Delta R.T. -0.01 min
 Lab File: BF093849.D
 Acq: 22 Mar 2017 23:02




Instrument :
 BNA_F
 ClientSampleId :
 LOQ-WATER-06-QT2-2017

Tot Ion:184 Resp: 48098
 Ion Ratio Lower Upper
 184 100
 63 65.6 62.7 94.1
 154 62.9 81.1 121.7#



#54
 Dibenzofuran
 Concen: 22.26 ng
 RT: 9.89 min Scan# 1360
 Delta R.T. -0.01 min
 Lab File: BF093849.D
 Acq: 22 Mar 2017 23:02

Tgt Ion:168 Resp: 660924
 Ion Ratio Lower Upper
 168 100
 139 37.8 30.6 45.8
 169 13.3 10.8 16.2





#55

4-Nitrophenol

Concen: 20.46 ng

RT: 9.78 min Scan# 1342

Delta R.T. -0.01 min

Lab File: BF093849.D

Acq: 22 Mar 2017 23:02

Instrument :

BNA_F

ClientSampleId :

LOQ-WATER-06-QT2-2017


Tot Ion:139 Resp: 109232

Ion Ratio Lower Upper

139 100

109 62.8 34.1 74.1




65 85.3 31.4 71.4#



Abundance Ion 139.00 (138.70 to 139.70): E

Ion 109.00 (108.70 to 109.70): E

Ion 65.00 (64.70 to 65.70): BF0



#56

2,4-Dinitrotoluene

Concen: 22.27 ng

RT: 9.87 min Scan# 1357

Delta R.T. -0.01 min

Lab File: BF093849.D

Acq: 22 Mar 2017 23:02

Tgt Ion:165 Resp: 162439


Ion Ratio Lower Upper

165 100

63 41.7 25.4 38.0#

89 68.8 46.4 69.6

182 2.5 1.8 2.6






Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BF0

Ion 89.00 (88.70 to 89.70): BF0

Ion 182.00 (181.70 to 182.70): E





#57

Fluorene

Concen: 21.74 ng

RT: 10.31 min Scan# 1432

Delta R.T. -0.01 min

Lab File: BF093849.D

Acq: 22 Mar 2017 23:02

Instrument :

BNA_F

ClientSampleId :

LOQ-WATER-06-QT2-2017


Tot Ion:166 Resp: 535398

Ion Ratio Lower Upper

166 100

165 101.1 78.8 118.2




167 13.2 10.6 16.0



Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E



#58

2,3,4,6-Tetrachlorophenol

Concen: 21.47 ng

RT: 10.03 min Scan# 1385

Delta R.T. -0.01 min

Lab File: BF093849.D

Acq: 22 Mar 2017 23:02

Tgt Ion:232 Resp: 109893


Ion Ratio Lower Upper

232 100

131 57.3 44.8 67.2

130 2.5 2.3 3.5

166 36.9 29.0 43.4






Abundance Ion 232.00 (231.70 to 232.70): E

Ion 131.00 (130.70 to 131.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 166.00 (165.70 to 166.70): E









#59
Diethylphthalate
Concen: 21.85 ng
RT: 10.19 min Scan# 1411
Delta R.T. -0.01 min
Lab File: BF093849.D
Acq: 22 Mar 2017 23:02

Instrument :
BNA_F
ClientSampleId :
LOQ-WATER-06-QT2-2017

Tot Ion:149 Resp: 546940
Ion Ratio Lower Upper
149 100
177 21.0 16.7 25.1
150 12.1 10.2 15.2




Abundance Ion 149.00 (148.70 to 149.70): E
Ion 177.00 (176.70 to 177.70): E
Ion 150.00 (149.70 to 150.70): E






#60
4-Chlorophenyl-phenylether
Concen: 22.26 ng
RT: 10.32 min Scan# 1433
Delta R.T. -0.01 min
Lab File: BF093849.D
Acq: 22 Mar 2017 23:02

Tgt Ion:204 Resp: 233507
Ion Ratio Lower Upper
204 100
206 31.7 26.2 39.2
141 64.9 60.8 91.2



Abundance Ion 204.00 (203.70 to 204.70): E
Ion 206.00 (205.70 to 206.70): E
Ion 141.00 (140.70 to 141.70): E





#61

4-Nitroaniline

Concen: 19.91 ng

RT: 10.33 min Scan# 1435

Delta R.T. -0.01 min

Lab File: BF093849.D

Acq: 22 Mar 2017 23:02

Instrument :

BNA_F

ClientSampleId :

LOQ-WATER-06-QT2-2017


Tot Ion:138 Resp: 130268

Ion Ratio Lower Upper

138 100

92 44.2 21.8 61.8




108 65.7 42.3 82.3



Abundance Ion 138.00 (137.70 to 138.70): E

Ion 92.00 (91.70 to 92.70): BF0

Ion 108.00 (107.70 to 108.70): E



#62

Azobenzene

Concen: 21.73 ng

RT: 10.51 min Scan# 1466

Delta R.T. -0.01 min

Lab File: BF093849.D

Acq: 22 Mar 2017 23:02

Tgt Ion: 77 Resp: 514454


Ion Ratio Lower Upper

77 100

182 22.4 0.1 40.1

105 23.0 3.6 43.6

51 27.6 9.4 49.4






Abundance Ion 77.00 (76.70 to 77.70): BF0

Ion 182.00 (181.70 to 182.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 51.00 (50.70 to 51.70): BF0









#63
Phenanthrene-d10
Concen: 20.00 ng
RT: 11.46 min Scan# 1627
Delta R.T. -0.01 min
Lab File: BF093849.D
Acq: 22 Mar 2017 23:02

Instrument :
BNA_F
ClientSampleId :
LOQ-WATER-06-QT2-2017

Tot Ion:188 Resp: 613127
Ion Ratio Lower Upper
188 100
94 11.9 9.4 14.2
80 12.6 10.2 15.2




Abundance Ion 188.00 (187.70 to 188.70): E
Ion 94.00 (93.70 to 94.70): BF0
Ion 80.00 (79.70 to 80.70): BF0






#64
4,6-Dinitro-2-methylphenol
Concen: 19.16 ng
RT: 10.37 min Scan# 1442
Delta R.T. -0.01 min
Lab File: BF093849.D
Acq: 22 Mar 2017 23:02

Tgt Ion:198 Resp: 71961
Ion Ratio Lower Upper
198 100
51 56.0 53.2 93.2
105 51.7 45.8 85.8



Abundance Ion 198.00 (197.70 to 198.70): E
Ion 51.00 (50.70 to 51.70): BF0
Ion 105.00 (104.70 to 105.70): E









#65
 n-Nitrosodiphenylamine
 Concen: 22.00 ng
 RT: 10.46 min Scan# 1458
 Delta R.T. -0.01 min
 Lab File: BF093849.D
 Acq: 22 Mar 2017 23:02




Instrument :
 BNA_F
 ClientSampleId :
 LOQ-WATER-06-QT2-2017


Tot Ion:169 Resp: 466501
 Ion Ratio Lower Upper
 169 100
 168 68.3 53.2 79.8
 167 37.0 29.5 44.3



#66
 4-Bromophenyl-phenylether
 Concen: 21.68 ng
 RT: 10.92 min Scan# 1535
 Delta R.T. -0.01 min
 Lab File: BF093849.D
 Acq: 22 Mar 2017 23:02

Tgt Ion:248 Resp: 130747
 Ion Ratio Lower Upper
 248 100
 250 98.2 76.8 115.2
 141 86.4 68.6 103.0









#67
Hexachlorobenzene
Concen: 21.97 ng
RT: 10.99 min Scan# 1547
Delta R.T. -0.01 min
Lab File: BF093849.D
Acq: 22 Mar 2017 23:02

Instrument :
BNA_F
ClientSampleId :
LOQ-WATER-06-QT2-2017

Tot Ion:284 Resp: 134121
Ion Ratio Lower Upper
284 100
142 41.2 22.8 34.2#
249 31.7 24.0 36.0




Abundance Ion 283.80 (283.50 to 284.50): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E






#68
Atrazine
Concen: 21.65 ng
RT: 11.13 min Scan# 1571
Delta R.T. -0.01 min
Lab File: BF093849.D
Acq: 22 Mar 2017 23:02

Tgt Ion:200 Resp: 137470
Ion Ratio Lower Upper
200 100
173 27.5 6.3 46.3
215 44.2 27.2 67.2



Abundance Ion 200.00 (199.70 to 200.70): E
Ion 173.00 (172.70 to 173.70): E
Ion 215.00 (214.70 to 215.70): E









#69
 Pentachlorophenol
 Concen: 20.18 ng
 RT: 11.22 min Scan# 1587
 Delta R.T. -0.01 min
 Lab File: BF093849.D
 Acq: 22 Mar 2017 23:02




Instrument :
 BNA_F
 ClientSampleId :
 LOQ-WATER-06-QT2-2017


Tot Ion:266 Resp: 76690
 Ion Ratio Lower Upper
 266 100
 268 62.1 51.1 76.7
 264 64.5 51.7 77.5



#70
 Phenanthrene
 Concen: 22.30 ng
 RT: 11.49 min Scan# 1632
 Delta R.T. -0.01 min
 Lab File: BF093849.D
 Acq: 22 Mar 2017 23:02

Tgt Ion:178 Resp: 760455
 Ion Ratio Lower Upper
 178 100
 176 19.8 16.2 24.4
 179 15.1 12.6 19.0









#71
 Anthracene
 Concen: 22.25 ng
 RT: 11.55 min Scan# 1643
 Delta R.T. -0.01 min
 Lab File: BF093849.D
 Acq: 22 Mar 2017 23:02




Instrument :
 BNA_F
 ClientSampleId :
 LOQ-WATER-06-QT2-2017


Tot Ion:178 Resp: 781528
 Ion Ratio Lower Upper
 178 100
 176 19.1 15.8 23.8
 179 15.5 12.7 19.1



#72
 Carbazole
 Concen: 21.56 ng
 RT: 11.74 min Scan# 1676
 Delta R.T. -0.01 min
 Lab File: BF093849.D
 Acq: 22 Mar 2017 23:02

Tgt Ion:167 Resp: 776225
 Ion Ratio Lower Upper
 167 100
 166 21.6 18.1 27.1
 139 13.9 11.7 17.5









#73
Di-n-butylphthalate
Concen: 22.85 ng
RT: 12.20 min Scan# 1754
Delta R.T. -0.00 min
Lab File: BF093849.D
Acq: 22 Mar 2017 23:02




Instrument :
BNA_F
ClientSampleId :
LOQ-WATER-06-QT2-2017


Tot Ion:149 Resp: 855592
Ion Ratio Lower Upper
149 100
150 9.1 7.7 11.5
104 4.5 4.0 6.0



#74
Fluoranthene
Concen: 21.85 ng
RT: 12.95 min Scan# 1881
Delta R.T. -0.01 min
Lab File: BF093849.D
Acq: 22 Mar 2017 23:02

Tgt Ion:202 Resp: 763137
Ion Ratio Lower Upper
202 100
101 13.3 0.0 33.2
203 17.6 0.0 38.1





#75

Chrysene-d12

Concen: 20.00 ng

RT: 14.71 min Scan# 2180

Delta R.T. -0.01 min

Lab File: BF093849.D

Acq: 22 Mar 2017 23:02

Instrument :

BNA_F

ClientSampleId :

LOQ-WATER-06-QT2-2017


Tot Ion:240 Resp: 504924

Ion Ratio Lower Upper

240 100

120 16.2 5.8 8.8#




236 26.0 20.5 30.7



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#76

Benzidine

Concen: 19.49 ng

RT: 13.12 min Scan# 1909

Delta R.T. -0.01 min

Lab File: BF093849.D

Acq: 22 Mar 2017 23:02


Tgt Ion:184 Resp: 389560

Ion Ratio Lower Upper

184 100

185 24.7 15.5 23.3#



183 12.6 9.8 14.6




Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E









#77
Pvrene
Concen: 21.66 ng
RT: 13.23 min Scan# 1928
Delta R.T. -0.01 min
Lab File: BF093849.D
Acq: 22 Mar 2017 23:02




Instrument :
BNA_F
ClientSampleId :
LOQ-WATER-06-QT2-2017


Tot Ion:202 Resp: 793579
Ion Ratio Lower Upper
202 100
200 21.8 17.6 26.4
203 17.6 14.4 21.6



#78
Terphenyl-d14
Concen: 86.89 ng
RT: 13.44 min Scan# 1965
Delta R.T. 0.01 min
Lab File: BF093849.D
Acq: 22 Mar 2017 23:02

Tgt Ion:244 Resp: 1957358
Ion Ratio Lower Upper
244 100
212 7.1 6.0 9.0
122 12.8 10.3 15.5









#79
Butylbenzylphthalate
Concen: 21.76 ng
RT: 14.04 min Scan# 2067
Delta R.T. -0.00 min
Lab File: BF093849.D
Acq: 22 Mar 2017 23:02




Instrument :
BNA_F
ClientSampleId :
LOQ-WATER-06-QT2-2017

Tot Ion:149 Resp: 339745
Ion Ratio Lower Upper
149 100
91 68.1 45.0 67.4#
206 18.2 17.0 25.6



#80
Benzo(a)anthracene
Concen: 21.40 ng
RT: 14.70 min Scan# 2178
Delta R.T. -0.01 min
Lab File: BF093849.D
Acq: 22 Mar 2017 23:02

Tgt Ion:228 Resp: 629990
Ion Ratio Lower Upper
228 100
226 27.6 22.6 33.8
229 20.0 16.3 24.5









#81
 3,3'-Dichlorobenzidine
 Concen: 20.43 ng
 RT: 14.67 min Scan# 2173
 Delta R.T. -0.01 min
 Lab File: BF093849.D
 Acq: 22 Mar 2017 23:02

Instrument :
 BNA_F
 ClientSampleId :
 LOQ-WATER-06-QT2-2017

Tot Ion	Ratio	Lower	Upper
252	100		
254	67.0	51.5	77.3
126	15.6	15.8	23.8#




Abundance Ion 252.00 (251.70 to 252.70): E
 Ion 254.00 (253.70 to 254.70): E
 Ion 126.00 (125.70 to 126.70): E






#82
 Chrysene
 Concen: 22.30 ng
 RT: 14.74 min Scan# 2186
 Delta R.T. -0.01 min
 Lab File: BF093849.D
 Acq: 22 Mar 2017 23:02

Tgt Ion	Ratio	Lower	Upper
228	100		
226	30.5	24.9	37.3
229	19.6	15.8	23.6



Abundance Ion 228.00 (227.70 to 228.70): E
 Ion 226.00 (225.70 to 226.70): E
 Ion 229.00 (228.70 to 229.70): E









#83
 Bis(2-ethylhexyl)phthalate
 Concen: 23.62 ng
 RT: 14.76 min Scan# 2189
 Delta R.T. -0.00 min
 Lab File: BF093849.D
 Acq: 22 Mar 2017 23:02




Instrument :
 BNA_F
 ClientSampleId :
 LOQ-WATER-06-QT2-2017


Tot Ion	Ratio	Lower	Upper
149	100		
167	26.8	21.0	31.4
279	3.0	1.9	2.9#



#84
 Di-n-octyl phthalate
 Concen: 22.10 ng
 RT: 15.52 min Scan# 2317
 Delta R.T. -0.01 min
 Lab File: BF093849.D
 Acq: 22 Mar 2017 23:02

Tgt Ion	Ratio	Lower	Upper
149	100		
167	1.4	1.2	1.8
43	10.1	8.5	12.7









#85
 Indeno(1,2,3-cd)pyrene
 Concen: 18.83 ng
 RT: 17.73 min Scan# 2694
 Delta R.T. -0.01 min
 Lab File: BF093849.D
 Acq: 22 Mar 2017 23:02




Instrument :
 BNA_F
 ClientSampleId :
 LOQ-WATER-06-QT2-2017


Tot Ion:276 Resp: 445702
 Ion Ratio Lower Upper
 276 100
 138 37.4 27.8 41.6
 277 25.1 20.2 30.4



#86
 Perylene-d12
 Concen: 20.00 ng
 RT: 16.35 min Scan# 2459
 Delta R.T. -0.00 min
 Lab File: BF093849.D
 Acq: 22 Mar 2017 23:02

Tgt Ion:264 Resp: 399468
 Ion Ratio Lower Upper
 264 100
 260 22.7 19.5 29.3
 265 20.6 17.2 25.8





#87

Benzo(b)fluoranthene

Concen: 21.58 ng

RT: 15.93 min Scan# 2388

Delta R.T. -0.01 min

Lab File: BF093849.D

Acq: 22 Mar 2017 23:02

Instrument :

BNA_F

ClientSampleId :

LOQ-WATER-06-QT2-2017


Tot Ion:252 Resp: 528406

Ion Ratio Lower Upper

252 100

253 22.5 18.1 27.1




125 12.2 9.8 14.8



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#88

Benzo(k)fluoranthene

Concen: 22.01 ng

RT: 15.96 min Scan# 2393

Delta R.T. -0.01 min

Lab File: BF093849.D

Acq: 22 Mar 2017 23:02


Tgt Ion:252 Resp: 531238

Ion Ratio Lower Upper

252 100

253 21.7 18.4 27.6



125 12.9 13.1 19.7#




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E





#89

Benzo(a)pyrene

Concen: 21.30 ng

RT: 16.29 min Scan# 2448

Delta R.T. -0.01 min

Lab File: BF093849.D

Acq: 22 Mar 2017 23:02

Instrument :

BNA_F

ClientSampleId :

LOQ-WATER-06-QT2-2017


Tot Ion:252 Resp: 480244

Ion Ratio Lower Upper

252 100

253 22.4 17.5 26.3




125 14.4 11.0 16.4



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#90

Dibenzo(a,h)anthracene

Concen: 19.95 ng

RT: 17.76 min Scan# 2698

Delta R.T. -0.01 min

Lab File: BF093849.D

Acq: 22 Mar 2017 23:02


Tgt Ion:278 Resp: 381010

Ion Ratio Lower Upper

278 100

139 24.4 16.6 24.8



279 24.8 19.3 28.9



Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E





#91

Benzo(a,h,i)perylene

Concen: 19.27 ng

RT: 18.14 min Scan# 2764

Delta R.T. -0.02 min

Lab File: BF093849.D

Acq: 22 Mar 2017 23:02

Instrument :

BNA_F

ClientSampleId :

LOQ-WATER-06-QT2-2017


Tot Ion:276 Resp: 370801

Ion Ratio Lower Upper

276 100

277 23.8 18.6 28.0

138 31.5 25.4 38.0



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

