

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA_F\DATA\BF032218\
 Data File : BF103857.D
 Acq On : 22 Mar 2018 17:39
 Operator : JU/SJ
 Sample : J1997-01 2X
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DGA-TOP-LAYER-1-A

Manual Integrations
 APPROVED

Sohil
 3/23/2018 2:46:43 PM

Quant Time: Mar 23 01:19:08 2018
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF031518.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 22 11:56:49 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.97	152	141984	20.00	ng	0.00
21) Naphthalene-d8	8.26	136	564475	20.00	ng	0.00
38) Acenaphthene-d10	10.02	164	244880	20.00	ng	0.00
63) Phenanthrene-d10	11.50	188	368426	20.00	ng	0.00
75) Chrysene-d12	14.16	240	297823	20.00	ng	0.00
86) Perylene-d12	15.70	264	225765	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.58	112	17751	1.90	ng	-0.01
7) Phenol-d6	6.58	99	80992	7.09	ng	-0.02
23) Nitrobenzene-d5	7.53	82	51344	6.34	ng	-0.01
41) 2,4,6-Tribromophenol	10.80	330	542	0.26	ng	0.00
44) 2-Fluorobiphenyl	9.33	172	116581	7.59	ng	0.00
78) Terphenyl-d14	13.09	244	86465	6.74	ng	0.00
Target Compounds						
70) Phenanthrene	11.53	178	79968	3.968	ng	98
74) Fluoranthene	12.72	202	85856	4.135	ng	98
77) Pyrene	12.96	202	72232	3.302	ng	98
82) Chrysene	14.19	228	37029	2.061	ng	98
87) Benzo(b)fluoranthene	15.24	252	35604m	2.496	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA_F\DATA\BF032218\
Data File : BF103857.D
Acq On : 22 Mar 2018 17:39
Operator : JU/SJ
Sample : J1997-01 2X
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DGA-TOP-LAYER-1-A

Manual Integrations
APPROVED

Sohil
3/23/2018 2:46:43 PM

Quant Time: Mar 23 01:19:08 2018
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF031518.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 22 11:56:49 2018
Response via : Initial Calibration

