




Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA F\DATA\BF032318\
Data File : BF103892.D
Acq On : 23 Mar 2018 21:26
Operator : JU/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 24 04:07:49 2018
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF031518.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 23 13:26:17 2018
Response via : Initial Calibration



TIC: BF103892.D

(69) Pentachlorophenol (C)




11.310min (-0.000) 0.00ng

response 28426

Ion	Exp%	Act%
265.70	100	100
268.00	63.90	70.33
264.00	61.90	63.98
0.00	0.00	0.00

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA F\DATA\BF032318\
 Data File : BF103892.D
 Acq On : 23 Mar 2018 21:26
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 24 04:07:49 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF031518.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 23 13:26:17 2018
 Response via : Initial Calibration



TIC: BF103892.D

(76) Benzidine

12.851min (-0.000) 58184.41ng

response 297791

Ion	Exp%	Act%
184.00	100	100
185.00	15.70	14.40
183.00	11.60	11.62
0.00	0.00	0.00