

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032321\
 Data File : BF123415.D
 Acq On : 23 Mar 2021 21:30
 Operator : JU/CG
 Sample : M1712-01
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 EXTERNAL-PAD-1

Quant Time: Mar 24 00:27:27 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF032321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 23 15:58:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.916	152	226286	20.00	ng	0.00
21) Naphthalene-d8	8.198	136	858625	20.00	ng	# 0.00
39) Acenaphthene-d10	9.957	164	467941	20.00	ng	0.00
64) Phenanthrene-d10	11.451	188	706874	20.00	ng	0.00
76) Chrysene-d12	14.098	240	367750	20.00	ng	0.00
86) Perylene-d12	15.603	264	421953	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.534	112	1155733	83.55	ng	0.01
7) Phenol-d6	6.533	99	1526583	82.91	ng	0.00
23) Nitrobenzene-d5	7.475	82	1047867	65.87	ng	0.00
42) 2,4,6-Tribromophenol	10.751	330	378280	72.31	ng	0.00
45) 2-Fluorobiphenyl	9.280	172	1933622	59.03	ng	0.00
79) Terphenyl-d14	13.045	244	1436284	66.19	ng	0.00
Target Compounds						
10) Phenol	6.545	94	46228	2.45	ng	98
50) Dimethylphthalate	9.669	163	71275	2.14	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032321\
Data File : BF123415.D
Acq On : 23 Mar 2021 21:30
Operator : JU/CG
Sample : M1712-01
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
EXTERNAL-PAD-1

Quant Time: Mar 24 00:27:27 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF032321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 23 15:58:35 2021
Response via : Initial Calibration

