

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032321\
 Data File : BF123427.D
 Acq On : 24 Mar 2021 7:32
 Operator : JU/CG
 Sample : M1712-05
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 RED-BRICK-1

Quant Time: Mar 24 08:13:14 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF032321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 23 15:58:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.916	152	261595	20.00	ng	0.00
21) Naphthalene-d8	8.198	136	999247	20.00	ng	# 0.00
39) Acenaphthene-d10	9.957	164	551768	20.00	ng	0.00
64) Phenanthrene-d10	11.451	188	873774	20.00	ng	0.00
76) Chrysene-d12	14.098	240	418981	20.00	ng	0.00
86) Perylene-d12	15.598	264	424398	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.534	112	1291822	80.79	ng	0.01
7) Phenol-d6	6.534	99	1646793	77.36	ng	0.00
23) Nitrobenzene-d5	7.475	82	1145810	61.89	ng	0.00
42) 2,4,6-Tribromophenol	10.745	330	432405	70.10	ng	-0.01
45) 2-Fluorobiphenyl	9.275	172	2135300	55.29	ng	0.00
79) Terphenyl-d14	13.045	244	1794224	72.58	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032321\
Data File : BF123427.D
Acq On : 24 Mar 2021 7:32
Operator : JU/CG
Sample : M1712-05
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
RED-BRICK-1

Quant Time: Mar 24 08:13:14 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF032321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 23 15:58:35 2021
Response via : Initial Calibration

