


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032321\
 Data File : BF123418.D
 Acq On : 24 Mar 2021 00:42
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF032321.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Mar 24 12:51:15 2021






AutoFind: Scans 1635, 1636, 1637; Background Corrected with Scan 1630

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	50.0	160219	PASS
68	69	0.00	2	1.7	2066	PASS
69	198	0.00	100	38.9	124511	PASS
70	69	0.00	2	0.6	700	PASS
127	198	10	80	48.5	155203	PASS
197	198	0.00	2	0.7	2302	PASS
198	198	100	100	100.0	320235	PASS
199	198	5	9	6.6	21172	PASS
275	198	10	60	27.7	88571	PASS
365	198	1	100	3.5	11330	PASS
441	198	0.01	100	14.8	47235	PASS
442	442	50	100	100.0	304701	PASS
443	442	15	24	18.6	56811	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032321\
 Data File : BF123418.D
 Acq On : 24 Mar 2021 00:42
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Quant Time: Mar 24 02:05:30 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF032321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 23 15:58:35 2021
 Response via : Initial Calibration



TIC: BF123418.D\data.ms

(70) Pentachlorophenol (C)

11.251min (-0.000) 24590.16 ng




response 230965

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.10	62.27
264.00	62.30	64.45
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032321\
Data File : BF123418.D
Acq On : 24 Mar 2021 00:42
Operator : JU/CG
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Mar 24 02:05:30 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF032321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 23 15:58:35 2021
Response via : Initial Calibration



TIC: BF123418.D\data.ms

(77) Benzidine

12.792min (+ 0.006) 0.00 ng

response 1145840

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.70	14.55
183.00	12.10	11.50
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
03/23/2021	BNA_F	<u>BF123418.D</u>
Compound Name	Response	Retention Time
DDT	402138	13.621
DDD	95093	13.316
DDE	4926	12.98
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
100019	502157	19.92