

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF032415\
 Data File : BF077924.D
 Acq On : 24 Mar 2015 21:16
 Operator : TP/IZ
 Sample : G1613-18
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-12

Quant Time: Mar 25 01:58:55 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF031115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 25 00:51:55 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.12	152	50322	20.00	ng	0.00
21) Naphthalene-d8	8.69	136	198622	20.00	ng	0.00
38) Acenaphthene-d10	10.86	164	98295	20.00	ng	0.00
63) Phenanthrene-d10	12.69	188	207666	20.00	ng	0.00
75) Chrysene-d12	15.97	240	226771	20.00	ng	0.00
86) Perylene-d12	17.70	264	207592	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.40	112	279080	96.80	ng	0.00
7) Phenol-d6	6.69	99	374165	105.05	ng	0.00
23) Nitrobenzene-d5	7.81	82	204835	67.60	ng	0.00
41) 2,4,6-Tribromophenol	11.84	330	85929	91.37	ng	0.00
44) 2-Fluorobiphenyl	10.04	172	469375	70.18	ng	0.00
78) Terphenyl-d14	14.69	244	582554	62.75	ng	0.00
Target Compounds						
49) Dimethylphthalate	10.54	163	15724	2.16	ng	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF032415\
Data File : BF077924.D
Acq On : 24 Mar 2015 21:16
Operator : TP/IZ
Sample : G1613-18
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-12

Quant Time: Mar 25 01:58:55 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF031115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 25 00:51:55 2015
Response via : Initial Calibration

