

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF032415\
 Data File : BF077931.D
 Acq On : 25 Mar 2015 2:40
 Operator : TP/IZ
 Sample : G1613-01 5X
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-1-1

Quant Time: Mar 25 04:58:29 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF031115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 25 04:36:48 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.12	152	39120	20.00	ng	0.00
21) Naphthalene-d8	8.69	136	163906	20.00	ng	0.00
38) Acenaphthene-d10	10.86	164	83761	20.00	ng	0.00
63) Phenanthrene-d10	12.69	188	169785	20.00	ng	0.00
75) Chrysene-d12	15.97	240	188496	20.00	ng	-0.01
86) Perylene-d12	17.69	264	175689	20.00	ng	-0.06
System Monitoring Compounds						
5) 2-Fluorophenol	5.40	112	43645	19.47	ng	0.00
7) Phenol-d6	6.69	99	57114	20.63	ng	0.00
23) Nitrobenzene-d5	7.81	82	29952	11.98	ng	0.00
41) 2,4,6-Tribromophenol	11.84	330	12747	15.91	ng	0.00
44) 2-Fluorobiphenyl	10.04	172	83120	14.58	ng	0.00
78) Terphenyl-d14	14.68	244	102463	13.28	ng	-0.01

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF032415\
Data File : BF077931.D
Acq On : 25 Mar 2015 2:40
Operator : TP/IZ
Sample : G1613-01 5X
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-1-1

Quant Time: Mar 25 04:58:29 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF031115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 25 04:36:48 2015
Response via : Initial Calibration

