

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF032415\
 Data File : BF077929.D
 Acq On : 25 Mar 2015 1:43
 Operator : TP/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 25 02:38:09 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF031115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 25 00:51:55 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
2	1,4-Dioxane	0.520	0.493	5.2	93	0.00
3	Pyridine	1.398	1.141	18.4	83	0.00
4	n-Nitrosodimethylamine	0.609	0.532	12.6	80	0.00
5 S	2-Fluorophenol	1.146	1.114	2.8	99	0.00
6	Aniline	2.025	1.963	3.1	99	0.00
7 S	Phenol-d6	1.416	1.412	0.3	99	0.00
8	2-Chlorophenol	1.364	1.303	4.5	99	0.00
9	Benzaldehyde	0.580	0.660	-13.8	113	0.00
10 C	Phenol	1.673	1.644	1.7	100	0.00
11	bis(2-Chloroethyl)ether	1.253	1.215	3.0	97	0.00
12	1,3-Dichlorobenzene	1.517	1.468	3.2	99	0.00
13 C	1,4-Dichlorobenzene	1.576	1.535	2.6	102	0.00
14	1,2-Dichlorobenzene	1.445	1.422	1.6	102	0.00
15	Benzyl Alcohol	1.105	1.010	8.6	91	0.00
16	2,2'-oxybis(1-Chloropropane	1.689	1.651	2.2	99	0.00
17	2-Methylphenol	1.110	1.089	1.9	98	0.00
18	Hexachloroethane	0.517	0.518	-0.2	105	0.00
19 P	n-Nitroso-di-n-propylamine	0.979	0.959	2.0	100	0.00
20	3+4-Methylphenols	1.457	1.409	3.3	99	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	99	0.00
22	Acetophenone	0.443	0.440	0.7	100	0.00
23 S	Nitrobenzene-d5	0.305	0.309	-1.3	104	0.00
24	Nitrobenzene	0.329	0.343	-4.3	110	0.00
25	Isophorone	0.616	0.597	3.1	97	0.00
26 C	2-Nitrophenol	0.161	0.150	6.8	88	0.00
27	2,4-Dimethylphenol	0.293	0.276	5.8	94	0.00
28	bis(2-Chloroethoxy)methane	0.374	0.374	0.0	102	0.00
29 C	2,4-Dichlorophenol	0.279	0.264	5.4	94	0.00
30	1,2,4-Trichlorobenzene	0.313	0.297	5.1	95	0.00
31	Naphthalene	1.027	0.982	4.4	97	0.00
32	Benzoic acid	0.141	0.123	12.8	79	0.00
33	4-Chloroaniline	0.417	0.393	5.8	93	0.00
34 C	Hexachlorobutadiene	0.177	0.175	1.1	101	0.00
35	Caprolactam	0.082	0.077	6.1	93	0.00
36 C	4-Chloro-3-methylphenol	0.281	0.284	-1.1	103	0.00
37	2-Methylnaphthalene	0.690	0.653	5.4	93	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	97	0.00
39	1,2,4,5-Tetrachlorobenzene	0.597	0.560	6.2	92	0.00
40 P	Hexachlorocyclopentadiene	0.254	0.263	-3.5	92	0.00
41 S	2,4,6-Tribromophenol	0.191	0.173	9.4	87	0.00
42 C	2,4,6-Trichlorophenol	0.413	0.381	7.7	86	0.00
43	2,4,5-Trichlorophenol	0.446	0.431	3.4	95	0.00
44 S	2-Fluorobiphenyl	1.361	1.351	0.7	100	0.00

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF032415\
 Data File : BF077929.D
 Acq On : 25 Mar 2015 1:43
 Operator : TP/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 25 02:38:09 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF031115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 25 00:51:55 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.599	1.562	2.3	94	0.00
46	2-Chloronaphthalene	1.299	1.290	0.7	99	0.00
47	2-Nitroaniline	0.323	0.330	-2.2	94	0.00
48	Acenaphthylene	2.161	2.127	1.6	98	0.00
49	Dimethylphthalate	1.483	1.497	-0.9	100	0.00
50	2,6-Dinitrotoluene	0.307	0.322	-4.9	101	0.00
51 C	Acenaphthene	1.239	1.160	6.4	91	0.00
52	3-Nitroaniline	0.357	0.358	-0.3	95	0.00
53 P	2,4-Dinitrophenol	0.093	0.087	6.5	90	0.00
54	Dibenzofuran	1.837	1.727	6.0	92	0.00
55 P	4-Nitrophenol	0.265	0.234	11.7	80	0.00
56	2,4-Dinitrotoluene	0.401	0.397	1.0	96	0.00
57	Fluorene	1.526	1.487	2.6	94	0.00
58	2,3,4,6-Tetrachlorophenol	0.344	0.332	3.5	93	0.00
59	Diethylphthalate	1.445	1.398	3.3	94	0.00
60	4-Chlorophenyl-phenylether	0.696	0.647	7.0	92	0.00
61	4-Nitroaniline	0.356	0.376	-5.6	101	0.00
62	Azobenzene	1.381	1.392	-0.8	91	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	97	0.00
64	4,6-Dinitro-2-methylphenol	0.078	0.089	-14.1	97	0.00
65 c	n-Nitrosodiphenylamine	0.666	0.672	-0.9	96	0.00
66	4-Bromophenyl-phenylether	0.213	0.211	0.9	94	0.00
67	Hexachlorobenzene	0.235	0.233	0.9	95	0.00
68	Atrazine	0.187	0.168	10.2	83	0.00
69 C	Pentachlorophenol	0.104	0.102	1.9	87	0.00
70	Phenanthrene	1.148	1.179	-2.7	100	0.00
71	Anthracene	1.134	1.170	-3.2	103	0.00
72	Carbazole	1.079	1.132	-4.9	106	0.00
73	Di-n-butylphthalate	1.210	1.242	-2.6	100	0.00
74 C	Fluoranthene	1.266	1.301	-2.8	94	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	102	0.00
76	Benzidine	0.546	0.491	10.1	98	0.00
77	Pyrene	1.258	1.217	3.3	91	0.00
78 S	Terphenyl-d14	0.819	0.763	6.8	90	0.00
79	Butylbenzylphthalate	0.496	0.477	3.8	97	0.00
80	Benzo(a)anthracene	1.186	1.189	-0.3	106	0.00
81	3,3'-Dichlorobenzidine	0.391	0.368	5.9	101	0.00
82	Chrysene	1.066	1.094	-2.6	105	0.00
83	Bis(2-ethylhexyl)phthalate	0.751	0.741	1.3	103	0.00
84 c	Di-n-octyl phthalate	1.284	1.183	7.9	97	0.00
85	Indeno(1,2,3-cd)pyrene	1.331	1.314	1.3	109	0.00
86 I	Perylene-d12	1.000	1.000	0.0	105	0.00
87	Benzo(b)fluoranthene	1.306	1.124	13.9	91	0.00

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF032415\
Data File : BF077929.D
Acq On : 25 Mar 2015 1:43
Operator : TP/IZ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Mar 25 02:38:09 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF031115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 25 00:51:55 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.020	1.185	-16.2	128	0.00
89 C	Benzo(a)pyrene	1.089	1.109	-1.8	108	0.00
90	Dibenzo(a,h)anthracene	1.081	1.093	-1.1	108	0.00
91	Benzo(g,h,i)perylene	1.100	1.097	0.3	105	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0