

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF032416\
 Data File : BF085706.D
 Acq On : 24 Mar 2016 10:23
 Operator : UM/SJ
 Sample : SSTDIC010
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDIC010

Quant Time: Mar 24 11:51:06 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF032416.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 24 11:35:38 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.82	152	110771	20.00	ng	0.00
21) Naphthalene-d8	8.10	136	426810	20.00	ng	-0.01
38) Acenaphthene-d10	9.86	164	218641	20.00	ng	0.00
63) Phenanthrene-d10	11.34	188	406190	20.00	ng	-0.01
75) Chrysene-d12	13.98	240	331160	20.00	ng	0.00
86) Perylene-d12	15.40	264	221218	20.00	ng	0.00

System Monitoring Compounds

5) 2-Fluorophenol	5.41	112	140209	21.26	ng	0.00
7) Phenol-d6	6.45	99	176273	20.84	ng	-0.01
23) Nitrobenzene-d5	7.38	82	163565	22.24	ng	-0.01
41) 2,4,6-Tribromophenol	10.65	330	39122	18.27	ng	0.00
44) 2-Fluorobiphenyl	9.19	172	297762	21.78	ng	0.00
78) Terphenyl-d14	12.93	244	326713	23.74	ng	0.00

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) 1,4-Dioxane	2.39	88	31335	9.77	ng	99
3) Pyridine	3.14	79	70527	8.96	ng	97
4) n-Nitrosodimethylamine	3.05	42	25582	7.78	ng	97
6) Aniline	6.48	93	118535	10.43	ng	98
8) 2-Chlorophenol	6.61	128	78224	10.25	ng	94
9) Benzaldehyde	6.37	77	55272	13.00	ng	99
10) Phenol	6.46	94	97638	10.13	ng	96
11) bis(2-Chloroethyl)ether	6.55	93	73347	10.10	ng	89
12) 1,3-Dichlorobenzene	6.77	146	85307	10.26	ng	99
13) 1,4-Dichlorobenzene	6.84	146	86138	10.16	ng	98
14) 1,2-Dichlorobenzene	7.00	146	84290	11.19	ng	98
15) Benzyl Alcohol	6.96	79	59697	10.30	ng	95
16) 2,2'-oxybis(1-Chloropropan	7.10	45	102292	10.71	ng	100
17) 2-Methylphenol	7.08	107	61928	9.79	ng	95
18) Hexachloroethane	7.34	117	31728	10.36	ng	99
19) n-Nitroso-di-n-propylamine	7.24	70	56189	11.12	ng	97
20) 3+4-Methylphenols	7.24	107	83638	11.42	ng	# 87
22) Acetophenone	7.24	105	119515	11.80	ng	# 94
24) Nitrobenzene	7.41	77	85855	10.67	ng	91
25) Isophorone	7.65	82	161594	10.98	ng	96
26) 2-Nitrophenol	7.73	139	41708	10.49	ng	96
27) 2,4-Dimethylphenol	7.76	122	71554	10.17	ng	98
28) bis(2-Chloroethoxy)methane	7.86	93	87607	10.46	ng	99
29) 2,4-Dichlorophenol	7.97	162	63315	10.90	ng	99
30) 1,2,4-Trichlorobenzene	8.05	180	66418	10.21	ng	98
31) Naphthalene	8.13	128	236263	10.97	ng	98
32) Benzoic acid	7.85	122	45473	7.73	ng	95
33) 4-Chloroaniline	8.18	127	99751	11.30	ng	97
34) Hexachlorobutadiene	8.25	225	37783	10.90	ng	98
35) Caprolactam	8.53	113	21208	10.61	ng	95
36) 4-Chloro-3-methylphenol	8.66	107	75113	11.07	ng	98
37) 2-Methylnaphthalene	8.82	142	159795	12.17	ng	99
39) 1,2,4,5-Tetrachlorobenzene	9.00	216	61838	9.32	ng	100
40) Hexachlorocyclopentadiene	8.97	237	11339	3.79	ng	99

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF032416\
 Data File : BF085706.D
 Acq On : 24 Mar 2016 10:23
 Operator : UM/SJ
 Sample : SSTDIC010
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDIC010

Quant Time: Mar 24 11:51:06 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF032416.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 24 11:35:38 2016
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
42) 2,4,6-Trichlorophenol	9.10	196	41118	9.06	ng	96
43) 2,4,5-Trichlorophenol	9.14	196	47098	10.19	ng #	94
45) 1,1'-Biphenyl	9.28	154	192442	10.20	ng	99
46) 2-Chloronaphthalene	9.30	162	136294	9.89	ng #	89
47) 2-Nitroaniline	9.41	65	43074	10.33	ng	99
48) Acenaphthylene	9.73	152	234664	10.52	ng	98
49) Dimethylphthalate	9.58	163	160540	9.73	ng #	99
50) 2,6-Dinitrotoluene	9.65	165	38007	10.91	ng	87
51) Acenaphthene	9.90	154	140849	10.07	ng	100
52) 3-Nitroaniline	9.82	138	41769	9.57	ng	83
53) 2,4-Dinitrophenol	9.93	184	8503	3.98	ng #	78
54) Dibenzofuran	10.07	168	191832	11.26	ng	96
55) 4-Nitrophenol	9.98	139	26218	7.78	ng	88
56) 2,4-Dinitrotoluene	10.05	165	45511	9.98	ng #	43
57) Fluorene	10.41	166	156136	11.00	ng	97
58) 2,3,4,6-Tetrachlorophenol	10.18	232	30748	9.24	ng	96
59) Diethylphthalate	10.29	149	166974	10.18	ng	96
60) 4-Chlorophenyl-phenylether	10.40	204	80796	11.65	ng	95
61) 4-Nitroaniline	10.42	138	40243	9.36	ng	94
62) Azobenzene	10.56	77	161983	10.29	ng	96
64) 4,6-Dinitro-2-methylphenol	10.46	198	20408	7.85	ng #	77
65) n-Nitrosodiphenylamine	10.52	169	139184	11.00	ng	99
66) 4-Bromophenyl-phenylether	10.89	248	44244	10.90	ng	94
67) Hexachlorobenzene	10.96	284	45480	10.55	ng	92
68) Atrazine	11.04	200	45198	11.73	ng	95
69) Pentachlorophenol	11.16	266	19028	7.27	ng	99
70) Phenanthrene	11.36	178	241173	11.26	ng	99
71) Anthracene	11.42	178	261500	11.64	ng	98
72) Carbazole	11.58	167	214360	11.06	ng	97
73) Di-n-butylphthalate	11.91	149	283476	11.68	ng	98
74) Fluoranthene	12.55	202	246983	11.01	ng	95
76) Benzidine	12.68	184	102606	9.21	ng	98
77) Pyrene	12.78	202	258679	10.88	ng	99
79) Butylbenzylphthalate	13.40	149	105640	9.32	ng #	67
80) Benzo(a)anthracene	13.97	228	187003	9.73	ng	99
81) 3,3'-Dichlorobenzidine	13.93	252	127047	18.03	ng #	96
82) Chrysene	14.00	228	196428	10.62	ng	100
83) Bis(2-ethylhexyl)phthalate	13.96	149	138720	9.85	ng	99
84) Di-n-octyl phthalate	14.57	149	191787	8.36	ng	98
85) Indeno(1,2,3-cd)pyrene	16.77	276	122262	7.91	ng	99
87) Benzo(b)fluoranthene	14.99	252	283381	19.63	ng	99
88) Benzo(k)fluoranthene	14.99	252	283381	23.83	ng #	97
89) Benzo(a)pyrene	15.34	252	126312	10.32	ng #	96
90) Dibenzo(a,h)anthracene	16.78	278	102255	10.02	ng	97
91) Benzo(g,h,i)perylene	17.19	276	100764	10.20	ng	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF032416\
 Data File : BF085706.D
 Acq On : 24 Mar 2016 10:23
 Operator : UM/SJ
 Sample : SSTDICC010
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_F
 Client Sampled :
 SSTDICC010

Quant Time: Mar 24 11:51:06 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF032416.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 24 11:35:38 2016
 Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 20.00 ng
 RT: 6.82 min Scan# 397
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23




Instrument :
 BNA_F
ClientSampled :
 SSTDICC010


Tgt Ion	Resp	Lower	Upper
152	110771		
150	158.3	124.2	186.2
115	61.5	47.0	70.6



#2
 1,4-Dioxane
 Concen: 9.77 ng
 RT: 2.39 min Scan# 9
 Delta R.T. 0.01 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Tgt Ion	Resp	Lower	Upper
88	31335		
58	68.0	54.2	81.2
43	25.4	19.3	28.9








#3
 Pyridine
 Concen: 8.96 ng
 RT: 3.14 min Scan# 75
 Delta R.T. 0.05 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23





Instrument :
 BNA_F
ClientSampleId :
 SSTDICC010


Tgt Ion	Resp	Lower	Upper
79	70527		
52	63.8	49.8	74.6
51	33.8	24.7	37.1



#4
 n-Nitrosodimethylamine
 Concen: 7.78 ng
 RT: 3.05 min Scan# 67
 Delta R.T. 0.01 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Tgt Ion	Resp	Lower	Upper
42	25582		
74	156.1	121.4	182.2
44	7.2	5.6	8.4









#5
 2-Fluorophenol
 Concen: 21.26 ng
 RT: 5.41 min Scan# 273
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC010


Tgt Ion	Resp	Lower	Upper
112	140209		
64	59.4	39.9	59.9
63	30.7	20.2	30.2#



#6
 Aniline
 Concen: 10.43 ng
 RT: 6.48 min Scan# 367
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Tgt Ion	Resp	Lower	Upper
93	118535		
66	38.1	31.4	47.2
65	19.8	15.7	23.5









#7
 Phenol-d6
 Concen: 20.84 ng
 RT: 6.45 min Scan# 364
 Delta R.T. -0.01 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23




Instrument :
 BNA_F
 ClientSampleID :
 SSTDICC010


Tgt Ion	Resp	Lower	Upper
99	176273		
42	16.7	11.1	16.7#
71	32.8	24.9	37.3



#8
 2-Chlorophenol
 Concen: 10.25 ng
 RT: 6.61 min Scan# 378
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Tgt Ion	Resp	Lower	Upper
128	78224		
130	32.8	12.9	52.9
64	33.6	20.2	60.2









#9
Benzaldehyde
Concen: 13.00 ng
RT: 6.37 min Scan# 357
Delta R.T. -0.00 min
Lab File: BF085706.D
Acq: 24 Mar 2016 10:23




Instrument :
BNA_F
ClientSampled :
SSTDIC010


Tgt Ion	Resp	Lower	Upper
77	100		
105	98.9	80.1	120.1
106	94.9	75.0	115.0



#10
Phenol
Concen: 10.13 ng
RT: 6.46 min Scan# 365
Delta R.T. -0.01 min
Lab File: BF085706.D
Acq: 24 Mar 2016 10:23

Tgt Ion	Resp	Lower	Upper
94	100		
65	30.6	8.6	48.6
66	42.8	20.0	60.0








#11
 bis(2-Chloroethyl)ether
 Concen: 10.10 ng
 RT: 6.55 min Scan# 373
 Delta R.T. -0.01 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23





Instrument :
 BNA_F
ClientSampled :
 SSTDICC010


Tgt Ion	Resp	Lower	Upper
93	73347		
93	100		
63	83.4	50.9	90.9
95	31.9	13.3	53.3



#12
 1,3-Dichlorobenzene
 Concen: 10.26 ng
 RT: 6.77 min Scan# 392
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Tgt Ion	Resp	Lower	Upper
146	85307		
146	100		
148	66.1	52.3	78.5
75	22.6	17.4	26.2









#13
 1,4-Dichlorobenzene
 Concen: 10.16 ng
 RT: 6.84 min Scan# 398
 Delta R.T. -0.01 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23




Instrument :
 BNA_F
 ClientSampleID :
 SSTDICC010


Tgt Ion	Resp	Lower	Upper
146	86138		
148	60.0	50.6	76.0
111	47.2	37.8	56.6



#14
 1,2-Dichlorobenzene
 Concen: 11.19 ng
 RT: 7.00 min Scan# 412
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Tgt Ion	Resp	Lower	Upper
146	84290		
148	64.1	51.4	77.2
111	40.5	34.6	52.0









#15
 Benzyl Alcohol
 Concen: 10.30 ng
 RT: 6.96 min Scan# 409
 Delta R.T. -0.01 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC010


Tgt Ion	Resp	Lower	Upper
79	100		
108	70.2	61.8	92.6
77	61.7	49.6	74.4



#16
 2,2'-oxybis(1-Chloropropane)
 Concen: 10.71 ng
 RT: 7.10 min Scan# 421
 Delta R.T. -0.01 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Tgt Ion	Resp	Lower	Upper
45	100		
77	15.9	0.0	35.7
79	12.4	0.0	32.2









#17
 2-Methylphenol
 Concen: 9.79 ng
 RT: 7.08 min Scan# 419
 Delta R.T. -0.01 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC010


Tgt Ion	Resp	Lower	Upper
107	100		
108	117.3	91.4	137.0
77	50.8	36.2	54.2
79	48.6	35.8	53.8



#18
 Hexachloroethane
 Concen: 10.36 ng
 RT: 7.34 min Scan# 442
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Tgt Ion	Resp	Lower	Upper
117	100		
119	95.0	76.7	115.1
201	103.6	83.3	124.9









#19
 n-Nitroso-di-n-propylamine
 Concen: 11.12 ng
 RT: 7.24 min Scan# 433
 Delta R.T. -0.01 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23




Instrument :
 BNA_F
ClientSampled :
 SSTDICC010


Tgt Ion	Resp	Lower	Upper
70	56189		
70	100		
42	49.2	37.1	55.7
101	10.5	8.6	12.8
130	23.3	19.6	29.4



#20
 3+4-Methylphenols
 Concen: 11.42 ng
 RT: 7.24 min Scan# 433
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Tgt Ion	Resp	Lower	Upper
107	83638		
107	100		
108	88.6	69.3	109.3
77	149.9	101.7	141.7#
79	27.8	8.3	48.3









#21
 Naphthalene-d8
 Concen: 20.00 ng
 RT: 8.10 min Scan# 509
 Delta R.T. -0.01 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Instrument :
 BNA_F
ClientSampled :
 SSTDICC010

Tgt Ion:136 Resp: 426810




Ion	Ratio	Lower	Upper
136	100		
137	11.6	9.1	13.7
54	9.4	7.4	11.0
68	7.3	5.8	8.8




#22
 Acetophenone
 Concen: 11.80 ng
 RT: 7.24 min Scan# 433
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Tgt Ion:105 Resp: 119515

Ion	Ratio	Lower	Upper
105	100		
71	7.4	3.6	5.4#
51	26.6	25.4	38.0
120	21.0	16.9	25.3









#23
 Nitrobenzene-d5
 Concen: 22.24 ng
 RT: 7.38 min Scan# 446
 Delta R.T. -0.01 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23




Instrument :
 BNA_F
 ClientSampleID :
 SSTDICC010


Tgt Ion	Resp	Lower	Upper
82	163565		
128	41.6	41.8	62.8#
54	47.3	33.4	50.0



#24
 Nitrobenzene
 Concen: 10.67 ng
 RT: 7.41 min Scan# 448
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Tgt Ion	Resp	Lower	Upper
77	85855		
123	50.8	35.0	52.4
65	12.6	10.6	16.0









#25
 Isophorone
 Concen: 10.98 ng
 RT: 7.65 min Scan# 469
 Delta R.T. -0.01 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23




Instrument :
 BNA_F
 ClientSampled :
 SSTDICC010


Tgt Ion	Resp	Lower	Upper
82	161594		
95	6.9	5.4	8.2
138	18.2	12.9	19.3



#26
 2-Nitrophenol
 Concen: 10.49 ng
 RT: 7.73 min Scan# 476
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Tgt Ion	Resp	Lower	Upper
139	41708		
109	28.2	22.1	33.1
65	46.5	33.9	50.9








#27
 2,4-Dimethylphenol
 Concen: 10.17 ng
 RT: 7.76 min Scan# 479
 Delta R.T. -0.01 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23





Instrument :
 BNA_F
ClientSampled :
 SSTDICC010


Tgt Ion	Resp	Lower	Upper
122	71554		
122	100		
107	118.5	93.0	139.6
121	58.1	45.9	68.9



#28
 bis(2-Chloroethoxy)methane
 Concen: 10.46 ng
 RT: 7.86 min Scan# 488
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Tgt Ion	Resp	Lower	Upper
93	87607		
93	100		
95	32.0	26.4	39.6
123	12.5	9.8	14.6









#29
 2,4-Dichlorophenol
 Concen: 10.90 ng
 RT: 7.97 min Scan# 497
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23




Instrument :
 BNA_F
 ClientSampleID :
 SSTDICC010


Tgt Ion	Resp	Lower	Upper
162	63315		
164	100	43.7	83.7
98	42.6	20.4	60.4



#30
 1,2,4-Trichlorobenzene
 Concen: 10.21 ng
 RT: 8.05 min Scan# 504
 Delta R.T. -0.01 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Tgt Ion	Resp	Lower	Upper
180	66418		
182	91.7	73.8	110.8
145	38.3	28.4	42.6









#31
 Naphthalene
 Concen: 10.97 ng
 RT: 8.13 min Scan# 511
 Delta R.T. -0.01 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Instrument :
 BNA_F
ClientSampled :
 SSTDICC010

Tgt Ion:128 Resp: 236263




Ion	Ratio	Lower	Upper
128	100		
129	11.1	9.4	14.0
127	12.9	11.0	16.6




#32
 Benzoic acid
 Concen: 7.73 ng
 RT: 7.85 min Scan# 487
 Delta R.T. -0.05 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Tgt Ion:122 Resp: 45473

Ion	Ratio	Lower	Upper
122	100		
105	117.6	102.9	142.9
77	86.3	71.8	111.8









#33
 4-Chloroaniline
 Concen: 11.30 ng
 RT: 8.18 min Scan# 516
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23




Instrument :
 BNA_F
ClientSampled :
 SSTDICC010


Tgt Ion	Resp	Lower	Upper
127	100		
129	31.6	26.1	39.1
65	26.4	19.5	29.3
92	17.3	14.6	22.0



#34
 Hexachlorobutadiene
 Concen: 10.90 ng
 RT: 8.25 min Scan# 522
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Tgt Ion	Resp	Lower	Upper
225	100		
223	61.9	50.4	75.6
227	62.3	51.4	77.2









#35
 Caprolactam
 Concen: 10.61 ng
 RT: 8.53 min Scan# 546
 Delta R.T. -0.03 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23




Instrument :
 BNA_F
ClientSampled :
 SSTDICC010


Tgt Ion	Resp	Lower	Upper
113	100		
55	152.3	139.4	179.4
56	115.9	100.4	140.4



#36
 4-Chloro-3-methylphenol
 Concen: 11.07 ng
 RT: 8.66 min Scan# 558
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Tgt Ion	Resp	Lower	Upper
107	100		
144	24.7	19.3	28.9
142	79.1	61.4	92.0









#37
 2-Methylnaphthalene
 Concen: 12.17 ng
 RT: 8.82 min Scan# 572
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23




Instrument :
 BNA_F
 ClientSampleID :
 SSTDICC010


Tgt Ion	Resp	Lower	Upper
142	159795		
141	88.6	71.6	107.4
115	32.7	26.8	40.2



#38
 Acenaphthene-d10
 Concen: 20.00 ng
 RT: 9.86 min Scan# 663
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Tgt Ion	Resp	Lower	Upper
164	218641		
162	104.3	82.1	123.1
160	47.0	37.4	56.2









#39
 1,2,4,5-Tetrachlorobenzene
 Concen: 9.32 ng
 RT: 9.00 min Scan# 587
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23




Instrument :
 BNA_F
 ClientSampled :
 SSTDICC010


Tgt Ion	Resp	Lower	Upper
216	61838		
214	78.8	63.1	94.7
179	23.0	18.2	27.2
108	22.3	17.3	25.9



#40
 Hexachlorocyclopentadiene
 Concen: 3.79 ng
 RT: 8.97 min Scan# 585
 Delta R.T. -0.01 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Tgt Ion	Resp	Lower	Upper
237	11339		
235	62.9	43.7	83.7
272	11.4	0.0	31.9









#41
 2,4,6-Tribromophenol
 Concen: 18.27 ng
 RT: 10.65 min Scan# 732
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23




Instrument :
 BNA_F
ClientSampled :
 SSTDICC010


Tgt Ion	Resp	Lower	Upper
330	100		
332	95.7	78.2	117.2
141	45.9	31.5	47.3



#42
 2,4,6-Trichlorophenol
 Concen: 9.06 ng
 RT: 9.10 min Scan# 596
 Delta R.T. -0.01 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Tgt Ion	Resp	Lower	Upper
196	100		
198	90.0	75.3	112.9
200	28.6	23.7	35.5









#43
 2,4,5-Trichlorophenol
 Concen: 10.19 ng
 RT: 9.14 min Scan# 600
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Instrument :
 BNA_F
ClientSampleId :
 SSTDICC010

Tgt Ion	Resp	Lower	Upper
196	47098		
198	93.6	76.6	114.8
97	48.4	30.5	45.7#
132	28.1	20.2	30.2




Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 97.00 (96.70 to 97.70): BF0
 Ion 132.00 (131.70 to 132.70): E






#44
 2-Fluorobiphenyl
 Concen: 21.78 ng
 RT: 9.19 min Scan# 604
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Tgt Ion	Resp	Lower	Upper
172	297762		
171	33.9	29.4	44.2
170	23.1	19.8	29.6



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E









#45
 1,1'-Biphenyl
 Concen: 10.20 ng
 RT: 9.28 min Scan# 612
 Delta R.T. -0.01 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23




Instrument :
 BNA_F
 ClientSampled :
 SSTDICC010


Tgt Ion	Resp	Lower	Upper
154	192442		
153	40.6	20.8	60.8
76	15.3	0.0	36.7



#46
 2-Chloronaphthalene
 Concen: 9.89 ng
 RT: 9.30 min Scan# 614
 Delta R.T. -0.01 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Tgt Ion	Resp	Lower	Upper
162	136294		
127	50.4	31.7	47.5#
164	32.4	27.3	40.9









#47
 2-Nitroaniline
 Concen: 10.33 ng
 RT: 9.41 min Scan# 623
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23




Instrument :
 BNA_F
ClientSampled :
 SSTDICC010


Tgt Ion	Resp	Lower	Upper
65	43074		
92	73.5	59.7	89.5
138	113.8	91.4	137.0



#48
 Acenaphthylene
 Concen: 10.52 ng
 RT: 9.73 min Scan# 651
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Tgt Ion	Resp	Lower	Upper
152	234664		
151	19.8	16.8	25.2
153	12.8	10.9	16.3









#49
 Dimethylphthalate
 Concen: 9.73 ng
 RT: 9.58 min Scan# 638
 Delta R.T. -0.01 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23




Instrument :
 BNA_F
 ClientSampleID :
 SSTDICC010


Tgt Ion	Resp	Lower	Upper
163	100		
194	2.9	2.9	4.3#
164	10.6	8.3	12.5



#50
 2,6-Dinitrotoluene
 Concen: 10.91 ng
 RT: 9.65 min Scan# 644
 Delta R.T. -0.01 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Tgt Ion	Resp	Lower	Upper
165	100		
63	52.6	51.8	77.8
89	58.7	53.7	80.5









#51
 Acenaphthene
 Concen: 10.07 ng
 RT: 9.90 min Scan# 666
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC010


Tgt Ion	Resp	Lower	Upper
154	100		
153	112.8	90.1	135.1
152	54.9	44.2	66.2



#52
 3-Nitroaniline
 Concen: 9.57 ng
 RT: 9.82 min Scan# 659
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Tgt Ion	Resp	Lower	Upper
138	100		
108	10.2	8.7	13.1
92	96.9	93.3	139.9









#53
 2,4-Dinitrophenol
 Concen: 3.98 ng
 RT: 9.93 min Scan# 669
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23




Instrument :
 BNA_F
ClientSampled :
 SSTDICC010


Tgt Ion	Resp	Lower	Upper
184	100		
63	55.2	69.5	104.3#
154	65.9	58.3	87.5



#54
 Dibenzofuran
 Concen: 11.26 ng
 RT: 10.07 min Scan# 681
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Tgt Ion	Resp	Lower	Upper
168	100		
139	37.2	31.9	47.9
169	12.8	11.0	16.6








#55
 4-Nitrophenol
 Concen: 7.78 ng
 RT: 9.98 min Scan# 673
 Delta R.T. -0.01 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23




Instrument :
 BNA_F
ClientSampled :
 SSTDICC010


Tgt Ion	Resp	Lower	Upper
139	100		
109	75.5	49.2	89.2
65	101.7	66.9	106.9



#56
 2,4-Dinitrotoluene
 Concen: 9.98 ng
 RT: 10.05 min Scan# 679
 Delta R.T. -0.01 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Tgt Ion	Resp	Lower	Upper
165	100		
63	56.9	24.9	37.3#
89	97.9	41.0	61.6#
182	2.5	2.1	3.1








#57
 Fluorene
 Concen: 11.00 ng
 RT: 10.41 min Scan# 711
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23





Instrument :
 BNA_F
 ClientSampled :
 SSTDICC010


Tgt Ion	Resp	Lower	Upper
166	156136		
165	102.2	79.4	119.0
167	13.5	11.0	16.4



#58
 2,3,4,6-Tetrachlorophenol
 Concen: 9.24 ng
 RT: 10.18 min Scan# 691
 Delta R.T. -0.01 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Tgt Ion	Resp	Lower	Upper
232	30748		
231	59.1	44.9	67.3
130	2.5	2.3	3.5
166	36.9	28.1	42.1









#59
 Diethylphthalate
 Concen: 10.18 ng
 RT: 10.29 min Scan# 700
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23




Instrument :
 BNA_F
ClientSampled :
 SSTDICC010


Tgt Ion	Resp	Lower	Upper
149	166974		
177	24.6	17.9	26.9
150	11.8	10.2	15.2



#60
 4-Chlorophenyl-phenylether
 Concen: 11.65 ng
 RT: 10.40 min Scan# 710
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Tgt Ion	Resp	Lower	Upper
204	80796		
206	32.9	26.6	39.8
141	66.2	57.4	86.0








#61
 4-Nitroaniline
 Concen: 9.36 ng
 RT: 10.42 min Scan# 712
 Delta R.T. -0.01 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Instrument :
 BNA_F
ClientSampleId :
 SSTDICC010

Tgt Ion: 138 Resp: 40243



Ion	Ratio	Lower	Upper
138	100		
92	48.8	26.4	66.4
108	75.3	49.0	89.0




#62
 Azobenzene
 Concen: 10.29 ng
 RT: 10.56 min Scan# 724
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Tgt Ion: 77 Resp: 161983

Ion	Ratio	Lower	Upper
77	100		
182	25.5	1.9	41.9
105	24.1	3.5	43.5
51	24.8	6.3	46.3









#63
 Phenanthrene-d10
 Concen: 20.00 ng
 RT: 11.34 min Scan# 792
 Delta R.T. -0.01 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23




Instrument :
 BNA_F
 ClientSampled :
 SSTDICC010


Tgt Ion	Resp	Lower	Upper
188	406190		
94	13.1	5.4	8.0#
80	14.8	5.5	8.3#



#64
 4,6-Dinitro-2-methylphenol
 Concen: 7.85 ng
 RT: 10.46 min Scan# 715
 Delta R.T. -0.01 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Tgt Ion	Resp	Lower	Upper
198	20408		
51	44.8	45.4	85.4#
105	50.1	45.9	85.9









#65
 n-Nitrosodiphenylamine
 Concen: 11.00 ng
 RT: 10.52 min Scan# 720
 Delta R.T. -0.01 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC010


Tgt Ion	Resp	Lower	Upper
169	139184		
168	66.6	54.4	81.6
167	37.1	29.4	44.0



#66
 4-Bromophenyl-phenylether
 Concen: 10.90 ng
 RT: 10.89 min Scan# 753
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Tgt Ion	Resp	Lower	Upper
248	44244		
250	96.2	77.4	116.0
141	67.4	63.6	95.4









#67
 Hexachlorobenzene
 Concen: 10.55 ng
 RT: 10.96 min Scan# 759
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC010


Tgt Ion	Resp	Lower	Upper
284	45480		
142	23.2	23.1	34.7
249	27.5	24.2	36.2



#68
 Atrazine
 Concen: 11.73 ng
 RT: 11.04 min Scan# 766
 Delta R.T. -0.01 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Tgt Ion	Resp	Lower	Upper
200	45198		
173	29.2	7.7	47.7
215	41.4	25.4	65.4









#69
 Pentachlorophenol
 Concen: 7.27 ng
 RT: 11.16 min Scan# 776
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23




Instrument :
 BNA_F
ClientSampled :
 SSTDICC010


Tgt Ion	Resp	Lower	Upper
266	19028		
268	100	51.5	77.3
264	63.0	50.6	76.0



#70
 Phenanthrene
 Concen: 11.26 ng
 RT: 11.36 min Scan# 794
 Delta R.T. -0.01 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Tgt Ion	Resp	Lower	Upper
178	241173		
176	100	16.3	24.5
179	19.7	12.6	18.8









#71
 Anthracene
 Concen: 11.64 ng
 RT: 11.42 min Scan# 799
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Instrument :
 BNA_F
 ClientSampled :
 SSTDICC010

Tgt Ion: 178 Resp: 261500




Ion	Ratio	Lower	Upper
178	100		
176	19.2	15.9	23.9
179	15.3	13.0	19.4




#72
 Carbazole
 Concen: 11.06 ng
 RT: 11.58 min Scan# 813
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Tgt Ion: 167 Resp: 214360

Ion	Ratio	Lower	Upper
167	100		
166	21.1	17.8	26.8
139	12.8	11.6	17.4









#73
 Di-n-butylphthalate
 Concen: 11.68 ng
 RT: 11.91 min Scan# 842
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23




Instrument :
 BNA_F
ClientSampled :
 SSTDICC010


Tgt Ion	Resp	Lower	Upper
149	100		
150	9.0	7.6	11.4
104	4.3	3.8	5.8



#74
 Fluoranthene
 Concen: 11.01 ng
 RT: 12.55 min Scan# 898
 Delta R.T. -0.01 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Tgt Ion	Resp	Lower	Upper
202	100		
101	12.8	0.0	36.6
203	17.3	0.0	38.1









#75
 Chrysene-d12
 Concen: 20.00 ng
 RT: 13.98 min Scan# 1023
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23




Instrument :
 BNA_F
 ClientSampled :
 SSTDICC010


Tgt Ion	Resp	Lower	Upper
240	331160		
120	10.3	13.8	20.8#
236	26.8	20.6	30.8



#76
 Benzidine
 Concen: 9.21 ng
 RT: 12.68 min Scan# 909
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Tgt Ion	Resp	Lower	Upper
184	102606		
185	14.6	12.4	18.6
183	11.3	9.4	14.2









#77
 Pyrene
 Concen: 10.88 ng
 RT: 12.78 min Scan# 918
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC010

Tgt Ion: 202 Resp: 258679




Ion	Ratio	Lower	Upper
202	100		
200	21.6	17.6	26.4
203	17.8	14.0	21.0




#78
 Terphenyl-d14
 Concen: 23.74 ng
 RT: 12.93 min Scan# 931
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Tgt Ion: 244 Resp: 326713

Ion	Ratio	Lower	Upper
244	100		
212	7.0	6.1	9.1
122	8.3	10.2	15.4#









#79
 Butylbenzylphthalate
 Concen: 9.32 ng
 RT: 13.40 min Scan# 972
 Delta R.T. -0.01 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23




Instrument :
 BNA_F
 ClientSampled :
 SSTDICC010


Tgt Ion	Resp	Lower	Upper
149	105640		
91	85.9	45.8	68.8#
206	13.9	17.8	26.8#



#80
 Benzo(a)anthracene
 Concen: 9.73 ng
 RT: 13.97 min Scan# 1022
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Tgt Ion	Resp	Lower	Upper
228	187003		
226	27.8	22.1	33.1
229	19.8	16.2	24.4









#81
 3,3'-Dichlorobenzidine
 Concen: 18.03 ng
 RT: 13.93 min Scan# 1019
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC010


Tgt Ion	Resp	Lower	Upper
252	127047		
254	64.9	50.7	76.1
126	9.6	11.8	17.6#



#82
 Chrysene
 Concen: 10.62 ng
 RT: 14.00 min Scan# 1025
 Delta R.T. -0.01 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Tgt Ion	Resp	Lower	Upper
228	196428		
226	31.0	25.0	37.6
229	19.4	15.7	23.5









#83
 Bis(2-ethylhexyl)phthalate
 Concen: 9.85 ng
 RT: 13.96 min Scan# 1021
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC010


Tgt Ion	Resp	Lower	Upper
149	138720		
167	24.0	19.4	29.2
279	1.8	1.5	2.3



#84
 Di-n-octyl phthalate
 Concen: 8.36 ng
 RT: 14.57 min Scan# 1075
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Tgt Ion	Resp	Lower	Upper
149	191787		
167	1.4	1.2	1.8
43	9.8	7.1	10.7









#85
 Indeno(1,2,3-cd)pyrene
 Concen: 7.91 ng
 RT: 16.77 min Scan# 1267
 Delta R.T. -0.01 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23




Instrument :
 BNA_F
 ClientSampled :
 SSTDICC010


Tgt Ion	Resp	Lower	Upper
276	100		
138	29.3	24.2	36.4
277	24.6	20.0	30.0



#86
 Perylene-d12
 Concen: 20.00 ng
 RT: 15.40 min Scan# 1147
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Tgt Ion	Resp	Lower	Upper
264	100		
260	24.3	19.4	29.2
265	21.2	17.2	25.8









#87
 Benzo(b)fluoranthene
 Concen: 19.63 ng
 RT: 14.99 min Scan# 1111
 Delta R.T. -0.01 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23




Instrument :
 BNA_F
 ClientSampleId :
 SSTDIC010


Tgt Ion	Resp	Lower	Upper
252	100		
253	21.6	17.6	26.4
125	11.9	10.3	15.5



#88
 Benzo(k)fluoranthene
 Concen: 23.83 ng
 RT: 14.99 min Scan# 1111
 Delta R.T. -0.03 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Tgt Ion	Resp	Lower	Upper
252	100		
253	21.6	17.8	26.6
125	11.9	7.7	11.5#









#89
 Benzo(a)pyrene
 Concen: 10.32 ng
 RT: 15.34 min Scan# 1142
 Delta R.T. -0.00 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC010


Tgt Ion	Resp	Lower	Upper
252	126312		
253	21.5	17.4	26.0
125	8.7	10.3	15.5#



#90
 Dibenzo(a,h)anthracene
 Concen: 10.02 ng
 RT: 16.78 min Scan# 1268
 Delta R.T. -0.01 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Tgt Ion	Resp	Lower	Upper
278	102255		
139	19.9	14.4	21.6
279	22.6	19.1	28.7





#91
 Benzo(g,h,i)perylene
 Concen: 10.20 ng
 RT: 17.19 min Scan# 1304
 Delta R.T. -0.01 min
 Lab File: BF085706.D
 Acq: 24 Mar 2016 10:23

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC010

Tgt Ion	Resp	Lower	Upper
276	100764		
277	23.8	18.8	28.2
138	23.5	22.6	34.0

