

Data Path : Z:\HPCHEM1\BNA F\DATA\BF032417\
 Data File : BF093907.D
 Acq On : 24 Mar 2017 14:40
 Operator : SJ/MA
 Sample : I2367-11
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 E2-P14-NSW

Manual Integrations
 APPROVED

mohammad
 3/27/2017 12:56:06 PM

Quant Time: Mar 25 00:22:09 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF032217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 23 02:02:22 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	5.97	152	184008	20.00	ng	-0.02
21) Naphthalene-d8	7.47	136	728049	20.00	ng	-0.02
38) Acenaphthene-d10	9.62	164	306028	20.00	ng	-0.02
63) Phenanthrene-d10	11.44	188	474627	20.00	ng	-0.02
75) Chrysene-d12	14.70	240	329004	20.00	ng	-0.02
86) Perylene-d12	16.34	264	273716	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	4.50	112	492659	45.22	ng	0.00
7) Phenol-d6	5.56	99	719109	53.36	ng	0.00
23) Nitrobenzene-d5	6.61	82	490170	38.42	ng	-0.03
41) 2,4,6-Tribromophenol	10.59	330	181640	75.11	ng	-0.02
44) 2-Fluorobiphenyl	8.80	172	1091779	54.42	ng	-0.02
78) Terphenyl-d14	13.42	244	935730	63.75	ng	-0.02
Target Compounds						Qvalue
49) Dimethylphthalate	9.29	163	97122	4.46	ng	99
51) Acenaphthene	9.65	154	76120	4.06	ng	95
57) Fluorene	10.29	166	53449	2.53	ng	98
70) Phenanthrene	11.47	178	850729	32.22	ng	99
71) Anthracene	11.53	178	210171	7.73	ng	98
72) Carbazole	11.73	167	126415	4.53	ng	96
74) Fluoranthene	12.94	202	1428743	52.85	ng	98
77) Pyrene	13.22	202	1243173	52.07	ng	97
80) Benzo(a)anthracene	14.69	228	604065	31.49	ng	99
82) Chrysene	14.73	228	598561	33.62	ng	97
85) Indeno(1,2,3-cd)pyrene	17.72	276	252380	16.37	ng	94
87) Benzo(b)fluoranthene	15.93	252	693902m	41.36	ng	
88) Benzo(k)fluoranthene	15.95	252	268275m	16.34	ng	
89) Benzo(a)pyrene	16.27	252	449558	29.10	ng	96
90) Dibenzo(a,h)anthracene	17.74	278	66592	5.09	ng	95
91) Benzo(g,h,i)perylene	18.13	276	234155	17.76	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF032417\
Data File : BF093907.D
Acq On : 24 Mar 2017 14:40
Operator : SJ/MA
Sample : I2367-11
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
E2-P14-NSW

Manual Integrations
APPROVED

mohammad
3/27/2017 12:56:06 PM

Quant Time: Mar 25 00:22:09 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF032217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 23 02:02:22 2017
Response via : Initial Calibration

