

Data Path : Z:\HPCHEM1\BNA F\DATA\BF032417\
 Data File : BF093909.D
 Acq On : 24 Mar 2017 15:34
 Operator : SJ/MA
 Sample : I2367-08
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 E2-BB10-NSW

Quant Time: Mar 25 00:25:33 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF032217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 23 02:02:22 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	5.97	152	175582	20.00	ng	-0.02
21) Naphthalene-d8	7.47	136	661611	20.00	ng	-0.02
38) Acenaphthene-d10	9.62	164	262427	20.00	ng	-0.02
63) Phenanthrene-d10	11.45	188	382951	20.00	ng	-0.02
75) Chrysene-d12	14.74	240	296947	20.00	ng	0.02
86) Perylene-d12	16.36	264	232031	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	4.50	112	1212118	116.60	ng	0.00
7) Phenol-d6	5.57	99	1467583	114.12	ng	0.00
23) Nitrobenzene-d5	6.62	82	898334	77.49	ng	-0.02
41) 2,4,6-Tribromophenol	10.59	330	234946	113.30	ng	-0.02
44) 2-Fluorobiphenyl	8.80	172	1498381	87.09	ng	-0.02
78) Terphenyl-d14	13.45	244	761749	57.50	ng	0.01
Target Compounds						Qvalue
9) Benzaldehyde	5.46	77	25028	2.88	ng	95
31) Naphthalene	7.50	128	133580	4.20	ng	99
37) 2-Methylnaphthalene	8.34	142	60373	2.85	ng	98
49) Dimethylphthalate	9.29	163	141015	7.56	ng	98
54) Dibenzofuran	9.87	168	47345	2.16	ng	# 97
57) Fluorene	10.29	166	45915	2.53	ng	# 94
70) Phenanthrene	11.47	178	143339	6.73	ng	98
74) Fluoranthene	12.96	202	57420	2.63	ng	# 71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF032417\
Data File : BF093909.D
Acq On : 24 Mar 2017 15:34
Operator : SJ/MA
Sample : I2367-08
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
E2-BB10-NSW

Quant Time: Mar 25 00:25:33 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF032217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 23 02:02:22 2017
Response via : Initial Calibration

