

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032422\
 Data File : BF127495.D
 Acq On : 24 Mar 2022 11:21
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 24 15:19:55 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF032122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 22 02:16:33 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	133	0.00
2	1,4-Dioxane	40.000	46.623	-16.6	168	0.00
3	Pyridine	40.000	40.583	-1.5	136	0.00
4	n-Nitrosodimethylamine	40.000	40.614	-1.5	137	0.00
5 S	2-Fluorophenol	80.000	77.890	2.6	132	0.00
6	Aniline	40.000	38.734	3.2	129	0.00
7 S	Phenol-d6	80.000	80.565	-0.7	137	0.00
8	2-Chlorophenol	40.000	39.258	1.9	131	0.00
9	Benzaldehyde	40.000	43.989	-10.0	141	0.00
10 C	Phenol	40.000	39.322	1.7	131	0.00
11	bis(2-Chloroethyl)ether	40.000	40.653	-1.6	137	0.00
12	1,3-Dichlorobenzene	40.000	38.221	4.4	130	0.00
13 C	1,4-Dichlorobenzene	40.000	38.724	3.2	131	0.00
14	1,2-Dichlorobenzene	40.000	37.107	7.2	133	0.00
15	Benzyl Alcohol	40.000	39.708	0.7	132	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.247	-0.6	135	0.00
17	2-Methylphenol	40.000	39.188	2.0	133	0.00
18	Hexachloroethane	40.000	38.785	3.0	132	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.871	2.8	134	0.00
20	3+4-Methylphenols	40.000	38.513	3.7	131	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	134	0.00
22	Acetophenone	40.000	38.034	4.9	133	0.00
23 S	Nitrobenzene-d5	80.000	77.687	2.9	134	0.00
24	Nitrobenzene	40.000	39.916	0.2	136	0.00
25	Isophorone	40.000	41.371	-3.4	138	0.00
26 C	2-Nitrophenol	40.000	40.786	-2.0	138	0.00
27	2,4-Dimethylphenol	40.000	40.018	-0.0	133	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.896	-2.2	140	0.00
29 C	2,4-Dichlorophenol	40.000	40.100	-0.3	136	0.00
30	1,2,4-Trichlorobenzene	40.000	38.794	3.0	134	0.00
31	Naphthalene	40.000	38.641	3.4	133	0.00
32	Benzoic acid	40.000	41.851	-4.6	144	0.02
33	4-Chloroaniline	40.000	39.971	0.1	136	0.00
34 C	Hexachlorobutadiene	40.000	38.694	3.3	133	0.00
35	Caprolactam	40.000	42.904	-7.3	140	0.01
36 C	4-Chloro-3-methylphenol	40.000	40.420	-1.1	137	0.00
37	2-Methylnaphthalene	40.000	39.403	1.5	135	0.00
38	1-Methylnaphthalene	40.000	38.053	4.9	132	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	134	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.639	5.9	131	0.00
41 P	Hexachlorocyclopentadiene	40.000	37.058	7.4	126	0.00
42 S	2,4,6-Tribromophenol	80.000	77.055	3.7	135	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.835	0.4	135	0.00
44	2,4,5-Trichlorophenol	40.000	41.162	-2.9	134	0.00
45 S	2-Fluorobiphenyl	80.000	72.722	9.1	130	0.00
46	1,1'-Biphenyl	40.000	37.531	6.2	133	0.00
47	2-Chloronaphthalene	40.000	38.087	4.8	133	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032422\
 Data File : BF127495.D
 Acq On : 24 Mar 2022 11:21
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 24 15:19:55 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF032122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 22 02:16:33 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.175	-0.4	138	0.00
49	Acenaphthylene	40.000	37.432	6.4	129	0.00
50	Dimethylphthalate	40.000	39.568	1.1	138	0.00
51	2,6-Dinitrotoluene	40.000	39.112	2.2	135	0.00
52 C	Acenaphthene	40.000	38.672	3.3	133	0.00
53	3-Nitroaniline	40.000	38.809	3.0	133	0.00
54 P	2,4-Dinitrophenol	40.000	57.062	-42.7#	213	0.00
55	Dibenzofuran	40.000	39.457	1.4	129	0.00
56 P	4-Nitrophenol	40.000	42.846	-7.1	136	0.00
57	2,4-Dinitrotoluene	40.000	37.106	7.2	128	0.00
58	Fluorene	40.000	39.676	0.8	134	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.214	2.0	138	0.00
60	Diethylphthalate	40.000	38.858	2.9	138	0.00
61	4-Chlorophenyl-phenylether	40.000	36.508	8.7	135	0.00
62	4-Nitroaniline	40.000	38.540	3.7	135	0.00
63	Azobenzene	40.000	38.802	3.0	137	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	127	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.332	-5.8	130	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.018	-0.0	133	0.00
67	4-Bromophenyl-phenylether	40.000	40.473	-1.2	133	0.00
68	Hexachlorobenzene	40.000	39.934	0.2	132	0.00
69	Atrazine	40.000	41.073	-2.7	131	0.00
70 C	Pentachlorophenol	40.000	42.653	-6.6	139	0.00
71	Phenanthrene	40.000	38.916	2.7	131	0.00
72	Anthracene	40.000	38.972	2.6	130	0.00
73	Carbazole	40.000	38.040	4.9	126	0.00
74	Di-n-butylphthalate	40.000	41.389	-3.5	136	0.00
75 C	Fluoranthene	40.000	36.392	9.0	120	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	110	0.00
77	Benzydine	40.000	20.817	48.0#	64	0.00
78	Pyrene	40.000	41.169	-2.9	120	0.00
79 S	Terphenyl-d14	80.000	81.852	-2.3	119	0.00
80	Butylbenzylphthalate	40.000	42.621	-6.6	117	0.00
81	Benzo(a)anthracene	40.000	39.263	1.8	112	0.00
82	3,3'-Dichlorobenzidine	40.000	40.331	-0.8	115	0.00
83	Chrysene	40.000	39.650	0.9	113	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.944	-4.9	115	0.00
85 c	Di-n-octyl phthalate	40.000	50.404	-26.0#	139	0.00
86 I	Perylene-d12	20.000	20.000	0.0	148	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.318	-0.8	141	0.01
88	Benzo(b)fluoranthene	40.000	36.966	7.6	136	0.00
89	Benzo(k)fluoranthene	40.000	35.913	10.2	144	0.00
90 C	Benzo(a)pyrene	40.000	39.486	1.3	149	0.00
91	Dibenzo(a,h)anthracene	40.000	40.868	-2.2	142	0.00
92	Benzo(g,h,i)perylene	40.000	39.609	1.0	138	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF032422\
Data File : BF127495.D
Acq On : 24 Mar 2022 11:21
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Mar 24 15:19:55 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF032122.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 22 02:16:33 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 1