

Data Path : Z:\HPCHEM1\BNA F\DATA\BF032515\
 Data File : BF077956.D
 Acq On : 25 Mar 2015 17:30
 Operator : TP/IZ
 Sample : G1612-01
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 MW-1

Quant Time: Mar 26 03:49:33 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF031115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 26 00:45:24 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.09	152	34825	20.00	ng	0.00
21) Naphthalene-d8	8.67	136	146146	20.00	ng	0.00
38) Acenaphthene-d10	10.84	164	76579	20.00	ng	0.00
63) Phenanthrene-d10	12.67	188	147167	20.00	ng	-0.01
75) Chrysene-d12	16.02	240	167218	20.00	ng	0.05
86) Perylene-d12	17.94	264	165885	20.00	ng	0.19
System Monitoring Compounds						
5) 2-Fluorophenol	5.38	112	124255	62.28	ng	0.00
7) Phenol-d6	6.67	99	98812	40.09	ng	0.00
23) Nitrobenzene-d5	7.79	82	202592	90.87	ng	0.00
41) 2,4,6-Tribromophenol	11.83	330	92910	126.81	ng	0.00
44) 2-Fluorobiphenyl	10.02	172	438923	84.23	ng	0.00
78) Terphenyl-d14	14.68	244	554854	81.05	ng	0.01
Target Compounds						
2) 1,4-Dioxane	1.97	88	1212655	1338.74	ng	Qvalue # 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF032515\
Data File : BF077956.D
Acq On : 25 Mar 2015 17:30
Operator : TP/IZ
Sample : G1612-01
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
MW-1

Quant Time: Mar 26 03:49:33 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF031115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 26 00:45:24 2015
Response via : Initial Calibration

