

Data Path : Z:\HPCHEM1\BNA F\DATA\BF032516\
 Data File : BF085767.D
 Acq On : 25 Mar 2016 17:52
 Operator : UM/SJ
 Sample : H1855-03
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-1A

Manual Integrations
 APPROVED

Sohil
 3/28/2016 7:56:03 PM

Quant Time: Mar 25 23:20:07 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF032416.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 24 14:15:39 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.82	152	116117	20.00	ng	0.00
21) Naphthalene-d8	8.10	136	445625	20.00	ng	-0.01
38) Acenaphthene-d10	9.86	164	198930	20.00	ng	0.00
63) Phenanthrene-d10	11.35	188	314594	20.00	ng	0.00
75) Chrysene-d12	13.99	240	231516	20.00	ng	0.01
86) Perylene-d12	15.41	264	221713	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.42	112	68546	9.83	ng	0.01
7) Phenol-d6	6.46	99	87408	9.86	ng	0.00
23) Nitrobenzene-d5	7.38	82	43576	5.51	ng	-0.01
41) 2,4,6-Tribromophenol	10.65	330	13924	7.37	ng	0.00
44) 2-Fluorobiphenyl	9.18	172	80664	6.77	ng	-0.01
78) Terphenyl-d14	12.93	244	57681	5.12	ng	0.00
Target Compounds						Qvalue
49) Dimethylphthalate	9.58	163	131038	8.68	ng	99
51) Acenaphthene	9.90	154	78978	6.43	ng	99
54) Dibenzofuran	10.07	168	57622	3.62	ng	95
57) Fluorene	10.41	166	87083	6.59	ng	99
70) Phenanthrene	11.37	178	1203387	74.50	ng	99
71) Anthracene	11.42	178	375441	22.14	ng	99
72) Carbazole	11.58	167	80576	5.66	ng	98
74) Fluoranthene	12.56	202	1943764	113.21	ng	95
77) Pyrene	12.79	202	1563442	82.15	ng	98
80) Benzo(a)anthracene	13.98	228	975510	73.90	ng	97
82) Chrysene	14.01	228	802721m	60.37	ng	
85) Indeno(1,2,3-cd)pyrene	16.80	276	400243	38.98	ng	94
87) Benzo(b)fluoranthene	15.01	252	1027321m	73.55	ng	
88) Benzo(k)fluoranthene	15.03	252	378632m	30.07	ng	
89) Benzo(a)pyrene	15.35	252	707999m	57.09	ng	
90) Dibenzo(a,h)anthracene	16.80	278	113253	9.83	ng	95
91) Benzo(g,h,i)perylene	17.21	276	333579	28.59	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF032516\
Data File : BF085767.D
Acq On : 25 Mar 2016 17:52
Operator : UM/SJ
Sample : H1855-03
Misc :
ALS Vial : 53 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-1A

Manual Integrations
APPROVED

Sohil
3/28/2016 7:56:03 PM

Quant Time: Mar 25 23:20:07 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF032416.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 24 14:15:39 2016
Response via : Initial Calibration

