

Data Path : Z:\HPCHEM1\BNA F\DATA\BF032516\
 Data File : BF085779.D
 Acq On : 25 Mar 2016 23:38
 Operator : UM/SJ
 Sample : H1834-05
 Misc :
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 4

Manual Integrations
 APPROVED

Sohil
 3/28/2016 8:00:52 PM

Quant Time: Mar 28 14:56:40 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF032416.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 24 14:15:39 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.82	152	102206	20.00	ng	0.00
21) Naphthalene-d8	8.12	136	263224	20.00	ng	0.00
38) Acenaphthene-d10	9.89	164	115561	20.00	ng	0.02
63) Phenanthrene-d10	11.37	188	173195	20.00	ng	0.02
75) Chrysene-d12	13.99	240	202369	20.00	ng	0.01
86) Perylene-d12	15.40	264	175062	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.43	112	802240	130.72	ng	0.02
7) Phenol-d6	6.47	99	1021577	130.93	ng	0.01
23) Nitrobenzene-d5	7.40	82	538192	115.14	ng	0.00
41) 2,4,6-Tribromophenol	10.68	330	140108	127.61	ng	0.02
44) 2-Fluorobiphenyl	9.20	172	661817	95.68	ng	0.01
78) Terphenyl-d14	12.94	244	485695	49.36	ng	0.01
Target Compounds						
31) Naphthalene	8.15	128	1259327	94.95	ng	Qvalue 96
37) 2-Methylnaphthalene	8.86	142	2386975	285.81	ng	# 95
45) 1,1'-Biphenyl	9.30	154	315584	32.36	ng	95
49) Dimethylphthalate	9.60	163	124679	14.22	ng	# 84
51) Acenaphthene	9.92	154	146643	20.55	ng	# 65
57) Fluorene	10.44	166	281895	36.71	ng	# 92
70) Phenanthrene	11.40	178	965278	108.54	ng	97
77) Pyrene	12.80	202	189219	11.37	ng	# 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF032516\
Data File : BF085779.D
Acq On : 25 Mar 2016 23:38
Operator : UM/SJ
Sample : H1834-05
Misc :
ALS Vial : 46 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
4

Manual Integrations
APPROVED

Sohil
3/28/2016 8:00:52 PM

Quant Time: Mar 28 14:56:40 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF032416.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 24 14:15:39 2016
Response via : Initial Calibration

