

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF032520\
 Data File : BF119521.D
 Acq On : 26 Mar 2020 1:37
 Operator : CG/JU
 Sample : L2013-08 10X
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 GHD1-SB-01-0-52

Quant Time: Mar 26 07:17:00 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF031820.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 18 15:11:18 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.84	152	308628	20.00	ng	-0.02
21) Naphthalene-d8	8.13	136	1208597	20.00	ng	-0.02
39) Acenaphthene-d10	9.89	164	637458	20.00	ng	-0.02
64) Phenanthrene-d10	11.37	188	1189883	20.00	ng	-0.02
76) Chrysene-d12	14.01	240	767749	20.00	ng	-0.02
87) Perylene-d12	15.48	264	618306	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.44	112	214915	11.09	ng	-0.02
7) Phenol-d6	6.45	99	288513	11.65	ng	-0.02
23) Nitrobenzene-d5	7.40	82	169826	7.64	ng	-0.02
42) 2,4,6-Tribromophenol	10.67	330	78286	11.90	ng	-0.02
45) 2-Fluorobiphenyl	9.20	172	342190	7.95	ng	-0.02
79) Terphenyl-d14	12.96	244	361274	9.22	ng	-0.02
Target Compounds						
84) Bis(2-ethylhexyl)phthalate	14.00	149	79855	2.629	ng	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF032520\
Data File : BF119521.D
Acq On : 26 Mar 2020 1:37
Operator : CG/JU
Sample : L2013-08 10X
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
GHD1-SB-01-0-52

Quant Time: Mar 26 07:17:00 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF031820.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 18 15:11:18 2020
Response via : Initial Calibration

